1

PM-1 3½"x4" PHOTOGRAPHIC MICROCOPY TARGET
NBS 1010a ANSI/ISO #2 EQUIVALENT

PRECISIONSM RESOLUTION TARGETS

National Library
of Canada

Acquisitions and
Bibliographic Services Branch

395 Wellington Street
Ottawa, Ontario
K1A 0N4

Bibliothèque nationale
du Canada

Direction des acquisitions et
des services bibliographiques

395, rue Wellington
Ottawa (Ontario)
K1A 0N4

Your file Votre référence

Our file Notre référence

NOTICE

The quality of this microform is heavily dependent upon the quality of the original thesis submitted for microfilming. Every effort has been made to ensure the highest quality of reproduction possible.

If pages are missing, contact the university which granted the degree.

Some pages may have indistinct print especially if the original pages were typed with a poor typewriter ribbon or if the university sent us an inferior photocopy.

Reproduction in full or in part of this microform is governed by the Canadian Copyright Act, R.S.C. 1970, c. C-30, and subsequent amendments.

AVIS

La qualité de cette microforme dépend grandement de la qualité de la thèse soumise au microfilmage. Nous avons tout fait pour assurer une qualité supérieure de reproduction.

S'il manque des pages, veuillez communiquer avec l'université qui a conféré le grade.

La qualité d'impression de certaines pages peut laisser à désirer, surtout si les pages originales ont été dactylographiées à l'aide d'un ruban usé ou si l'université nous a fait parvenir une photocopie de qualité inférieure.

La reproduction, même partielle, de cette microforme est soumise à la Loi canadienne sur le droit d'auteur, SRC 1970, c. C-30, et ses amendements subséquents.

The Effects Of Hyperthermia On Tissue Blood Flow

By

Ufuk Akyürekli

A thesis submitted to the Faculty of Graduate Studies and Research in partial fulfillment of the requirements for the degree of Doctor of Philosophy

Department of Physics
Carleton University
Ottawa Ontario
Canada
August, 1995

© copyright

National Library
of Canada

Acquisitions and
Bibliographic Services Branch

395 Wellington Street
Ottawa, Ontario
K1A 0N4

Bibliothèque nationale
du Canada

Direction des acquisitions et
des services bibliographiques

395, rue Wellington
Ottawa (Ontario)
K1A 0N4

Your file Votre référence

Our file Notre référence

The author has granted an irrevocable non-exclusive licence allowing the National Library of Canada to reproduce, loan, distribute or sell copies of his/her thesis by any means and in any form or format, making this thesis available to interested persons.

L'auteur a accordé une licence irrévocable et non exclusive permettant à la Bibliothèque nationale du Canada de reproduire, prêter, distribuer ou vendre des copies de sa thèse de quelque manière et sous quelque forme que ce soit pour mettre des exemplaires de cette thèse à la disposition des personnes intéressées.

The author retains ownership of the copyright in his/her thesis. Neither the thesis nor substantial extracts from it may be printed or otherwise reproduced without his/her permission.

L'auteur conserve la propriété du droit d'auteur qui protège sa thèse. Ni la thèse ni des extraits substantiels de celle-ci ne doivent être imprimés ou autrement reproduits sans son autorisation.

ISBN 0-612-08833-2

Canada

Name U. M. I.

Dissertation Abstracts International is arranged by broad, general subject categories. Please select the one subject which most nearly describes the content of your dissertation. Enter the corresponding four-digit code in the spaces provided.

U·M·I

SUBJECT TERM

SUBJECT CODE

Subject Categories

THE HUMANITIES AND SOCIAL SCIENCES

COMMUNICATIONS AND THE ARTS

Architecture 0729
Art History 0377
Cinema 0900
Dance 0378
Fine Arts 0357
Information Science 0723
Journalism 0391
Library Science 0399
Mass Communications 0708
Music 0413
Speech Communication 0459
Theater 0465

EDUCATION

General 0515
Administration 0514
Adult and Continuing 0516
Agricultural 0517
Art 0273
Bilingual and Multicultural 0282
Business 0688
Community College 0275
Curriculum and Instruction 0727
Early Childhood 0518
Elementary 0524
Finance 0277
Guidance and Counseling 0519
Health 0680
Higher 0745
History of 0520
Home Economics 0278
Industrial 0521
Language and Literature 0279
Mathematics 0280
Music 0522
Philosophy of 0998
Physical 0523

Psychology 0525
Reading 0535
Religious 0527
Sciences 0714
Secondary 0533
Social Sciences 0534
Sociology of 0340
Special 0529
Teacher Training 0530
Technology 0710
Tests and Measurements 0288
Vocational 0747

LANGUAGE, LITERATURE AND LINGUISTICS

Language
General 0679
Ancient 0289
Linguistics 0290
Modern 0291
Literature
General 0401
Classical 0294
Comparative 0295
Medieval 0297
Modern 0298
African 0316
American 0591
Asian 0305
Canadian (English) 0352
Canadian (French) 0355
English 0593
Germanic 0311
Latin American 0312
Middle Eastern 0315
Romance 0313
Slavic and East European 0314

PHILOSOPHY, RELIGION AND THEOLOGY

Philosophy 0422
Religion
General 0318
Biblical Studies 0321
Clergy 0319
History of 0320
Philosophy of 0322
Theology 0469

SOCIAL SCIENCES

American Studies 0323
Anthropology
Archaeology 0324
Cultural 0326
Physical 0327
Business Administration
General 0310
Accounting 0272
Banking 0770
Management 0454
Marketing 0338
Canadian Studies 0385
Economics
General 0501
Agricultural 0503
Commerce-Business 0505
Finance 0508
History 0509
Labor 0510
Theory 0511
Folklore 0358
Geography 0366
Gerontology 0351
History
General 0578

Ancient 0579
Medieval 0581
Modern 0582
Black 0328
African 0331
Asia, Australia and Oceania 0332
Canadian 0334
European 0335
Latin American 0336
Middle Eastern 0333
United States 0337
History of Science 0585
Law 0398
Political Science
General 0615
International Law and Relations 0616
Public Administration 0617
Recreation 0914
Social Work 0452
Sociology
General 0626
Criminology and Penology 0627
Demography 0938
Ethnic and Racial Studies 0631
Individual and Family Studies 0628
Industrial and Labor Relations 0629
Public and Social Welfare 0630
Social Structure and Development 0700
Theory and Methods 0344
Transportation 0709
Urban and Regional Planning 0999
Women's Studies 0453

THE SCIENCES AND ENGINEERING

BIOLOGICAL SCIENCES

Agriculture
General 0473
Agronomy 0285
Animal Culture and Nutrition 0475
Animal Pathology 0476
Food Science and Technology 0359
Forestry and Wildlife 0478
Plant Culture 0479
Plant Pathology 0480
Plant Physiology 0817
Range Management 0777
Wood Technology 0746
Biology
General 0306
Anatomy 0287
Biostatistics 0308
Botany 0309
Cell 0379
Ecology 0329
Entomology 0353
Genetics 0369
Limnology 0793
Microbiology 0410
Molecular 0307
Neuroscience 0317
Oceanography 0416
Physiology 0433
Radiation 0821
Veterinary Science 0778
Zoology 0472
Biophysics
General 0786
Medical 0760
EARTH SCIENCES
Biogeochemistry 0425
Geochemistry 0996

Geodesy 0370
Geology 0372
Geophysics 0373
Hydrology 0388
Mineralogy 0411
Paleobotany 0345
Paleoecology 0426
Paleontology 0418
Paleozoology 0985
Palynology 0427
Physical Geography 0368
Physical Oceanography 0415

HEALTH AND ENVIRONMENTAL SCIENCES

Environmental Sciences 0768
Health Sciences
General 0566
Audiology 0300
Chemotherapy 0992
Dentistry 0567
Education 0350
Hospital Management 0769
Human Development 0758
Immunology 0982
Medicine and Surgery 0564
Mental Health 0347
Nursing 0569
Nutrition 0570
Obstetrics and Gynecology 0380
Occupational Health and Therapy 0354
Ophthalmology 0381
Pathology 0571
Pharmacology 0419
Pharmacy 0572
Physical Therapy 0382
Public Health 0573
Radiology 0574
Recreation 0575

Speech Pathology 0460
Toxicology 0383
Home Economics 0386

PHYSICAL SCIENCES

Pure Sciences
Chemistry
General 0485
Agricultural 0749
Analytical 0486
Biochemistry 0487
Inorganic 0488
Nuclear 0738
Organic 0490
Pharmaceutical 0491
Physical 0494
Polymer 0495
Radiation 0754
Mathematics 0405
Physics
General 0605
Acoustics 0986
Astronomy and Astrophysics 0606
Atmospheric Science 0608
Atomic 0748
Electronics and Electricity 0607
Elementary Particles and High Energy 0798
Fluid and Plasma 0759
Molecular 0609
Nuclear 0610
Optics 0752
Radiation 0756
Solid State 0611
Statistics 0463
Applied Sciences
Applied Mechanics 0346
Computer Science 0984

Engineering
General 0537
Aerospace 0538
Agricultural 0539
Automotive 0540
Biomedical 0541
Chemical 0542
Civil 0543
Electronics and Electrical 0544
Heat and Thermodynamics 0348
Hydraulic 0545
Industrial 0546
Marine 0547
Materials Science 0794
Mechanical 0548
Metallurgy 0743
Mining 0551
Nuclear 0552
Packaging 0549
Petroleum 0765
Sanitary and Municipal 0554
System Science 0790
Geotechnology 0428
Operations Research 0796
Plastics Technology 0795
Textile Technology 0994

PSYCHOLOGY

General 0621
Behavioral 0384
Clinical 0622
Developmental 0620
Experimental 0623
Industrial 0624
Personality 0625
Physiological 0989
Psychobiology 0349
Psychometrics 0632
Social 0451

1995. Ufuk Akyürekli

The undersigned recommend to the Faculty of Graduate Studies and Research
acceptance of the thesis

" The Effects Of Hyperthermia On Tissue Blood Flow"

submitted by

Ufuk Akyürekli, B.Sc., M.Sc.

in partial fulfillment for the degree of Doctor of Philosophy

Chairman, Department of Physics

Thesis Supervisor

External Examiner

Carleton University
August, 1995

Abstract

This study investigates the changes in normal porcine tissue blood flow before, during and following a period of regional microwave hyperthermia. In order to isolate the effects of heat on tissue blood flow, control experiments were undertaken to establish the vascular changes occurring as a consequence of the surgical protocol. The vascular effect of heat was investigated in kidney, brain and muscle. Blood flow was measured using the radiolabeled microsphere reference sample technique.

An important part of this investigation was the development of a technique for preparing standardized tissue samples for microsphere assay of tissue blood flow. The validity and reproducibility of this technique has been validated for planar configured gamma-ray detectors using an *in-situ* porcine kidney model. A positive correlation coefficient of $R=0.99$ was observed between the total kidney blood flow as measured by a calibrated ultrasonic flow probe and the microsphere reference sample technique.

Each of the tissues investigated responded differently to thermal stress, where only muscle showed the classical response to heat. The effect of heat on renal blood flow was examined in both an autoperfused and an *in-situ* kidney model. Both models showed a dramatic decrease in blood flow during heating. The response of regional cerebral blood flow (rCBF) to the tissue temperature was determined for both modest and severe hyperthermia. Results of our study show that when heated with modest power levels (23.5 Watts) temperatures in the range 42.6 - 43.8°C are reached, resulting in a gradual increase in blood flow over the entire heating period. In contrast, cerebral blood flow measured in animals heated at higher power levels

(27 Watts) to temperatures $>46^{\circ}\text{C}$, rCBF decreased. In the case of muscle, blood flow increased by an average factor of 4 within 15-30 minutes of heating at 45°C and decreased with further heating. The greatest increase in blood flow was observed just beneath the center of the applicator. The increase in blood flow appears only in the region of elevated temperatures, with tissues at the higher temperature receiving greater flow than those at the lower temperatures.

Finally, a simple model, namely, the one dimensional bioheat transfer equation, was used to examine the effects of blood perfusion and surface cooling on the tissue temperature distribution. The calculated tissue temperature history was compared to an *in-vivo* muscle temperature measurement made under hyperthermic conditions generated by a 915 MHz microwave applicator. Blood flow parameters used in the model were also obtained experimentally. The results show that the simple bioheat transfer equation can quantitatively reproduce the steady-state features of the temperature pattern seen experimentally.

Acknowledgments

I would like to express my sincere gratitude to my thesis supervisors Dr. L. H. Gerig and Dr. G. P. Raaphorst for their guidance, support and patience throughout the years. I am grateful to all my Professors at Carleton university for all their help throughout the many years. I am very thankful to Jack Scott from the National Research Council Institute of Biodiagnostics for all his invaluable help. I would also like to express my deep appreciation to everyone at the National Research Council Ionizing Radiation Standards group, especially, Dr. L. van der Zwan, Leo Hisek, Dr. D. Santry, and Dr. C. Ross for their invaluable help. Many thanks to Dr. D Salhani from the Ottawa Regional Cancer Centre for his help and words of encouragement, also to Mo, and Yeve from dosimetry, for secretly finding the time to CT the all too many carcasses. On a personal note, I am indebted to Esra, Tibet, Yasar, Halide, Sema, Engin, and Deniz for their patience, understanding, and encouragement.

Preface

The use of hyperthermia for the treatment of cancer is certainly not new. The first record of a patient treated with hyperthermia is found in the Edwin Smith Surgical Papyrus, an Egyptian papyrus, which dates back more than 5,000 years. However, the greatest interest and activity regarding the use of heat in the treatment of human malignancies has developed over the last two decades. Laboratory studies with cells and animals conducted in this period have firmly established the potential of hyperthermia as an anti-cancer agent. A growing body of clinical data strongly suggest that the use of hyperthermia as an adjuvant to radiation or chemotherapy can lead to significantly improved tumor control with only minimal concomitant increase in normal tissue side effects. Despite the optimism, hyperthermia's full potential will not be realized until heat transfer models predicting the tissue temperature distribution become accurate. The lack of data regarding the magnitude, distribution, and response of tissue perfusion and the probable nature of any changes occurring in response to therapy, limit our ability to either confirm, modify or exploit current heat transport theories. Therefore, an understanding of the heat-induced blood flow changes in normal tissues as well as in tumors is important for constructing and evaluating theoretical models for predicting tissue temperature distributions.

The intent of this project was to measure the vascular response of various porcine tissues to hyperthermia. The results of this work have been presented at several scientific meetings of the Radiation Research Society and the North American Hyperthermia Society (New Orleans 1990, Toronto 1991, Dallas 1993, Nashville 1994), the annual meeting of the Canadian Organization of Medical

Physicists (Ottawa 1993) and at the local level, at several meetings of the Medical Physics Organized Research Unit (MPORU) and physics graduate seminars.

The study used a total of 140 pigs weighing between 20 and 65 kg with a mean weight of 40 kg and generated over 3000 tissue and blood samples, a similar number of gamma-ray spectra and over 1000 transverse CT images. This thesis is a compilation of manuscripts submitted to several scientific journals for review, with the exception of chapters one, seven and eight. Each chapter constitutes a distinct manuscript. The manuscripts have been incorporated directly into the thesis as follows:

1. Thesis chapter 2, "A tissue preparation technique for microsphere assays of blood flow" to be published in *Animal Laboratory Science* Vol. 45, No. 4 August 1995.
2. Thesis chapter 3, "Changes in blood flow in a hyperthermic autoperfused kidney", submitted to *European Journal of Hyperthermia*.
3. Thesis chapter 4, "Changes in renal blood flow during regional hyperthermia", submitted to *International Journal of Hyperthermia*.
4. Thesis chapter 5, "Changes in cerebral blood flow during interstitial hyperthermia", submitted to *Journal of Neurosurgery*.
5. Thesis chapter 6, "Porcine muscle blood flow distribution during hyperthermia", submitted to *International Journal of Hyperthermia*.

Contents

	Page
Acceptance Sheet	ii
Abstract	iii
Acknowledgments	v
Preface	vi
Table of Contents	viii
List of Figures	xii
List of Tables	xviii
 1 General overview	
1.1 Introduction	1
1.2 Rationale for hyperthermia	3
1.2.1 Cellular Mechanisms of Heat Inactivation	3
1.2.2 Environmental Factors	4
1.3 Objectives and experimental rationale	6
1.4 Materials and methods	8
1.4.1 Animals	8
1.4.2 Surgical preparation	8
1.4.3 Blood flow measurements	9
1.4.4 Measurement of γ -ray spectra	11
1.4.5 Calculation of blood flow	12
1.4.6 Electromagnetic flow probe	12
1.4.6 Ultrasonic flow probe	13
1.4.7 Hyperthermic system	13

2 A tissue preparation technique for microsphere assays of blood flow

2.1	Introduction	18
2.2	Materials and Methods	19
2.2.1	Animal preparation for microsphere measurements	19
2.2.2	Calculation of blood flow	21
2.2.3	Preparation of counting samples	21
2.2.4	Microsphere leaching and dispersion	22
2.2.5	Geometric reproducibility of centrifuged samples	23
2.2.6	Validation of the tissue preparation technique	23
2.2.7	Measurement of γ -ray spectra	23
2.3	Results	24
2.4	Discussion	25

3 Measurement of blood flow in an autoperfused kidney

3.1	Introduction	33
3.2	Materials And Methods	34
3.2.1	Animals	34
3.2.2	Hyperthermia system	34
3.2.3	Kidney retrieval and re-perfusion	35
3.2.4	Microsphere blood flow measurement	36
3.3	Results	37
3.3.1	Effect of Anesthesia alone	37
3.3.2	Normothermic autoperfused kidney	37
3.3.3	Hyperthermic autoperfused group	38
3.3.4	Comparison of blood flow measurements	40
3.4	Discussion	40

4 Changes in renal blood flow during regional hyperthermia

4.1	Introduction	58
4.2	Materials And Methods	59
4.2.1	Animals	59
4.2.2	Hyperthermia system	59
4.2.3	Preparation for heating	60
2.2.4	Preparation for microsphere measurements	60
4.3	Results	61
4.3.1	Effect of prolonged anesthesia	61
4.3.2	Effect of thermocouple insertion	61
4.3.3	Effect of hyperthermia without thermocouple insertion	62
4.3.4	Hyperthermia with thermocouple insertion	63
4.4	Discussion	64

5 Changes in cerebral blood flow during interstitial hyperthermia

5.1	Introduction	82
5.2	Materials And Methods	83
5.2.1	Animals	83
5.2.2	Hyperthermia system	83
5.2.3	Preparation for heating	84
5.2.4	Blood flow measurement	84
5.3	Results	85
5.4	Discussion	87

6 Changes in muscle blood flow during regional hyperthermia

6.1	Introduction	98
6.2	Materials And Methods	99
6.2.1	Animals	99
6.2.2	Hyperthermia system	99
6.2.3	Preparation for heating	100
6.2.4	Preparation for microsphere measurements	100
6.3	Results	101
6.3.1	Effect of anesthesia	101
6.3.2	Effect of thermocouple insertion	102
6.3.3	Effect of hyperthermia after thermocouple insertion	102
6.4	Discussion	104

7 The effect of blood flow on microwave heating of tissue

7.1	Introduction	121
7.2	The animal model	123
7.3	Semi-infinite slab model	123
7.3.1	The one-dimensional bioheat equation	123
7.3.2	Steady-state solution when $q=0$	125
7.3.3	Steady-state solution when $q \neq 0$	126
7.3.4	The transient solution	128
7.3.5	The finite element solution	129
7.3.6	Comparison with experimental results	129
7.3.7	Comparison with the two dimensional finite element model	130
7.4	Discussion	130

8 Conclusion 146

Appendix A	151
References	157

List of Figures

	Page
Figure 1.1 Flow chart illustrating the heating and control experiments.	15
Figure 1.2 This spectrum is derived from a mixture of routinely used radiolabelled microspheres. A Ge(Li) detector was used to obtain the spectrum.	16
Figure 1.3 Longitudinal axis power deposition profile of the 915 MHz microwave applicator.	17
Figure 2.1. Illustration of the source-detector arrangement showing an enlarged view of the counting vial. The reproducibility of the source position is assured by mechanically connecting the sample holder to the detector housing.	29
Figure 2.2. Comparison of the total kidney blood flow as measured by the reference sample microsphere technique and a Transonic flow probe.	30
Figure 2.3. Variation in the standard error σ_s/S for an isotropic point source, as a function of detector to source distance d (plotted from Eq. 2.4) for various count rates in the particular case $R_D = 2.33\text{cm}$, $\sigma_d = 0.05\text{cm}$, $\epsilon_i(E, \Omega) = 0.06$ and for a count time of $t=10$ minutes.	31
Figure 3.1. Diagram of the extracorporeal circuit for the autoperfused porcine kidney. Arterial blood perfusing the isolated kidney is heated using a heat exchanger regulated by a constant temperature circulating water bath. Venous out-flow from the perfused organ is returned to the systemic circulation by way of the femoral vein using a roller pump.	46
Figure 3.2. Illustration of the time sequence of blood flow measurements.	47

- Figure 3.3.** Illustration showing the sectioning of the kidney into transverse tissue samples. Both the autoperfused and contralateral control kidney were removed and sliced into seven or eight 1.5 cm thick transverse sections and weighed. 48
- Figure 3.4** Changes in renal blood flow while anesthetized with halothane. Blood flow remained unchanged in both the left and right kidneys presence of 1.0-1.5% halothane. Mean of 7 determinations $\pm$ S.E.M are shown. 49
- Figure 3.5.** Total blood flow (mean $\pm$ S.E.M.) to the normothermic (38°C) autoperfused kidneys (n=6) and the contralateral control kidneys plotted against the elapsed time. 50
- Figure 3.6.** An example of blood perfusion in the eight tissue samples of the kidney contralateral to the autoperfused normothermic (38°C) kidney plotted against the elapsed time. 51
- Figure 3.7.** An example of blood perfusion in the eight tissue samples of the kidney contralateral to the autoperfused normothermic (38°C) kidney plotted against the elapsed time. 52
- Figure 3.8.** Changes in total blood flow (mean $\pm$ S.E.M.) to the autoperfused kidneys (n=6) heated to 43°C and the control contralateral kidneys plotted against the elapsed time. 53
- Figure 3.9.** An example of blood perfusion in the eight tissue samples of the kidney contralateral to the heated (43°C) autoperfused kidney. 54
- Figure 3.10.** An example of blood perfusion in the eight tissue samples of the kidney contralateral to the heated (43°C) autoperfused kidney. 55
- Figure 3.11.** Comparison of total renal artery blood flow as measured by the microsphere reference sample technique and a cannulating in-line electromagnetic flow probe. 56
- Figure 4.1.** Illustration of the 915 MHz microwave applicator, kidney and thermocouple orientation during hyperthermia treatment of the porcine kidney. 70
- Figure 4.2.** The transverse CT scan of the catheterized kidney where obtained in order to verify the position of the thermocouple arrays. 71

- Figure 4.3.** Illustration of the sequence in which radiolabeled tracer microspheres were administered. The numbers in the figure indicate the order in which microsphere blood flow measurements were made; (1) following mobilization of one kidney, (2) after implanting the thermocouple arrays, (3-5) during heating, and (6) post-hyperthermia. 72
- Figure 4.4.** Illustration showing the sectioning of the kidney into transverse tissue samples. Both the autoperfused and contralateral control kidney were removed and sliced into seven or eight 1.5 cm thick transverse sections and weighed. 73
- Figure 4.5.** Porcine renal blood flow under halothane anesthesia. Blood flow remained unchanged in both the left and right kidneys in the presence of 1.0-1.5% halothane. Mean of 7 determinations $\pm$ S.E.M are shown. 74
- Figure 4.6.** Effect of thermocouple insertion on total renal blood flow. Comparison of the normothermic kidney in which thermocouple were implanted and the contralateral control kidney. Mean of 3 determinations $\pm$ S.E.M are shown. 75
- Figure 4.7.** Changes in total blood flow to both the heated non-catheterized kidney and the contralateral control kidney. Mean of 3 determinations $\pm$ S.E.M. are shown. 76
- Figure 4.8.** An example of the temperature history in the porcine kidney during heating with a surface cooled 915 MHz microwave dielectrically loaded wave guide. The thermocouple was located 1 cm beneath the anterior surface of the kidney and coincident with the principal axis of the wave guide. 77
- Figure 4.9** An example of the thermal distribution in the porcine kidney during heating. The thermocouple array was located 1 cm beneath the anterior surface of the kidney. 78
- Figure 4.10** Changes in total blood flow to both the heated and contralateral control kidney. Mean of 12 determinations $\pm$ S.E.M are shown. 79
- Figure 4.11** An example of the blood flow distribution in six tissue samples of the contralateral control kidney during hyperthermia. 80
- Figure 4.12** An example of the blood flow distribution in the kidney heated to 43°C. 81
- Figure 5.1.** Schematic illustration of the 915 MHz helical antenna used to heat the porcine brain. 90

- Figure 5.2.** Schematic illustration of the 915 MHz helical antenna support and template. 91
- Figure 5.3.** Transverse CT image of the porcine brain showing the position of the thermocouple and antenna catheters. 92
- Figure 5.4.** Illustration of the sequence in which radiolabeled tracer microspheres were administered. 93
- Figure 5.5.** Schematic illustration of the superior aspect of the brain with position of thermocouples, microwave antenna and transverse tissue slices. 94
- Figure 5.6.** Mean blood flow for the control group of animals, and the group of animals heated with 23.5 Watts constant NFP and 27.0 Watts constant NFP. 95
- Figure 5.7.** A representative observation showing the gradual increase in blood flow and decrease in temperature for a normal porcine brain during a 60 minute heating with 23.5 Watts constant NFP. The temperature was monitored with the reference thermocouple 6.5 mm from the antenna catheter surface at the depth of the peak SAR. 96
- Figure 5.8.** A representative observation showing the gradual increase then decrease in blood flow resulting in the increase in tissue temperature for a normal porcine brain during a 60 minute heating with 27 Watts constant NFP. The temperature was monitored with the reference thermocouple 6.5 mm from the antenna catheter surface at the depth of the peak SAR. 97
- Figure 6.1.** Schematic illustration of the coordinate system used to describe the thermocouples' orientation and position during hyperthermia treatment. The muscle beneath the wave guide was sliced into five sections labeled 1 through to 5 and tissue samples labeled "A" and "B" dissected from each slice. 108
- Figure 6.2.** Illustration showing the sequence in which radiolabeled tracer microspheres were administered. The numbers in the figure indicate the order in which microsphere blood flow measurements were made; (1) during anesthesia alone (2) after implanting the thermocouple arrays, (3-5) during heating, and (6) post-hyperthermia. 109
- Figure 6.3.** An example of the transverse computed tomographic image of the treated limb used to verify the position of the thermocouple arrays. 110

- Figure 6.4.** The effects of halothane anesthesia on the muscle blood flow distribution in tissue samples labeled "A". Blood flow remained unchanged in presence of 1.0-1.5% halothane. Mean of 6 determinations $\pm$ S.E.M are shown. 111
- Figure 6.5.** The effects of halothane anesthesia on the muscle blood flow distribution in tissue samples labeled "B". Blood flow remained unchanged in presence of 1.0-1.5% halothane. Mean of 6 determinations $\pm$ S.E.M are shown. 112
- Figure 6.6.** The effects of thermocouple insertion during general anesthesia on the muscle blood flow distribution in the superficial "A" tissue samples. Blood flow remained unchanged after insertion of the thermocouples. Mean of 5 determinations $\pm$ S.E.M. are shown. 113
- Figure 6.7.** The effects of thermocouple insertion during anesthesia on the muscle blood flow distribution in the deeper "B" tissue the samples. Blood flow remained unchanged after insertion of the thermocouple arrays. Mean of 5 determinations $\pm$ S.E.M. are shown. 114
- Figure 6.8.** A representative observation of the temperature history for A and B tissue samples for thermocouples at a depth of 1.5 and 3.5 cm during heating with 40 Watts constant NFP. 115
- Figure 6.9.** The initial rate of heating in the superficial "A" tissue samples as a function of the thermocouple position. 116
- Figure 6.10.** The changes in blood flow distribution in "A" and "B" tissue samples from the volume of tissue beneath the microwave applicator. Mean of 21 determinations $\pm$ S.E.M are shown. 119
- Figure 6.11.** The effects of hyperthermia on the blood flow distribution to the contralateral control muscle in both superficial and deep tissue samples. Mean of 21 determinations $\pm$ S.E.M are shown. 120
- Figure 7.1.** Transverse section through the microwave applicator and muscle showing the coordinate system used to describe the thermocouples orientation and position during hyperthermia treatment. 134
- Figure 7.2.** The variation in the steady-state muscle temperature distribution with perfusion prior to applying power ($q=0$) for the case when $h=500$. 136
- Figure 7.3.** The applied power density as given by equation 7.2 were the net forward power is 36 watts. 137

Figure 7.4. Solution to equation 7.8 (i.e. eqn. 7.9) illustrating the effect of perfusion on the steady-state temperature distribution in microwave irradiated muscle when the heat transfer coefficient is $500 \text{ W/m}^2\text{C}$.	138
Figure 7.5. Solution to equation 7.8 describing the variation in the steady-state muscle temperature distribution with the heat transfer coefficient when perfusion is constant at 10 ml/min/100g .	139
Figure 7.6. Solution to equation 7.8 describing the variation in the steady-state muscle temperature distribution with the heat transfer coefficient when perfusion is constant at 40 ml/min/100g .	140
Figure 7.7. Plot of equation 7.11 describing the variation in the muscle temperature profile with time when perfusion is constant at $w=10\text{ml/min/100g}$, and the heat transfer coefficient is $h=500 \text{ W/m}^2\text{C}$.	141
Figure 7.8. Plot of equation 7.11 describing the variation in the muscle temperature profile with time when perfusion is constant at $w=40\text{ml/min/100g}$, and the heat transfer coefficient is $h=500 \text{ W/m}^2\text{C}$.	142
Figure 7.9. Comparison of the transient solution and measured temperature histories in muscle for a thermocouple 1.5 cm beneath the tissue surface.	143
Figure 7.10. The difference between the transient solution and measured temperature history for a thermocouple 1.5 cm beneath the muscle surface.	144
Figure 7.11 Comparison of the 1 dimensional transient analytic solution to a two dimensional transient finite element solution (Akyürekli 1988).	145

List of Tables

Table 2.1 The percent of total sample activity found in the supernatant before and after stock solutions of microspheres were heated at 47°C for 5 hours in a 2M solution of KOH.	32
Table 2.2 The reproducibility in the gamma-ray counts after centrifuging samples at 6000g for 30 minutes.	32
Table 3.1. Ratio of the normothermic kidney (n=6) mean total blood flow to that of the contralateral kidney prior to the laparotomy, after mobilizing the renal vessels and after re-perfusion.	57
Table 3.2. Ratio of the hyperthermic kidney (n=6) mean total blood flow to that of the contralateral control kidney prior to the laparotomy, after mobilizing the renal vessels, after re-perfusion and during hyperthermia treatment.	57
Table 6.1 Temperature of the superficial "A" tissue samples and corresponding blood flow measured during regional microwave hyperthermia.	117
Table 6.2 Temperature of the deeper "B" tissue samples and corresponding blood flow measured during regional microwave hyperthermia.	118
Table 7.1. Thermo-physical properties of tissue/organs used in calculations of tissue temperature (Chen <i>et al.</i> 1990).	135
Table 7.2. Experimentally obtained blood flow values both baseline and during hyperthermia treatment.	135

Chapter 1

General overview

1.1 Introduction

Many solid tumors have proven to be extremely resistant to conventional treatment modalities such as chemotherapy and ionizing radiation. The failure of these modalities is thought to be related to many factors including tumor metabolism and immunological differences between and within the various types of cancer. In addition, the problems posed by cellular diversity are compounded by such factors as tumor location, vascularity, and metastatic potential. For different cancers, the limiting factor in prolonging the length of quality survival may be control of metastatic disease, while in others the critical issue is control of the primary disease.

The lack of local tumor control is often associated with treatment failure. Many modalities have been investigated as an adjuvant to ionizing radiation, including hypoxic cell sensitizers, hyperbaric oxygen, and modified fractionation schedules, all aimed at increasing tumor control rates without introducing unacceptable normal tissue damage. Many of these modalities have shown great promise in laboratory studies but have failed clinically. During the last two decades clinical investigations have suggested that hyperthermia, particularly in combination with radiotherapy, may be a promising modality in the treatment of cancer (Kim *et al.* 1982, Perez *et al.* 1983, Luk *et al.* 1984, Oleson *et al.* 1984, Scott *et al.* 1984, Arcangeli *et al.* 1985, van der Zee *et al.* 1985, Lindholm *et al.* 1987). Hyperthermia has considerable potential to act synergistically with radiation and chemotherapy, but its full potential will not be realized until tissue heating techniques become predictable

and controllable. This will require the further development of heat transfer models used to predict the tissue temperature distribution during hyperthermia.

Many physical and physiological factors combine to determine the resulting tissue temperature distribution achieved during hyperthermia treatments. These include the intrinsic specific absorption rate (SAR) pattern of the applicator, the energy absorption characteristics of different tissues as a function of the power source, the tissues' intrinsic thermal conductivity as well as blood perfusion (Jain 1980, Bowman 1982). As the principal mechanism for heat dissipation in tissue, blood flow assumes a critical role in determining not only physiologically the success of hyperthermia treatment (Gerweck 1985, Reinhold *et al.* 1985, Reinhold and Endrich 1986) but is also the physically dominant factor determining the ability to achieve satisfactory temperature distributions (Patterson and Strang 1979, Hahn 1982, Lagendijk and Mooibroek 1986). It is thus apparent that in order to plan and control hyperthermia treatment it is imperative to have an understanding of local and regional blood flow in normal and malignant tissues during hyperthermia.

The effects of hyperthermia on blood flow in human cutaneous tissues (Barcroft and Edholm 1946, Johnson *et al.* 1976, Nagasaka *et al.* 1987) and in the skin and muscle of animals (Lokshina *et al.* 1985, Milligan and Panjehpour 1985, Shrivastav *et al.* 1983, and Song *et al.* 1987a, 1987b) have been extensively studied. At the same time the vascular effects of hyperthermia on visceral organs and the central nervous system have received relatively little attention in the literature. Similarly little information is available regarding the vascular response of critical organs in large animals. Moreover, most existing data describing the effects of hyperthermia on blood flow detail only the average blood flow in a single heated volume of tissue and little is known about the heat-induced changes in the spatial distribution of blood flow.

The purpose of this study was to examine the spatial distribution of blood flow in various normal porcine tissues before, during and following a period of regional microwave hyperthermia. In order to better isolate and understand the effect of heat

on tissue blood flow, control studies were undertaken to establish the effect of anesthesia and thermocouple insertion on the muscle blood flow.

1.2 Rationale For Hyperthermia

In recent years the potential usefulness of hyperthermia as a technique to treat cancer has been demonstrated (Abe *et al.* 1986). Many studies of superficial human tumors treated with a combination of hyperthermia and radiotherapy have been reported, as reviewed by Arcangeli *et al.* (1988). The biological rationale for treating malignant tumors with hyperthermia has been provided by a great number of studies in recent years. The degree of cell killing depends on thermal dose, which is a function of temperature and time. The exact mechanism for the synergism between radiation and hyperthermia is not well understood, but several mechanisms have been suggested including the inhibition of repair (Raaphorst *et al.* 1979a, Raaphorst and Feeley 1990). Heat can also act as a direct cytotoxic agent (Raaphorst *et al.* 1979b), in particular on tumor cells which are nutritionally deprived (Dewey *et al.* 1980, Overgaard 1980, 1981, 1985, Arcangeli *et al.* 1983, van der Zee *et al.* 1986, Perez *et al.* 1986). Heat can also increase the cytotoxicity of many chemotherapeutic agents (Hahn 1979, Marmor 1979, Thuning *et al.* 1980). Since chemotherapy is used primarily against metastatic disease, the benefits of combining chemotherapy with local hyperthermia may not be as obvious. Nevertheless, in cases where lesions do not respond to chemotherapy, a combination of hyperthermia and drug therapy may have considerable potential in inducing regression in these lesions. Extensive studies of the interactions between heat and chemotherapeutic agents have been carried out by Hahn (1982).

1.2.1 Cellular Mechanisms of Heat Inactivation

Four major targets have been proposed for heat induced cellular inactivation. These have been reviewed by Raaphorst 1990, and are as follows :

(1) During DNA synthesis (S-phase), heated cells appear to die as a result of chromosomal injury. However at other times during the cell cycle (i.e., G₁, G₂ and

mitosis) no evidence of DNA strand breaks or chromosome aberrations have been observed.

(2) At times other than S-phase the primary target is thought to be a membrane, possibly the cell (or plasma) membrane or internal membranes such as the cytoskeleton. Membrane injury causes cells to die and lyse rapidly during interphase.

(3) After exposure to heat there is an increase in the number of lysosomes in the cellular cytoplasm. Disintegration of these lysosome vesicles and damage by the digestive enzymes might be the cause of cellular death.

(4) There is considerable evidence that the critical target of hyperthermia may be a membrane bound protein.

1.2.2 Environmental Factors

In some clinical cases, hyperthermia has been shown to cause cell death and subsequent tumor regression, while inflicting a lesser and clinically acceptable degree of damage to the surrounding normal tissues. It is now thought that these phenomena are partially a function of the cellular environment of normal and malignant tissues as well as the differences in the blood perfusion response between these two tissues (see section 1.3) These conditions have been reviewed by Song (1984) and Reinhold (1988) and are given below :

Intra-tumor pH

Both normal and malignant cells are known to be more sensitive to hyperthermia at low pH. It is also well established that in general solid tumors have a lower pH than the surrounding normal tissues. Additionally, at the onset of tumor heating there is an initial rise in tissue pH followed by a decrease, often significantly below the preheating value. It has been postulated that the drop in pH is a phenomenon related to thermally induced changes in blood flow. It should be noted that these changes in pH are more pronounced in some tumor tissues than in normal tissues. Several mechanisms have been proposed to account for the increase in intra-tumor pH as a function of tissue heating. These include: (1) A increase in

anaerobic glycolysis as a result of reduced oxygen availability. (2) Reduced clearance of lactate, leading to an elevation of lactic acid levels in tumors or a buildup of other metabolic by-products such as carbon dioxide.

Blood Perfusion

Blood flow determines the availability of cellular nutrients and the rate at which waste products of cellular metabolism are removed. In highly perfused tissue there are both glucose and oxygen available, and the cells primarily utilize aerobic metabolism. Under such conditions the extracellular pH remains near that of normal tissue, and cells are relatively heat resistant. If however, blood flow is reduced, then there is inadequate supply of oxygen (hypoxia) and cells utilize anaerobic metabolism pathways resulting in the production of lactic acid. In addition, with reduced blood perfusion, metabolic products are not removed at a sufficient rate and accumulate in the cellular environment. Since many of these products are acidic (i.e. lactic acid) the tissue pH is significantly lowered. As indicated above, conditions of low pH make cells thermally sensitive.

The homeostatic response to hyperthermia is vasodilation and an increased permeability of the vascular wall in normal tissues like muscle and skin. This results in an increase in blood flow to regions where high temperature gradients exist, and hence a dissipation of thermal energy. The degree of pathophysiological changes in the vascular system in normal tissue is, of course dependent on the temperature and duration of heating as well as the tissue type. The response of the vasculature in tumors and normal tissues to thermal stress is quite different. When tumors are heated to temperatures of approximately 42°C, tumor blood flow either undergoes no significant change or increases slightly. It appears that the tumor vasculature is vulnerable to heat so that vascular stasis, occlusion, and hemorrhage are prone to develop at temperatures greater than 43°C (Song 1984, Vaupel and Kallinowski 1987, Reinhold 1988). Under such conditions, tumors become very easy to heat and their cells become very heat sensitive.

Thermal damage to areas peripheral to tumors can be difficult to achieve. These regions are rather well perfused and efficient heat dissipation may be the cause of the frequent failure of hyperthermia to damage the peripheral regions of solid tumors (Song 1984, Reinhold 1988). In contrast to tumors, the blood flow in normal tissues markedly increases upon heating and thermal energy is very efficiently dissipated. It is apparent that the relative difference in blood flow between tumors and surrounding normal tissues is the principal factor determining the temperature achieved in tumors and in the surrounding normal tissues.

1.3 Objectives and experimental rationale

The primary objective of this study was to measure the vascular response of various normal porcine tissues during hyperthermia. In addition, because surgical stress can alter organ and tissue blood flow in an animal model, this study also investigated the effects of the surgical protocol and anesthesia on blood flow in an effort to better understand and isolate the vascular effects of heat. Control experiments were conducted in order to determine whether or not insertion of thermocouples into the paranchyma would alter or modify the vascular response of the kidney during hyperthermia. Animals used to examine the effect of thermocouple insertion were prepared in the identical manner as those in the hyperthermic group, but underwent sham heating. In all cases, the native contralateral organ or tissue was used as an internal control, this provided a means of determining whether the changes in blood flow before, during and after hyperthermia were the outcome of systemic or local factors. Figure 1.1 shows a flow chart illustrating the experiments conducted in this study.

As part of this study, three independent techniques of measuring blood flow were employed. The principal method, the radioactive tracer microsphere technique, was used to measure total blood flow as well as blood flow distribution. Calibrated electromagnetic and ultrasound transit time flow probes were used to corroborate the microsphere data, which is important for scientific validation of the measurements. Furthermore, as part of our investigation we developed a tissue

preparation technique that produces tissue samples with the consistent, reproducible counting geometry necessary when planar configured Ge(Li) detectors are used to make multiple microsphere blood flow measurements.

Hyperthermia may improve the radiation response of certain abdominal tumors, such as ovarian carcinoma. Heating abdominal tumors may also inadvertently induce renal hyperthermia, therefore in order to optimally use hyperthermia and avoid complications it is important to quantify the renal vascular response. At present there is a lack of information regarding the effect of hyperthermia on kidney blood flow. As part of our investigation of the vascular response of the kidney to heat, we have measured the changes in renal blood flow in an autoperfused porcine kidney before, during and after a period of hot water immersion heating. The model was developed and used to investigate the denervated vascular response of the kidney to heat. An *in-situ* hyperthermic kidney model was developed in order to examine the vascular response of an enervated kidney to heat and as a comparison to the response of the denervated autoperfused kidney.

During the 1980's several institutions reported results of Phase I studies of interstitial irradiation in the treatment of malignant gliomas. The usefulness of this technique resides in the ability to boost the bulky portion of the tumor where the majority of the recurrences develop after external-beam radiotherapy. However, even interstitial implants, local recurrence continues to be the most common mode of failure. Hyperthermia as an adjuvant to interstitial radiotherapy has generated significant interest in the neuro-oncology community because of its potential synergistic effect with radiation. Therefore, it is important to understand the vascular response of the brain to interstitial heating. In his study we measured the changes in porcine brain blood flow before, during and after a period of interstitial hyperthermia.

The effects of hyperthermia on local and regional blood flow in the muscle of animals (Lokshina *et al.* 1985, Milligan and Panjehpour 1985, Shrivastav *et al.* 1983, Song *et al.* 1987a, 1987b) have been extensively investigated. However, most existing data describing the effects of hyperthermia on muscle blood flow

detail only the average blood flow in a single heated volume of tissue and little is known about the heat-induced changes in the spatial distribution of blood flow. In these experiments we examined the spatial distribution of blood flow in normal porcine muscle before, during and following a period of regional microwave hyperthermia. In order to isolate the effect of heat on tissue blood flow, control studies were undertaken to establish the effect of anesthesia and thermocouple insertion on the muscle blood flow.

Thermal models of tissue play an important part in hyperthermia cancer therapy, in particular for computerized simulations of hyperthermic treatments. In this study a simple model, namely, the one dimensional bioheat transfer equation, was used to examine the effects of blood perfusion and surface cooling on the tissue temperature distribution. The calculated tissue temperature history was compared to an *in-vivo* muscle temperature measurement made under hyperthermic conditions generated by a 915 MHz microwave applicator. Blood flow parameters used in the model were obtained experimentally. The results from the simple bioheat transfer equation were then compared to the temperature pattern seen experimentally.

1.4 Materials and methods

1.4.1 Animals

After obtaining protocol approval from the National Research Council of Canada Animal Care and Use Committee, acute experiments were conducted on 140 healthy pigs of either sex and of mixed breeding, weighing between 20 and 65 kg with a mean weight of 40 kg. The tissues and organs investigated in this study included the kidney muscle and brain.

1.4.2 Surgical preparation

Following premedication with 0.5-1.0 mg/kg of diazepam I.M. and 0.05 mg/kg of atropine sulfate I.M., surgical anesthesia was induced with a single dose of ketamine, 20-30 mg/kg I.M. While in dorsal recumbence a tracheotomy was

performed and a cuffed endotracheal tube placed in the trachea. The animal was mechanically ventilated (Ohio Respirator) and maintained under stable anesthesia for 7-8 hours with a 1-1.5% concentration of halothane in a 50/50 mixture of O₂/N₂. A continuous infusion of isotonic normal saline was administered intravenously and the esophageal temperature (midthoracic), left ventricle pressure, pH_a, arterial carbon-dioxide tension and hemoglobin-oxygen saturation were all monitored and well maintained throughout the experiment.

1.4.3 Blood flow measurements

The microsphere reference sample technique offers a convenient means of sequentially assessing both the local and regional distribution of blood flow to virtually all organs under a variety of hemodynamic conditions. Microspheres are discrete sources of radioactivity with a density slightly greater than that of red blood cells and a size ranging from 10-30µm in diameter. The fundamental assumption underlying the technique is as follows: Microspheres infused as a bolus into the left ventricle of the heart are homogeneously mixed and distributed throughout the systemic circulation. The microspheres embolize in the terminal vascular bed of tissue in proportion to blood flow, hence, the tissue radioactivity is proportional to tissue blood flow (Heymann *et al.* 1977).

In order to quantify the rate of blood flow a reference sample is withdrawn at a fixed known rate from an artery distal to the heart during infusion of the microspheres. The reference sample provides a quantification of the radioactivity for a fixed and known flow rate. At the conclusion of the experiment samples of tissue are dissected from the region of interest and their activity along with that of the reference sample quantified. Since the γ-ray spectrum of each isotope is individually resolvable, changes in blood flow can sequentially be measured by using microspheres with different isotopic labels. After measuring the activity of the tissue and reference samples blood flow to each tissue sample can then be calculated using equation 1.1. In this study microspheres (Dupont NEN-TRAC), 15 µm in diameter, and labeled with either ⁵⁷Co, ¹¹³Sn, ⁸⁵Sr, ⁴⁶Sc, ¹¹⁴In or ¹⁵³Gd were

used to measure blood flow at each of six distinct stages during the experiment. The detailed procedure for the measurement of blood flow with radioactive microspheres is reported in chapter 2.

Under general anesthesia with halothane, the left ventricle was catheterized by way of the right common carotid artery using a catheter sheath introducer (Cordis) and a number 8F pigtail catheter (Cordis). The catheter served as a site for the infusion of microspheres and for monitoring left ventricle pressure. The left femoral artery was cannulated by way of a superficial inguinal branch using a 16 gauge, 30 cm Dannula cut-down catheter (Medinique International Inc.) advanced superiorly towards the abdominal aorta. This catheter served to monitor the systemic arterial pressure and as a site for obtaining arterial microsphere blood reference samples. Similarly, the left femoral vein was cannulated using an 18 gauge, 30 cm Dannula cut-down catheter advanced superiorly towards the abdominal vena cava and served to monitor the central venous pressure. The animals' core temperature was maintained between 38-39°C.

Each microsphere infusion contained approximately 1.2×10^6 microspheres suspended in 10 ml of isotonic normal saline with 0.01% polysorbate (Tween-80). The microspheres were vortexed for 5 minutes, then infused into the left ventricle at a rate of 1ml/sec over 10 seconds followed by a 10 ml flush with warm isotonic saline. Blood reference sampling started 5 seconds before microsphere infusion and was collected over a period of 1 minute at a rate of 24.7 ml/min from the abdominal aorta using a constant flow withdrawal pump (Harvard Apparatus Inc.). Blood reference samples were then transferred to 50 ml heparinized centrifuge tubes (Corning).

Animals were euthanized by inducing cardiac arrest with an injection of 30 ml of a saturated KCl solution into the left ventricle. Each tissue sample was weighed and identified with respect to the thermocouple positions within the sample. The tissue and blood reference samples were prepared for gamma-ray counting and their activity measured as described in chapter 2.

1.4.4 Measurement of γ -ray spectra

In order to determine tissue blood flow the radioactivity of each tissue and blood reference sample must be determined. When blood flow measurements are made using multiple microsphere isotope labels a complex gamma ray spectrum is generated. In this study gamma-ray spectra were measured with a Ortec (Model 8III-15200-S) cylindrical Ge(Li) detector with a crystal diameter of 46.6 mm and a length of 57.0 mm. Pulses from the detector were amplified and shaped by a Ortec model 572 spectroscopy amplifier and routed through a Canberra model 8076 analog-to-digital converter to a Canberra Series 90 multichannel analyzer with 2046 channel resolution. The detector was calibrated using six isotope point sources (Laboratoire de Métrologie des Rayonnements Ionisants, France) covering the energy range from 122 keV to 1120 keV and located 3.00 ± 0.05 cm from the detector face. Figure 1.2 shows an example of multiple isotope gamma ray spectrum measured using the Ge(Li) detector.

The area under each isotope photopeak was integrated in order to determine the activity. The isotope photopeak activities were calculated using a data analysis program (*Spectrum Analysis Program*) provided by the Canadian National Research Council, Institute of Radiation Standards and run on a Digital PDP 11/44 computer. The program fits spectral data using the Levenberg-Marquardt method of non-linear least squares such that for large photopeaks the Gaussian parameters are established and the residual background counts subtracted from the spectrum. Hence, in this manner the activity of each individual isotope can be determined.

During counting samples were held in position with respect to the detector face by a precision sample holder, mechanically connected to the detector assembly (see chapter 2). In order to minimize photon scatter, the sample holder was constructed from Lucite. The end cap of the detector was surrounded by a four inch lead castle reducing the background radiation count rate to less than 0.03 cpm. The counting vials were positioned such that the vertex of the vials were at a distance $d = 3.00 \pm 0.05$ cm from the detector face.

1.4.5 Calculation of blood flow

The blood flow (ml/min/100g) to each tissue sample was calculated using the following equation :

$$Q_b = 100Q_R \frac{I_i}{I_a w_i} \quad (1.1)$$

where Q_b is the blood flow in tissue section i , Q_R the reference sample withdrawal rate, I_i the tissue activity, I_a arterial blood reference sample activity and w_i , the weight of the tissue section. Total organ blood flow Q_b , is then given by :

$$Q_b = \sum_{i=1}^N 100Q_R \frac{I_i}{I_a w_i} \quad (1.2)$$

where the sum is over the total number of tissue sections N , into which the organ was dissected.

1.4.6 Electromagnetic flow probe

The most commonly used technique for the measurement of arterial or venous blood flow during surgery is the electromagnetic flow probe. The operation of the flow probe is based on Faraday's law of electromagnetic induction

$$E = Blv \times 10^{-8}$$

where E is the electromotive force (volts), B is the magnetic field (gauss), l is the lumen diameter (cm) and v is the velocity of blood (cm/s). The law states that when a conductor (blood) moves through a magnetic field, an electromotive force is generated in a direction mutually perpendicular to both the magnetic field and the direction of motion. As blood moves through the magnetic field at a given velocity, a voltage (emf) proportional to the velocity is produced across two electrodes held rigidly in the lumen at an angle of 180° from each other (For further information consult Weber 1991).

1.4.7 Ultrasonic flow meter

Another commonly used technique for the measurement of arterial or venous blood flow during surgery is the transit time ultrasonic volume flow meter (Transonics Systems Inc.). The transit time flow meter has certain advantages over other blood flow meters such as the electromagnetic flow meter in that the detected signal gives a direct measurement of the net volume flow through the blood vessel. The measurement of blood flow by this technique is based on the principal that the transit time of a sound waves traveling between two points is dependent on the velocity of sound in blood plus the velocity of the blood in the direction of the propagating wave. The difference in the transit times between the upstream and down stream ultrasonic pulses can be related to the flow rate of the medium (For further information consult Weber 1991).

1.4.8 Hyperthermia system

Hyperthermia was induced by means of a single direct contact 915 MHz dielectrically loaded rectangular wave guide. The penetration or skin depth of the 915 MHz wave guide is 3 cm, hence at this distance the power density decreases to $1/e$ (37%) of its value at the surface. The applicator was constructed from WR 430 wave guide (internal dimensions 10.92×5.46 cm) with the feed structure based on an arrangement described by Guy (1978). Since the WR 430 wave guide is too small to propagate energy at 915 MHz, the applicator was filled with alumina powder and a 0.6 cm perspex window at the aperture of the applicator retained the powder. Longitudinal axis power deposition profile of the 915 MHz microwave applicator is shown in figure 1.3.

In order to improve the applicator-tissue coupling and to ensure effective surface cooling a temperature controlled de-ionized water bolus was attached to the aperture of the applicator. The bolus consisted of a water filled, 12×6 cm, thin rectangular silicone membrane with in-flow and out-flow connectors at the edges. De-ionized water was pumped through the bolus from a constant temperature water

bath. Forward and reflected power was monitored with a digital power meter (Bird Model 4381).

Interstitial heating was accomplished by means of a single 915 MHz interstitial helical coil microwave antenna inserted within a blunt close ended polypropylene catheter (Sovereign, 2.3 mm O.D., 1.9 mm I.D.). The design and SAR distribution of the helical antenna, has been described previously (Sato and Stauffer 1988). The antenna consists of an eight turn helical coil, 2.5 cm in length shorted to the antenna feed line at the tip with an outer diameter of 1.2 mm.

The thermometry system has been characterized and described in detail in the literature (Gerig *et al.* 1988, 1992). Four type K (Chromel-Alumel) thermocouple arrays in an extruded nylon sheath with an outer diameter of 0.81 mm and a wall thickness of 0.08 mm were used to measure temperature throughout the heating period. Each array consisted of 7 thermocouple junctions with a 10.0 ± 0.5 mm separation. Thermometry data was collected with a computer controlled data acquisition system (Keithley DAS 500) at a sampling rate of 3 kHz. Each reported temperature measurement was the average of 100 samples.

Figure 1.1 Flow chart illustrating the heating and control experiments.

Figure 1.2 This spectrum is derived from a mixture of routinely used radiolabelled microspheres. A Ge(Li) detector was used to obtain the spectrum.

Figure 1.3 Longitudinal axis power deposition profile of the 915 MHz microwave applicator.

Chapter 2

A Tissue Preparation Technique For Microsphere Assays Of Blood Flow

2.1 Introduction

The validity of the reference sample technique of blood flow measurement using radiolabeled microspheres has been well established in a variety of experimental models (Buckburg *et al.* 1971, Archie *et al.* 1973, Heymann *et al.* 1977, Dole *et al.* 1982, Bassingthwaite *et al.* 1987, Austin *et al.* 1989). Initially described by Makowski *et al.* 1968, the technique offers a convenient means of sequentially assessing both the local and regional distribution of blood flow to virtually all organs under a variety of hemodynamic conditions.

The fundamental assumption underlying the technique is that sphere embolization and therefore tissue sample radioactivity is proportional to blood flow (Heymann *et al.* 1977). However, systematic and random errors make this assumption only an approximation (Dole *et al.* 1982, Austin *et al.* 1989). When quantifying the radioactivity of tissue samples, systematic errors resulting from variation in the source-detector distance and the source distribution, contribute significantly to the total error of the microsphere measurements (Austin *et al.* 1989, Katz *et al.* 1972), especially when planar configured semiconductor detectors are utilized (Tsouflanidis 1983).

Microspheres are discrete sources of radioactivity with a density slightly greater than that of red blood cells (Goldman 1993). In liquid media such as whole blood, they settle to the bottom of the counting vial, whereas in solid tissues they are dispersed throughout the tissue in the counting vial (Heymann *et al.* 1977). If

meaningful comparisons of the gamma-ray activity from different types and configurations of tissue samples are to be made, the comparative counting efficiency relationships must be determined for each nuclide in each geometric configuration (Heymann *et al.* 1977, Mernagh *et al.* 1976, Debertin and Helmer 1988). Alternatively, the need for such relationships may be eliminated by preparing all tissue samples in a standardized manner which ensures a consistent counting geometry.

Nose *et al.* (1985) solubilized myocardial tissue samples in 10% NaOH and allowed the microspheres to precipitate to the bottom of the counting vial, such that tissue and blood samples would be in the same physical form for counting. However, a consistent microsphere counting geometry is not assured, since the microspheres will randomly distribute over the bottom of the counting vial. By solubilizing and then centrifuging tissue samples in conical shaped counting vials the microspheres are concentrated at the vertex of the vial, thereby obtaining tissue and blood reference samples with the same well defined, reproducible counting geometry.

In this work we report on a tissue preparation technique that produces samples with the consistent, reproducible counting geometry necessary when planar configured Ge(Li) detectors are used to make multiple microsphere blood flow measurements. Results of measurements made to verify the validity and reproducibility of this technique are also reported.

2.2 Materials and Methods

2.2.1 Animal preparation for microsphere measurements

After obtaining protocol approval from the National Research Council of Canada Animal Care and Use Committee, studies were conducted on 13 healthy pigs of either sex and of mixed breeding, weighing between 40 and 65 kg (mean $\pm$ SEM = 52 ± 2 kg). Following premedication with 0.5-1.0 mg/kg of diazepam i.m., surgical anesthesia was induced with a single dose of ketamine, 20-30 mg/kg i.m.. While in dorsal recumbence a tracheotomy was performed and a cuffed endotracheal tube

placed in the trachea. The animal was then mechanically ventilated (Ohio Respirator) and maintained under stable anesthesia for 7-8 hours with a 1-1.5% concentration of halothane to a 50/50 mixture of O₂/N₂.

Catheterization of the left ventricle was by way of the right common carotid artery using a number 8F pigtail catheter (Cordis). The left femoral artery was cannulated using a 12F catheter (Cordis) advanced towards the abdominal aorta. A 14F cannula, introduced through the left femoral vein and advanced towards the vena cava, served to monitor the central venous pressure (CVP). Isotonic normal saline was administered by way of the lateral auricular vein, while the core temperature was maintained between 37-38°C. Mean arterial blood pressure and CVP were continuously monitored using pressure transducers (Gould Statham, Model P23DB) and recorded on a multichannel strip chart recorder (Gould). Arterial PO₂, PCO₂, pH and base excess were measured using a blood gas analyzer (Instrumentation Laboratory IL System 1304).

A laparotomy was performed to expose the left (or right) kidney. Total blood flow to the kidney was then measured using a calibrated Transonics blood flow meter (Model T101) with a 4 mm R-series perivascular flow probe (Transonics Systems Inc.) placed over the renal artery. After a one to two hour stabilization period, sequential measurements of the kidney blood flow were made using 15µm microspheres (Dupont NEN-TRAC) labeled with ⁵⁷Co, ¹¹⁴In, ¹¹³Sn, ¹⁰³Ru, ⁸⁵Sr and ⁴⁶Sc.

Each aliquot, contained approximately 1.2×10^6 microspheres suspended in 10ml of warm isotonic normal saline with 0.01% Tween-80. The microspheres were Vortexed for 5 minutes, then infused into the left ventricle at a rate of 1ml/s over 10 seconds and then flushed at the same rate with 10 ml of warm isotonic normal saline. Blood reference sampling started 5 seconds before microsphere infusion and was collected over a period of 1 minute at a rate of 24.7 ml/min from the abdominal aorta using a withdrawal pump (Harvard Instruments). Blood reference samples were then transferred to heparinized 50 ml conical polypropylene centrifuge tubes (Corning).

At the conclusion of each experiment the animals were sacrificed by inducing cardiac arrest with an injection of approximately 30 ml of saturated KCl solution into the left ventricle, after which both kidneys were retrieved for gamma ray counting.

2.2.2 Calculation of blood flow

The blood flow (ml/min/100g) to each tissue sample was calculated using the following equation :

$$Q_i = 100Q_R \frac{I_i}{I_a w_i} \quad (2.1)$$

where Q_i is the blood flow in tissue section i , Q_R the reference sample withdrawal rate, I_i the tissue activity, I_a arterial blood reference sample activity and w_i the weight of the tissue section. Total organ blood flow Q_b , is then given by :

$$Q_b = \sum_{i=1}^N 100Q_R \frac{I_i}{I_a w_i} \quad (2.2)$$

where the sum is over the total number of tissue sections N , into which the retrieved tissue was dissected.

2.2.3 Preparation of the counting samples

In order to determine the spatial distribution of blood flow, as well as total organ flow, kidneys retrieved from the study were sliced transversely into 1 cm sections and further subdivided into posterior and anterior halves. Each tissue section was placed into pre-weighed 50 ml polypropylene centrifuge vials (Corning) labeled with the tissue section's anatomical site of origin, re-weighed, then minced inside the vial.

Blood samples were hemolyzed and tissue sections dissolved by adding a sufficient amount of 2M KOH (approximately 20 ml) to each sample vial, such that

the level of fluid in all the vials were identical. After approximately 5 hours at 47°C in a constant temperature agitating water bath, the tissue samples were fully dissolved. The samples were then centrifuged at 3000g for 5 minutes in a swinging bucket rotor in order to concentrate the microspheres at the vertex of the counting vial. To prevent migration of the microspheres within the vial, they were frozen using crushed dry ice.

2.2.4 Microsphere leaching and dispersion

Inherent to our technique of tissue preparation is the assumption that centrifuging a solubilized tissue sample concentrates all of the microspheres and therefore, all of the isotope at the vertex of the counting vial. However, isotope leaching and convective motion of the supernatant can alter this distribution and may make the assumption invalid. In order to assess the validity of this assumption, the degree to which each microsphere label could be concentrated at the vertex of the counting vial was determined.

Control microsphere samples were prepared for each nuclide label by suspending approximately 5×10^5 spheres in 30 ml of isotonic normal saline and 0.01% Tween-80 at 4°C. The samples were Vortex mixed for 5 minutes then centrifuged at 3000g for 5 minutes, after which 29 ml of the supernatant from each sample was transferred to a separate vial. The supernatants were Vortex mixed for 2 minutes, then 1 ml samples withdrawn and transferred to separate vials for γ -ray counting. The activity of the supernatant from the control sample I_{sc} , was determined by multiplying the activity of the 1 ml sample by the total volume of the supernatant. The activity of the precipitate I_{pc} , was then measured and the fraction I_{sc}/I_{pc} , calculated for each isotope label.

In order to determine the effect of the solubilization process on the isotope distribution within the counting vial, 29 ml of 2M KOH was added to each of the remaining 1 ml solutions containing the microsphere precipitates. These were then mixed and heated at 47°C in an agitating constant temperature bath for a period of 5 hours in order to simulate the tissue preparation procedure. The vials were then

allowed to cool to room temperature before being centrifuged at 3000g for 5 minutes. Subsequently, 29 ml of supernatant from each processed sample was withdrawn and transferred to separate centrifuge vials and their activity I_{ss} , determined as described above and the ratio I_{ss}/I_{pc} , calculated for each isotope label.

2.2.5 Geometric reproducibility of centrifuged samples

In order to assess the accuracy with which samples are positioned for counting, approximately 5×10^4 microspheres suspended in 30 ml of isotonic normal saline and 0.01% Tween-80 were vortex mixed for 5 minutes then centrifuged at 3000g for 5 minutes and placed in the sample holder (see Fig. 2.1) for γ -ray counting. This process was repeated 10 times for each isotope and results compared. The assessment was also repeated for an arterial blood reference sample and a kidney tissue sample containing all six isotope labels.

2.2.6 Validation of the tissue preparation technique

The validity and accuracy of our tissue preparation and counting technique was assessed by comparing measurements of total renal blood flow made using a calibrated Transonic flow probe (Weber 1991) to values simultaneously obtained by microspheres using our tissue preparation technique.

2.2.7 Measurement of γ -ray spectra

Gamma-ray spectra were measured with a Ortec (Model 8III-15200-S) cylindrical Ge(Li) detector with a crystal diameter of 46.6 mm and a length of 57.0 mm. Pulses from the detector were amplified and shaped by a Ortec model 572 spectroscopy amplifier and routed through a Canberra model 8076 analog-to-digital converter to a Canberra Series 90 multichannel analyzer with 2046 channel resolution. The detector was calibrated using six isotope point sources (Laboratoire de Métrologie des Rayonnements Ionisants, France) covering the energy range from 122 keV to 1120 keV and located 3.00 ± 0.05 cm from the detector face.

The isotope photopeak activities were calculated using a data analysis program (Spectrum Analysis Program) provided by the Canadian National Research

Council, Institute of Radiation Standards and run on a Digital PDP 11/44 computer. The program fits spectral data using the Levenberg-Marquardt method of non-linear least squares. Samples were held in position with respect to the detector face by a precision sample holder, mechanically connected to the detector assembly as illustrated in Fig. 2.1. In order to minimize photon scatter, the sample holder was constructed from Lucite. The end cap of the detector was surrounded by a four inch lead castle reducing the background radiation count rate to less than 0.03 cpm. The counting vials were positioned such that the vertex of the vials were at a distance $d=3.00 \pm 0.05$ cm from the detector face.

2.3 Results

The fraction of sample radioactivity I_{ss}/I_{pc} , found in the supernatant of microsphere samples having undergone the tissue preparation process are shown in Table 2.1. As can be seen the ratio appears to be isotope specific. For all the isotope labels tested the ratio I_{ss}/I_{pc} increased by a factor ranging from 0.5 to 20 above control values I_{sc}/I_{pc} which had an average value of just under 0.2%. The maximum increase in the fraction I_{ss}/I_{pc} after exposure to the tissue preparation process was seen in the ^{85}Sr sample and was found to be 1.58%.

In order to determine the counting accuracy of our detection system, the activity of samples having undergone the tissue preparation process were repeatedly measured. The results of these measurements are shown in table 2.2 and represent the percent standard error of the mean for $n=10$. Control microspheres samples have a reproducibility with a standard error of less than 2% at the 68% confidence level. Similarly, the arterial blood reference sample had a reproducibility of 2% while that of the tissue sample was approximately 4%.

The overall accuracy of the technique as determined independently by a calibrated Transonic flow probe is shown in Fig. 2.2. In this case, *in-situ* blood flow to a porcine kidney was measured simultaneously by the two techniques. Results of these two independent methods of flow measurement ($n=13$) are in good agreement, with a correlation coefficient of $R=0.991$.

2.4 Discussion

The fundamental assumption underlying the microsphere technique is that sphere entrapment and therefore tissue sample radioactivity is proportional to blood flow (Heymann *et al.* 1977). However, two inherent classes of error, stochastic and methodological make this assumption only an approximation (Dole *et al.* 1982, Austin *et al.* 1989). The embolization and decay of radionuclide labeled microspheres are both stochastic processes described by a Poisson distribution. In this study random errors were reduced to within 5% of the mean distribution at the 95% confidence level by infusing 1.2×10^6 microspheres to each animal for each flow measurement. This resulted in approximately 1×10^3 microspheres per gram of kidney per nuclide label, and approximately 1×10^4 microspheres in the arterial blood reference samples.

The discrete nature of microspheres can pose special problems in quantifying their radioactivity, particularly when using planar configured semiconductor detectors. In liquid media, such as blood, microspheres settle to the bottom of the counting vial, whereas in whole tissues they are dispersed randomly throughout (Heymann *et al.* 1977). Methodological errors introduced in quantifying the radioactivity of samples can contribute significantly to the total error of microsphere measurements (Katz and Blantz 1972).

The sources of detector error which are of importance include counting geometry, coincidence summing, and nuclide separation (Austin *et al.* 1989). Since typical Ge(Li) detectors have energy resolutions on the order of 0.1% at FWHM and Peak-to-Compton ratios near 50:1, the separation of multiple nuclide photopeaks can be achieved without resorting to complex serial stripping techniques. Furthermore, planar configured detectors are much less sensitive to coincidence summing than are well type detectors, mainly because of their smaller solid angle of acceptance (Debertin and Helmer 1988). However, special consideration must be given to the counting geometry, since the counting efficiency of planar configured semiconductor detectors are markedly sensitive to variations in

the source-detector distance and the source distribution (Tsouflanis 1983, Debertin and Helmer 1988).

The relation between the number of counts $N(E)$, under a γ -ray photo peak of energy E and the counting geometry is given by the equation

$$N(E) = \frac{A}{4\pi} S(E) \Omega \epsilon_i(E, \Omega) \quad (2.3)$$

where Ω is the solid angle (in steradians) subtended by the detector at the position of the source, $\epsilon_i(E, \Omega)$ the detector intrinsic peak efficiency, $S(E)$ the number of γ -rays emitted by the source during the counting period and A is a combination of correction factors that may be applied to the result (Tsouflanis 1983).

To illustrate the effects of source geometry consider the simplest case, namely a point source emitting γ -rays isotropically and positioned at a distance d from a circular detector of radius R_D and zero thickness. The uncertainty in $S(E)$, is given as

$$\frac{\sigma_s}{S} = \sqrt{\left(\frac{1}{\sqrt{N}}\right)^2 + \left(\frac{\sigma_\Omega}{\Omega}\right)^2} \quad (2.4)$$

where it is assumed that $\sigma_N = \sqrt{N}$, $\sigma_{\epsilon_i} = 0$ and $\sigma_A = 0$ and where the solid angle subtended by the detector at the position of the source is given as

$$\Omega = 2\pi \left(1 - \frac{d}{(d^2 - R_D^2)^{1/2}} \right) \quad (2.5)$$

The error in solid angle that results if there is a uncertainty in position of σ_d , is as follows

$$\frac{\sigma_{\Omega}}{\Omega} = \left\{ \left[1 + \frac{d^2}{R_D^2} \right] \left[\sqrt{1 + \frac{d^2}{R_D^2}} - \frac{d}{R_D} \right] \right\}^{-1} \frac{\sigma_D}{R_D} \quad (2.6)$$

The variation in the standard error σ_s/S is illustrated in Fig. 2.3 as a function of d for various count rates in the particular case $R_D = 2.33\text{cm}$, $\sigma_d = 0.05\text{cm}$, $\epsilon_s(E, \Omega) = 0.06$ and for a count time $t = 10$ minutes. As can be seen the source-detector distance which minimizes σ_s/S is primarily dependent on the source activity, however, other aspects must also be considered (Debertin and Helmer 1988).

After centrifuging a solubilized tissue sample, the microspheres may be assumed to form a thin circular disc shaped source at the bottom of the vial (see Fig. 2.1). The average solid angle fraction subtended by a circular detector of radius $R_D = 2.33\text{cm}$ at a distance $d = 3.0\text{ cm}$ from a disc source of diameter $R_S = 0.35 \pm 0.05\text{ cm}$ coaxial with the detector has been calculated by Gardner *et al.* (1980) to be 0.0970 and is approximately equal to that for a point source (0.1051) at the same distance. Hence, we may assume that centrifuged solubilized tissue samples are approximately "point like" sources for $d \geq 3\text{ cm}$ and positioning them between 3-4 cm from the detector face minimizes the overall uncertainty in $S(E)$ when the count rates are between 500-1000 cpm (see Fig. 2.3).

Comparing the supernatants withdrawn from microsphere samples exposed to the solubilization process to those not exposed (Table 2.1) indicates that the solubilization process does influence the distribution of the isotopes within the counting vials, however the increase in activity of the supernatant is not significant. The fraction of total sample activity found in the supernatant of microspheres suspended in isotonic normal saline was within the manufactures' specifications of $< 0.5\%$. It is clear that in centrifuging solubilized bulk tissue samples, the microspheres, and therefore the radiolabels are concentrated at the vertex of the counting vials.

This is supported by the following observations: (1) the amount of radioactivity found in the supernatant of all the isotope labels having undergone the solubilization process was less than 1%, except for ^{85}Sr which had an activity of $1.58 \pm 0.01\%$, and (2) microspheres suspended in a liquid with a density of up to 1.39 g/cm^3 will have a drift velocity of 0.3 cm/sec when the centrifugal force is $3000g$. Hence, when the liquid in the counting vials are at a height of approximately 6.5 cm , all the microspheres should sediment to the bottom of the counting vial in approximately 22 seconds.

The reproducibility in positioning microsphere samples suspended in isotonic normal saline and 0.01% Tween-80, reported as the percent standard error on the mean for $n=10$ was 2%. For arterial blood reference samples and tissue samples, the reproducibility was slightly $> 2\%$ of the mean and slightly $> 4\%$ of the mean, respectively. The reproducibility may be less for tissue samples because fibrous tissues remain in the samples after 5 hours at 47°C in 2M KOH . This prevents uniform packing of the microspheres at the bottom of the vial. Our results establish the reproducibility in centrifuging microsphere tissue samples to within 4% of the mean at the 68% confidence level.

Validation of our technique against a calibrated Transonic flow probe showed good correlation between the two methods with a correlation coefficient of 0.99. Using our technique of tissue preparation with a plane cylindrical Ge(Li) detector, we found that reproducibility of sample geometry was within 4% of the mean at the 68% confidence level. We would expect this to be reduced by the use of well type counters which have less sensitivity to variations in source geometry. The overall validity of the technique is supported by agreement with independent flow measurements.

Figure 2.1. Illustration of the source-detector arrangement showing an enlarged view of the counting vial. The reproducibility of the source position is assured by mechanically connecting the sample holder to the detector housing.

Figure 2.2. Comparison of the total kidney blood flow as measured by the reference sample microsphere technique and a Transonic flow probe.

Figure 2.3. Variation in the standard error σ_s/S for an isotropic point source, as a function of detector to source distance d (plotted from Eq. 2.4) for various count rates in the particular case $R_D = 2.33\text{cm}$, $\sigma_d = 0.05\text{cm}$, $\varepsilon_i(E, \Omega) = 0.06$ and for a count time of $t = 10$ minutes.

Isotope	% Leaching Before	% Leaching After
⁵⁷ Co	0.28 ± 0.02	0.79 ± 0.01
¹¹⁴ In	0.26 ± 0.06	0.33 ± 0.07
¹¹³ Sn	0.12 ± 0.04	0.16 ± 0.04
¹⁰³ Ru	0.39 ± 0.09	0.44 ± 0.13
⁸⁵ Sr	0.07 ± 0.02	1.58 ± 0.01
⁴⁶ Sc	0.10 ± 0.01	0.28 ± 0.01

Table 2.1. The percent of total sample activity found in the supernatant before and after stock solutions of microspheres were heated at 47°C for 5 hours in a 2M solution of KOH.

Nuclide Label	Energy KeV	Arterial Blood Reference sample	Microsphere Sample	Tissue Sample
⁵⁷ Co	122	1.3	1.5	4.2
¹¹⁴ In	192	2.2	1.8	3.8
¹¹³ Sn	392	2.2	1.5	3.3
¹⁰³ Ru	497	1.6	1.7	4.0
⁸⁵ Sr	514	2.0	1.5	3.7
⁴⁶ Sc	889	2.1	1.0	4.3

Table 2.2. The reproducibility in the gamma-ray counts after centrifuging samples at 6000g for 30 minutes. Results are represented as the percent standard error on the mean for n=10.

Chapter 3

Changes in blood flow in a hyperthermic autoperfused kidney

3.1 Introduction

The vascular effects of hyperthermia on visceral organs have received relatively little attention in the literature. In particular there is a lack of information regarding the effect of hyperthermia on renal blood flow. Although hyperthermia may improve the radiation response of certain abdominal tumors such as ovarian carcinoma, hyperthermia treatment may inadvertently heat the kidney. In order to avoid complications and for the effective use adjuvant hyperthermia it is important to quantify the renal vascular response to heat.

The isolated, autoperfused kidney is a suitable model for studying the vascular response of the kidney to hyperthermia under control conditions. Extracorporeal perfusion and heating of the kidney provides a means of thermally isolating and regulating the kidney temperature. By immersion heating of the kidney, it is possible to obtain a uniform temperature distribution. Furthermore, a better understanding of the vascular response of the kidney to hyperthermia can be attained by first examining the denervated kidney, since the response during heating will primarily be mediated by myogenic and systemic factors.

As part of our investigation of the vascular response of the kidney to hyperthermia, we have measured the changes in renal blood flow in an autoperfused porcine kidney before, during and after a period of hot water immersion heating. In order to isolate the effects of heat on tissue blood flow,

control experiments were undertaken to establish the vascular changes occurring as a consequence of the surgical protocol.

3.2 Materials And Methods

3.2.1 Animals

Acute experiments were performed on 19 healthy pigs of either sex and of mixed breeding, ranging in weight from 27 to 66 kg with a mean weight of 42 ± 4 kg (Mean $\pm$ S.E.M). Six animals were used to investigate the effect of hyperthermia on renal blood flow in a heated autoperfused kidney. A further six animals were used to examine blood flow in the autoperfused kidney under conditions of normothermia (38°C). The remaining seven animals were used to investigate the effect of prolonged anesthesia on renal blood flow.

3.2.2 Hyperthermia system

The autoperfused kidney selected for hyperthermia treatment was heated for 60 minutes by immersion in an isotonic saline bath, the temperature of which was maintained by a thermally regulated circulating bath (Haake Model F3-C) as shown in figure 3.1. Temperature throughout the bath was uniform ($\pm 0.1^{\circ}\text{C}$) and remained relatively stable throughout the course of the experiment. Arterial blood perfusing the isolated kidney was heated using a heat exchanger regulated by a second constant temperature circulating water bath (Haake Model F3-C), also shown in figure 3.1. The temperature of the autoperfused kidney was monitored throughout the heating period by type K (Chromel-Alumel) thermocouples placed within the catheters used to cannulate the renal artery and vein of the autoperfused kidney. The temperature of saline bath and the perfusing arterial blood were both elevated at a rate of approximately $0.5^{\circ}\text{C}/\text{min}$ to 43°C . At this rate of heating the temperature of the out-flowing venous blood was approximately equal ($\pm 0.2^{\circ}\text{C}$) to that of the perfusing arterial blood.

3.2.3 Kidney retrieval and re-perfusion

Kidney retrieval was carried out in median laparotomy. The right colon and small intestine were reflected medially and insulated with warm saline soaked towels. The kidney selected for hyperthermia treatment was mobilized, and after isolating the renal vessels, a 4 mm perivascular ultrasonic flow probe (Transonic Systems Inc. Ithaca NY) was placed over the artery and connected to a flow meter (Transonic Systems Inc. Ithaca NY, Model TI01) calibrated using the gravity-fed constant flow technique (Weber 1991). The flow probe provided continual monitoring of renal flow before the nephrectomy and was used to corroborate the microsphere measurements. Prior to transecting the renal vessels, the animal was administered a 250 unit/kg I.V. loading dose of heparin. Following the nephrectomy the kidney was placed in a cold saline bath at approximately 4°C and perfused with Belzers UW-CSS hypothermic flush solution (Dupont Pharmaceuticals, E. I. Dupont de Nemours and Company, Wilmington, Delaware) at a perfusion pressure of 60 cm of water until the renal artery and vein were cannulated and the effluent was clear. Catheterization of the renal vessels resulted in a cold ischemic time of less than 15 minutes.

The kidney was coupled to the extracorporeal perfusion circuit as illustrated in figure 3.1 and reperfused by the femoral artery via the renal artery cannula. Arterial blood flow to the isolated autoperfused kidney was continuously measured using a calibrated (Weber 1991) electromagnetic flow meter (Gould Statham Model SP2202) with a cannulating in-line flow probe (Gould Statham Model SP7509) as shown in figure 3.1. Renal venous blood drained into a thermally insulated reservoir and was returned to the femoral vein using a peristaltic pump. The volume of blood in the extracorporeal circuit was kept constant at approximately 180 ml. Following revascularization, the kidney was placed in the thermally regulated isotonic saline bath at 38°C and a 100 unit/kg/hr I.V. heparin infusion was initiated. In order to prevent renal edema and improve renal blood flow, 25 g of mannitol as a 20% solution was administered intravenously over 10 minutes, the dosage was subsequently reduced to 15 g per hour as a 20 % solution. Urine output from both

the autoperfused kidney and contralateral control kidney was measured every half hour.

The autoperfused kidneys from the group of animals used to investigate the effect of hyperthermia on renal blood flow were heated at rate of approximately $0.5^{\circ}\text{C}/\text{min}$ to 43°C . For the normothermic control group of animals the identical experimental protocol was followed with the exception that the kidneys were not heated, but maintained at constant temperature of 38°C after revascularization. The group of control animals used to measure renal blood flow during general anesthesia ($n=7$), did not undergo laparotomy, but were surgically prepared for the microsphere measurements as described below.

3.2.4 Microsphere blood flow measurements

After allowing approximately one hour to establish a stable renal hemodynamic state, renal blood flow (RBF) was sequentially measured using the radioactive microsphere reference sample technique as described in chapter 2. Microspheres were used to measure flow prior to the laparotomy, after mobilizing the renal artery and vein, after revascularization, and during hyperthermia treatment. Figure 3.2 illustrates the sequence of microsphere blood flow measurements. The detailed procedure for the measurement of blood flow with radioactive microspheres has been reported previously (Heymann *et al.*, 1977, Akyürekli *et al.* 1995a, see chapter 2).

After the animals were euthanized both the autoperfused and contralateral control kidney were removed and sliced into eight 1.5 cm thick transverse sections as illustrated in figure 3.3 and weighed. In this manner total and regional blood flow could be determined. The activity of the blood reference samples and each tissue section was counted and blood flow calculated as previously described (Akyurekli *et al.* 1995, Chapter 2).

3.3 Results

3.3.1 *Effect of anesthesia alone*

In the group of control animals used to investigate the effect of general anesthesia (n=7), a 1-1.5% concentration of halothane did not appear to alter renal blood flow over the measurement period. Figure 3.4 illustrates the effects of halothane anesthesia on blood flow to both the left and right kidneys. Although there were some variations, total blood flow to both kidneys remained approximately equal and constant at 240 ml/min over the measurement period. The maximum coefficient of variation observed for any kidney at the 95% confidence level was 26% over the measurement period.

3.3.2 *Normothermic autoperfused kidney*

Figure 3.5 describes the effects of the surgical protocol on renal blood flow before and after revascularization in the normothermic group of animals. Prior to the laparotomy blood flow to both kidneys was approximately equal and within the range previously observed for halothane anesthetized pigs. Following laparotomy and mobilization of the renal vessels of the kidney selected for autoperfusion, blood flow to both kidneys decreased, with the decrease being approximately uniform. Blood flow to the mobilized kidney declined from the pre-laparotomy level of 216 ± 17 ml/min to a value of 172 ± 31 ml/min, while flow to the contralateral control kidney decreased from a pre-laparotomy level of 212 ± 19 ml/min to a value of 184 ± 17 ml/min.

Following revascularization there was a gradual but progressive decline in total blood flow to both kidneys with the decrease in blood flow being more pronounced in the autoperfused than the contralateral kidney. Flow to the autoperfused kidney decreased by approximately 45% from the pre-laparotomy level to a value of 121 ± 24 ml/min immediately following revascularization. Blood flow to the autoperfused kidney decreased by approximately 70% from the pre-laparotomy level to a value of 65 ± 8 ml/min one hour after the first blood flow measurement following revascularization. In comparison, blood flow to the contralateral control

kidney decreased by 40% one hour after the first blood flow measurement following revascularization. The urine output of the autoperfused kidney was approximately 15-25 ml/hr which was lower than that of the contralateral control kidney (40-50 ml/hr).

Figure 3.6 describes the changes in blood flow distribution in a normothermic (38°C) autoperfused kidney, before and following revascularization. Blood perfusion is greater in the tissue samples nearer the superior and inferior poles of the kidney. The distribution shows a progressive decrease in blood flow following revascularization. An example of the blood flow distribution in the contralateral kidney is described in figure 3.7. The blood flow distribution appears similar to that of the autoperfused kidney with the exception that the reduction in flow following revascularization is not as great as in the autoperfused kidney. Table 3.1 shows the ratio of mean total blood flow to the normothermic autoperfused kidney to that of the contralateral control kidney, prior to laparotomy, after mobilizing the renal vessels and following revascularization. The ratio prior to the laparotomy was 1.02 and decreased to a value of 0.94 subsequent to mobilizing the renal vessels. Following revascularization the ratio decreased from an initial value of 0.74 to 0.52 by the end of the perfusion period.

3.3.3 Hyperthermic autoperfused kidney

Figure 3.8 shows the effects of the surgical protocol on renal blood flow before and after revascularization, as well as during heating. There is a significant difference between the vascular response of the autoperfused kidney and its contralateral control during normothermia and hyperthermia. This difference is evident when figures 3.5 and 3.8 are compared. During heating blood flow to the contralateral kidney did not decrease to the same extent as did the flow to the contralateral kidney in the normothermic group, even though flow to the heated kidney declined much more dramatically. Furthermore, the decrease in flow is less rapid in the contralateral kidney during heating than during normothermia.

Prior to the laparotomy total flow to both kidneys was approximately equal. Subsequent to mobilizing the renal vessels, blood flow to the surgically manipulated kidney decreased from the pre-laparotomy level of 225 ± 19 ml/min to a value of 164 ± 20 ml/min, while blood flow to the contralateral kidney decreased from 241 ± 20 ml/min to 171 ± 22 ml/min. This is identical to the response observed in the normothermic group of animals. Following revascularization blood flow to the contralateral kidney was 169 ± 30 while flow to the autoperfused kidney decreased by approximately 40% below the pre-laparotomy level to 132 ± 7 ml/min. The decrease is similar to that observed in the normothermic group. Heating resulted in a dramatic decline in blood flow to the autoperfused kidney, with the decrease in flow being significantly greater than that resulting from revascularization alone. This is evident when total blood flow following revascularization in the normothermic group of animals (Fig. 3.5) is compared to the total blood flow during heating in the hyperthermic group of animals (Fig. 3.8). After one hour of hyperthermia at 43°C , blood flow to the autoperfused kidney decreased by approximately 85% from the pre-laparotomy level to 33 ± 8 ml/min, while blood flow to the contralateral control kidney decreased by 40%. The urine output for the heated autoperfused kidneys was approximately 5-10 ml/hr, which was much lower than that of the contralateral control kidney (45-60 ml/hr) and the normothermic autoperfused kidney (15-25 ml/hr).

Figure 3.9 describes the changes in blood flow distribution to the heated (43°C) autoperfused kidney, before and following revascularization. Prior to heating, blood flow was greater in the tissue samples near the superior and inferior poles than near the central region of the kidney. The similar observation was made for the normothermic group of animals (Fig. 3.6). However, during heating the blood flow distribution changes, with tissue samples near the poles seeing a greater reduction in blood flow than samples near the central region of the kidney.

The blood flow distribution in the contralateral kidney is illustrated by the example shown in figure 3.10. Blood flow to tissue samples nearer the poles remained greater than in the central region of the kidney. The blood flow

distribution prior to heating appears similar to that of the autoperfused kidney, however, during heating flow does not decrease as dramatically as in the case of the autoperfused kidney. The similar observation was also noted in figure 3.8 describing total renal blood flow in the hyperthermic group of animals.

Table 3.2 enumerates the ratio of mean total blood flow to the heated autoperfused kidney to that of the contralateral control kidney, prior to the laparotomy, after mobilizing the renal vessels and after re-perfusion. During heating blood flow to the autoperfused kidney decreased substantially and continued to decrease, reaching flows as low as 20 ml/min. The ratio prior to the laparotomy was 0.93 and 0.92 subsequent to mobilization of the renal vessels. Following revascularization the ratio decreased from an initial value of 0.72 to 0.22 by the end of the hyperthermia treatment.

3.3.4 Comparison of blood flow measurements

There is close agreement between the microsphere measurements of blood flow and those obtained using the electromagnetic (Gould Statham) flow probe. The comparison of the two independent methods of flow measurement are shown in figure 3.11. Results of these two independent methods of flow measurement ($n=28$) are in good agreement, with a correlation coefficient of $R=0.98$. The comparison between microsphere measurements of blood flow and those obtained using the Transonic flow probe have been reported (Akyurekli *et al.* 1995, Chapter 2) and are also in good agreement with a correlation coefficient of $R=0.99$.

3.4 Discussion

The present study investigates the changes in renal blood flow in an autoperfused porcine kidney model under conditions of hot water immersion heating. In order to isolate the effect of heat on tissue blood flow, control experiments were undertaken to establish the vascular changes occurring as a consequence of the surgical protocol. Accordingly, this involved establishing the effect of anesthesia and autoperfusion on renal blood flow. The autoperfused porcine kidney provided a large animal organ model in which the organ was

thermally isolated and regulated while the contralateral intact kidney was used as a internal control. The kidney maintained relatively normal function since blood flow and urine output were sustained even though diminished.

In the group of control animals used to examine the effects of prolonged general anesthesia, a 1-1.5% concentration of halothane did not appear to alter renal blood flow over the measurement period (Fig. 3.4), total blood flow to both kidneys remained approximately uniform and constant although the initial effect was not measured. Moreover, the values of renal blood flow observed were well within the normal range for halothane anesthetized pigs (150-300 ml/min) (Tranquilli *et al.* 1982, Manohar and Parks 1984).

Prior to the laparotomy, blood flow to both kidneys was approximately equal and well within the range observed in the control group used to examine the effect of halothane on renal blood flow. Following the laparotomy and mobilization of one kidney, blood flow to both the surgically manipulated and contralateral control kidney decreased from the pre-laparotomy level. A uniform decrease in renal blood flow was observed in both the normothermic (Fig. 3.5) and hyperthermic group of animals (Fig. 3.8). The reasons for these changes in renal blood flow are probably multiple. Renal circulation is very sensitive to blood loss. Even relatively small hemorrhages, which do not decrease the systemic blood pressure, can result in striking decreases in the renal blood flow. The reduction affects primarily the superficial cortex. Other factors which may result in reduced renal blood flow include surgical stress and anesthesia, both of which can alter autonomic and neuroendocrine function. Norepinephrine and epinephrine, released by sympathetic postganglionic fibers and by the adrenal medulla, respectively, cause renal vasoconstriction and renin release from the juxtaglomerular apparatus. Secretion of renin leads to the production of angiotensin II, a potent renal vasoconstrictor. Anesthetic drugs may also alter renal blood flow not only by direct renal effects but also by producing changes in cardiovascular function and/or neuroendocrine activity. An anesthetic could alter renal blood flow indirectly by reducing myocardial contractility and changing the cardiac output, by activating the

sympathetic nervous system to increase renal vascular resistance, or by increasing endogenous secretion of ADH (antidiuretic hormone) to produce renal vasoconstriction. Halothane, at clinically relevant concentrations is known to decrease myocardial blood flow and cardiac output (Greenblatt *et al.* 1992).

A progressive decrease in blood flow following revascularization of the nephrectomized kidney was observed in the normothermic group of animals (Fig. 3.5). However, the decrease in blood flow was significantly greater in the heated autoperfused kidney in which case the decrease appears to be specific to the revascularized kidney, since the respective contralateral kidney did not experience a similar decrease. This is also supported by the fact that the decrease in blood flow following revascularization is more significant in the reperfused kidney than in the contralateral kidney and by the fact that the rate of blood flow decrease is greater in the revascularized than the contralateral kidney (see fig. 3.5). The decrease may be the result of ischemic damage. The continual decrease in blood flow to both kidneys following revascularization is probably the result of hemorrhage that resulted from filling the extracorporeal perfusion circuit (approximately 180 ml). In addition, blood loss at the operative site owing to systemic heparinization, would tend to increase renal vascular resistance as a result of various reactive mechanisms. In the normothermic group of animals the blood flow distribution within the kidney was observed to be similar in both the autoperfused and contralateral kidney before and following revascularization (see figures 3.6 and 3.7). This may indicate that the mechanism responsible for the decrease in blood flow is acting uniformly throughout the kidney.

During hyperthermia a significant difference was observed between the vascular response of the autoperfused kidney and its respective contralateral control kidney. Heating the autoperfused kidney resulted in a dramatic decrease in blood flow. A similar decrease was not observed in the contralateral kidney during heating, indicating that the mechanism responsible for the decrease in blood flow may be specific to the heated organ. Moreover, the observed decrease in blood flow to the autoperfused kidney during hyperthermia is also significantly different from that

observed following reperfusion in the normothermic group. In the case of the normothermic autoperfused kidney, the ratio of normothermic autoperfused blood flow to contralateral control kidney flow (Table 3.1) decreased from an initial value of 0.74 to a final value of 0.52. In contrast the ratio in the case of the heated autoperfused kidney decreased from 0.72 to 0.22.

When the blood flow distribution in the heated (Fig. 3.9) and contralateral (Fig. 3.10) kidney are compared the indication is that the effect of hyperthermia is specific to the heated kidney (as discussed above). More importantly the response of the kidney does not appear to be uniform. Assuming that the autoperfused kidney was uniformly heated, which is a reasonable assumption since heating was accomplished by hot water immersion and perfusion, then the greater reduction in flow observed at the poles of the kidney may indicate that hyperthermia results in reduced cortical blood flow. In contrast to tissue samples from the central region of the kidney which are a combination of cortical and medullary tissue. Tissue samples from the poles of the kidney are primarily cortical tissue. Furthermore, since cortical blood flow accounts for more than 90% of the total kidney blood flow, then it may be reasonable to assume that hyperthermia has a greater effect on cortical blood flow than medullary blood flow.

The observed decrease in blood flow to the autoperfused kidney during hyperthermia cannot be reflex in nature, since the perfused organ was denervated. Furthermore, the decrease in flow cannot be explained by the passive changes in viscosity, since an increase in the perfusate temperature should decrease blood viscosity (Brickford and Winton 1937, Harvy 1959). The decrease in blood flow may be the result of vascular damage or collapse caused by ischemia and compounded by the hyperthermia treatment. It is known that at extremely low levels of renal blood flow (less than approximately 75 ml/min/100g) no more oxygen can be extracted; consequently, renal cells become hypoxic and undergo ischemic damage. As a consequence of ischemic damage renin is secreted into the systemic circulation by the autoperfused kidney. Renin in turn causes the formation of angiotensin II, which circulates to the contralateral kidney to cause it also to retain

sodium and water. This consequently causes the arterial blood pressure to increase. The secretion of renin from the ischemic kidney never ceases, because the normal kidney will not retain quite enough sodium and water to allow the arterial pressure to rise high enough to return the blood flow to the ischemic kidney entirely to normal. However, modifying the effect of renin/angiotensin is the effect of mannitol. Mannitol increases plasma volume, filling pressures, and cardiac output, while decreasing systemic vascular resistance. Mannitol also increases renal blood flow to a greater extent than plasma volume is expanded (Gelman 1987). This may account for the greater flow in the contralateral kidney of the heated autoperfused kidney than that observed in the contralateral kidney of the normothermic autoperfused kidney.

The reasons for the changes in renal blood flow are probably multiple. This would not account, however for the differences in blood flow between the autoperfused and contralateral kidneys. Two major factors may account for these observed differences. First, there is the effect of passing blood through the perfusion system on the vasoconstrictive properties of the blood. Nizet (1967) has shown that simple stagnation of blood in a foreign vessel rapidly leads to release of vasoconstrictive substances chiefly from red blood cells. Others have implicated the release of vasoconstrictive substances from platelets during perfusion (Bracco and Curti 1954, Lustinec 1965, Nizet *et al* 1957). Denervation procedures may cause trauma and adversely affect renal function; even gentle manipulation of the kidney can have an effect on renal flow. Handley and Moyer (1954) demonstrated this response in the dog and Mercer (1971) in the rat. Both studies observed that renal blood flow and glomerular filtration rate decreased as much as 50% with renal manipulation. Cross-perfusion studies performed in the dog by Wexler and Kao (1970), Pearce (1968) and Takeuchi *et al.* (1971) have caused similar decreases in renal blood flow. The present study demonstrates a similar fall in renal blood flow, a 20-30% reduction following mobilization of the renal vessels.

Seely and Boulpaep (1971) observed a similar decrease in blood flow in a normothermic autoperfused dog kidney. Gillenwater *et al.* (1964) observed no

change or a slight increase in renal vascular resistance when the temperature of an autoperfused dog kidney was increased from 40 to 42°C. Although the experimental conditions were different from ours the effect of heating appears to be the similar in both studies.

The results presented in this study have demonstrated the effect of hyperthermia on an isolated autoperfused porcine kidney. Although the surgical protocol alone had a significant effect on blood flow to both the autoperfused and contralateral kidneys, heating resulted in a more dramatic decrease in blood flow. The reduction in blood flow to the heated autoperfused kidney was significantly greater than that observed in the normothermic control group following revascularization. Moreover, heating the autoperfused kidney resulted in reduced blood flow to both kidneys, although the decrease was more severe in the heated than the contralateral kidney. The exact mechanisms for this reduced blood flow are not clear, but it is evident that in addition to systemic effects there may also be local factors acting on the heated kidney. Since heating the kidney can result in significantly reduced blood flow, clinical heating for cancer therapy may be contraindicated. Although heating the kidney would be difficult, mainly due to the organ's large blood flow compared to surrounding tissues, caution should be observed when treating tissues in the immediate vicinity of the kidney in order to avoid renal damage.

Figure 3.1. Diagram of the extracorporeal circuit for the autoperfused porcine kidney. Arterial blood perfusing the isolated kidney is heated using a heat exchanger regulated by a constant temperature circulating water bath. Venous outflow from the perfused organ is returned to the systemic circulation by way of the femoral vein using a roller pump.

Figure 3.2. Illustration of the sequence in which radiolabeled tracer microspheres were administered.

Figure 3.3. Illustration showing the sectioning of the kidney into transverse tissue samples. Both the autoperfused and contralateral control kidney were removed and sliced into seven or eight 1.5 cm thick transverse sections and weighed.

Figure 3.4 Changes in renal blood flow while anesthetized with halothane. Blood flow remained unchanged in both the left and right kidneys presence of 1.0-1.5% halothane. Mean of 7 determinations $\pm$ S.E.M are shown.

Figure 3.5. Total blood flow (mean $\pm$ S.E.M.) to the normothermic (38°C) autoperfused kidneys (n=6) and the contralateral control kidneys plotted against the elapsed time.

Figure 3.6. An example of blood perfusion in the eight tissue samples of the normothermic (38°C) autoperfused kidney plotted against the elapsed time.

Figure 3.7. An example of blood perfusion in the eight tissue samples of the kidney contralateral to the autoperfused normothermic (38°C) kidney plotted against the elapsed time.

Figure 3.8. Changes in total blood flow (mean $\pm$ S.E.M.) to the autoperfused kidneys (n=6) heated to 43°C and the control contralateral kidneys plotted against the elapsed time.

Figure 3.9. An example of blood perfusion in the eight tissue samples of the heated (43°C) autoperfused kidney.

Figure 3.10. An example of blood perfusion in the eight tissue samples of the kidney contralateral to the heated (43°C) autoperfused kidney.

Figure 3.11. Comparison of total renal artery blood flow as measured by the microsphere reference sample technique and a cannulating in-line electromagnetic flow probe.

	Before Laparotomy	After Kidney Mobilization	After Re-Perfusion			
Time (min)	0	50	120	140	160	180
Flow Ratio Normothermic/ Contralateral	1.02	0.94	0.74	0.62	0.54	0.52

Table 3.1. Ratio of the normothermic kidney (n=6) mean total blood flow to that of the contralateral kidney prior to the laparotomy, after mobilizing the renal vessels and after re-perfusion.

	Before Laparotomy	After Kidney Mobilization	After Re-Perfusion			
				Hyperthermia		
Time (min)	0	50	120	140	160	180
Flow Ratio Hyperthermic/ Contralateral	0.93	0.92	0.72	0.30	0.28	0.22

Table 3.2. Ratio of the hyperthermic kidney (n=6) mean total blood flow to that of the contralateral control kidney prior to the laparotomy, after mobilizing the renal vessels, after re-perfusion and during hyperthermia treatment.

Chapter 4

Changes in renal blood flow during regional hyperthermia

4.1 Introduction

It has been long known that denervation has a profound effect on renal function which depends critically on the presence or absence of anesthesia and trauma due to the experimental surgery (Smith 1951). However, the role of the renal nerves in regulating renal blood flow has not been fully defined. In the previous chapter we examined the vascular response of the denervated autoperfused kidney to hyperthermia. The mediating role of the renal nerves in response to hyperthermia may be further explored by comparing the autoperfused kidney's response to that of an enervated in-situ kidney. In addition, in order to obtain data that relates more directly to physiological mechanisms controlling renal blood flow it is important to examine the effect of hyperthermia on the in-situ kidney. Renal blood flow is known to be strongly influenced by surgical stress and anesthesia (Prough and Foreman 1989), consequently, establishing the renal blood flow during anesthesia and following laparotomy and mobilization of the kidney is important to isolating and understanding the vascular response of the kidney during hyperthermia.

As part of our investigation of the vascular response of the kidney to hyperthermia, we have examined the changes in renal blood flow in a porcine kidney before, during and after a period of regional microwave hyperthermia, using two independent flow measurement techniques. In order to isolate the effect of heat on tissue blood flow, control studies were undertaken to establish the baseline renal blood flow during prolonged anesthesia and following thermocouple insertion.

4.2 Materials And Methods

4.2.1 Animals

Acute experiments were performed on 25 healthy pigs of either sex and of mixed breeding, weighing between 21 and 42 kg with a mean weight of $34 \text{ kg} \pm 1 \text{ kg}$ (Mean $\pm$ S.D.). Twelve of these animals were randomly selected for regional microwave hyperthermia treatment, while the remaining thirteen animals were used for control studies. From the control group, seven animals were used to establish the effect of a 7-8 hour exposure to the anesthetic on renal blood flow. These animals did not undergo laparotomy, but were surgically prepared for the microsphere measurements as described below.

Two sets of control experiments were conducted in order to determine whether or not insertion of thermocouples into the renal paranchyma would alter or modify the vascular response of the kidney during hyperthermia. Three animals were used to examine the effect of thermocouple insertion. These animals were prepared in the identical manner as those in the hyperthermic group, but underwent sham heating. A further three animals were used to determine the vascular response of the kidney to heat when thermocouples were not implanted. These animals were prepared in an identical manner to those in the hyperthermic group, with the exception that thermocouples were not inserted into the kidney. In all cases, the native contralateral kidney was used as an internal control.

4.2.2 Hyperthermia system

Kidneys were heated with constant net forward power (NFP) to 43°C for 60 minutes by means of a single direct contact 915 MHz dielectricly loaded rectangular wave guide as described in chapter 1. De-ionized water was pumped through the bolus from a constant temperature water bath, held at a temperature of 25°C .

The thermocouple arrays were oriented perpendicular to the E-field and along the longitudinal axis of the kidney in the sagittal plane as illustrated in figure 4.1. Thermometry data was collected with a computer controlled data acquisition system.

(Keithley DAS 500) at a sampling rate of 3 kHz. Each reported temperature measurement was the average of 100 samples.

4.2.3 Preparation for heating

The kidney was mobilized following median laparotomy. After isolating the renal artery a 4 mm R-series perivascular ultrasonic flow probe (Transonic Systems Inc. Ithaca NY) connected to a calibrated flow meter (Transonic Systems Inc. Ithaca NY, Model T101) was placed around the vessel. To provide a path for the thermocouple arrays, four 18 gauge blunt-end stainless-steel trocars, oriented along the longitudinal axis of the kidney in the sagittal plane (see Fig. 4.1) were introduced through the renal capsule into the parenchyma. Once the trocars were in position the stylets were withdrawn and the thermocouples placed within the trocars. The trocars were then retracted, leaving just the thermocouples in the tissue. The thermocouples were positioned and held securely in place with Teflon buttons fastened to the inferior and superior ends of the thermocouple arrays with clips. Thermocouple arrays were placed on the anterior surface of the kidney just beneath the renal capsule and at approximately 1, 2 and 3 cm below the anterior surface. At the conclusion of each experiment a series of contiguous 1.0 cm thick transverse computed tomographic images of the kidney were made to record and verify the position of the thermocouples (Fig. 4.2).

4.2.4 Preparation for microsphere measurements

The detailed procedure for the measurement of blood flow with radioactive microspheres has been reported previously (Akyürekli *et al.* 1995, Chapter 2). Microspheres were used to measure blood flow at each of six distinct stages during the experiment. Figure 4.3 illustrates the sequence of microsphere blood flow measurements. Prior to the first microsphere measurement the animals had been anesthetized for approximately four hours. Measurements were made following mobilization and placement of the flow probe over the renal artery, after insertion of the thermocouple arrays into the kidney, at approximately 20 and 40 minutes into

the heating phase, as well as at the end of the treatment period. The final microsphere measurement was made approximately 20 minutes after the cessation of the hyperthermia session. In the case of the control animals used to examine the effect of prolonged anesthesia, renal blood flow measurements were made during the final five hours of anesthesia in order to coincide with the time during which measurements were made in the hyperthermic group of animals.

After the animals were euthanized both kidneys were excised and sectioned into 1.5 cm thick transverse slices as illustrated in figure 4.4. Each tissue sample was weighed and identified with respect to the thermocouple positions within the sample. The tissue samples were prepared for gamma-ray counting as described by Akyurekli *et al.* (1995, Chapter 2) and activity of the tissue and blood reference samples were measured and the blood flow calculated.

4.3 Results

4.3.1 Effect of prolonged anesthesia

In the group of control animals used to investigate the effect of prolonged general anesthesia, a 1-1.5% concentration of halothane did not appear to alter total renal blood flow over the measurement period. Figure 4.5 shows the effects of halothane anesthesia on blood flow to both the left and right kidneys (mean $\pm$ S.E.M., n=7). Although there were some variations, total mean blood flow to both kidneys remained approximately constant and equal at 240 ml/min. The maximum coefficient of variation of blood flow to any kidney was 26% at the 95% confidence level. It should be noted that renal blood flow in the nonanesthetized animal was not measured, hence the initial effect of the anesthetic is not known.

4.3.2 Effect of thermocouple insertion

Figure 4.6 shows the mean total blood flow to both the mobilized and contralateral kidneys prior to and following insertion of thermocouple arrays into the mobilized kidney (mean $\pm$ S.E.M., n=3). When total blood flow in the anesthetic only control group (Fig. 4.5) is compared to that in the group of animals used to

investigate the effect of thermocouple insertion (Fig. 4.6), it is evident that subsequent to the laparotomy and mobilization of one kidney, blood flow to both kidneys decreased by approximately 30%. Note that the effect does not appear to be specific to the manipulated kidney, but is demonstrated in the control kidney as well.

The insertion of thermocouple arrays did not result in any significant blood loss and had no consequential effect on total blood flow to either kidney. Blood flow to both kidneys remained approximately constant and equal during the remainder of the observation period. The maximum coefficient of variation observed for any kidney following insertion of the thermocouples was 18% at the 95% confidence level. Figure 4.6 also demonstrates the close agreement between the two independent methods of total blood flow measurement.

4.3.3 Hyperthermia without thermocouple insertion

Figure 4.7 shows the changes in mean total blood flow to both the heated and contralateral control kidneys in the group of animals receiving hyperthermia without thermocouple insertion (mean $\pm$ S.E.M., n=3). Prior to heating blood flow to the mobilized and contralateral kidney were approximately constant and equal. During the heating phase blood flow to both kidneys decreased continuously, with no significant difference between the heated and the contralateral normothermic kidneys. Blood flow to both the heated and contralateral kidneys decreased continually from a pre-treatment value of 170 ± 3 ml/min to a post-hyperthermic value of 126 ± 10 ml/min. During the cooling phase blood flow to both kidneys remained depressed. There was some variation in the response from one experiment to next, but the general pattern was that of a progressive decrease in blood flow from the start of heating to the cessation of treatment. Figure 4.7 also shows the good agreement between the two independent flow measurement techniques, the Transonic flow probe used to measure total blood flow to the heated kidney and the microsphere reference sample technique.

4.3.4 Hyperthermia with thermocouple insertion

The final phase in determining the vascular response of the kidney to hyperthermia required examining the effect of heating the kidney in which thermocouple arrays were inserted. Figure 4.8 shows an example of temperature history monitored by a thermocouple located 1 cm beneath the anterior surface of the kidney and coincident with the principal axis of the wave guide. The temperature increased continuously during hyperthermia and returned to the pretreatment level following cessation of heating. Figure 4.9 shows the temperature profile monitored by a thermocouple array 1 cm beneath the anterior surface and positioned along the longitudinal axis (y-axis) of the kidney after 50 minutes of heating. The maximum temperature occurs along the principal axis of the wave guide (i.e. $x=0$, $y=0$ cm)

Figure 4.10 shows the effects of hyperthermia on the mean total blood flow to both the heated kidney in which thermocouples were inserted, and the contralateral control kidney (mean $\pm$ S.E.M., $n=12$). As can be seen there is good agreement between the two independent flow measurement techniques. Prior to insertion of the thermocouple arrays, total flow to both kidneys was approximately equal and consistent with data from the control group of animals used to examine the effect of heating without thermocouple insertion (See figure 4.7). During hyperthermia there was a continual decrease in total blood flow to both kidneys with flow to the heated kidney decreasing by approximately 20% below the level prior to inserting the thermocouples. Total blood flow to the contralateral kidney also decreased by approximately 20% to 132 ± 13 ml/min. After cessation of heating and as the tissue temperatures approached the pre-treatment levels, blood flow to the heated kidney was 123 ± 11 ml/min, approximately 25% below the value prior to thermocouple implantation, as was flow to the contralateral control kidney.

Figure 4.11 shows an example of the blood flow distribution within the contralateral kidney before, during and after heating the mobilized kidney. Prior to heating, blood flow in tissue samples obtained from the superior and inferior poles were greater than samples from the central region of the kidney. The similar

observation was made during heating. Although blood flow decreased there was no significant change in the distribution of blood flow. Figure 4.12 shows an example of the blood flow distribution in the mobilized kidney before, during and after heating with constant NFP. Prior to heating the blood flow distribution appears similar to that observed in the contralateral kidney. Blood flow was greater in the tissue samples near the superior and inferior poles than near the central region of the kidney. The insertion of thermocouples did not appear to alter the blood flow distribution in either the mobilized or contralateral kidney. However, during the heating phase blood flow to tissue samples nearer the central region of the kidney experienced a greater reduction in blood flow than tissue samples near the superior and inferior poles of the kidney, coinciding with the heating pattern. The decrease in blood flow appears to be specific only to the heated and not the contralateral kidney.

4.4 Discussion

It has been demonstrated by a number of investigators that blood flow in the skin and muscle of rodents, as well as the blood flow in the human skin, increases significantly when heated (Lokshina et al. 1985, Song 1984, Song *et al.* 1987a, Milligan *et al.* 1985). The present study demonstrates that the kidney's response to hyperthermia is markedly different from that of skin and muscle. The present study examined the changes in normal porcine kidney blood flow before, during and following a period of hyperthermia. An understanding of the vascular effects of the surgical protocol is essential to isolating the changes in tissue blood flow resulting from hyperthermia. Consequently, this study establishes the effect of prolonged anesthesia, kidney mobilization following laparotomy and the effect of thermocouple insertion, on total renal blood flow.

The effect of halothane anesthesia on renal blood flow has been extensively studied. The objective of this study was to examine renal blood flow over a prolonged period of anesthesia (approximately 7-8 hours). Continuous inhalation of a 1-1.5% concentration of halothane did not appear to alter total renal blood flow (Figure 4.5), although the initial effect of halothane was not determined. Total

blood flow to both kidneys remained approximately constant and uniform. Moreover, renal blood flow was well within the normal range for halothane anesthetized pigs (Tranquilli *et al.* 1982, Manohar and Parks 1984). Halothane is known to effect the glomerular filtration rate and urinary output (Tranquilli *et al.* 1982), however, studies of renal blood flow changes with halothane have yielded conflicting results. Reasons for this include difficulty in controlling other factors influencing renal blood flow, such as circulating catecholamine levels, as well as methodological problems with the measurement of renal blood flow. Initial studies, using techniques based on the clearance of inert filtered substances, concluded that halothane reduces renal blood flow. Later, direct measurement techniques indicated that clinical doses of halothane (1-2%) decrease renal vascular resistance but have little effect on total renal blood flow. Our observation that renal blood flow was well maintained during prolonged halothane anesthesia is in agreement with the findings of other investigators in dogs (Vatner and Smith 1974), monkeys (Amory *et al.* 1971) and rats (Miller *et al.* 1980).

Renal blood flow is known to be strongly influenced by surgical stress and anesthesia (Prough and Foreman 1989). Consequently, establishing the effect of kidney mobilization following laparotomy is important to isolating and understanding the vascular response of the kidney during hyperthermia. Mobilization of one kidney following the laparotomy resulted in a 30% reduction in total renal blood flow as is evident when figures 4.5 and 4.6 are compared. This decrease was not isolated to the mobilized kidney, but was also demonstrated in the contralateral control kidney as well. This is in agreement with a previous study by Akyürekli *et al.* (1995b, Chapter 3) which demonstrated a nonspecific decrease in blood flow to both the surgically manipulated as well as the contralateral kidney. The uniform reduction in total renal blood flow may be attributable to a number of mechanisms. Surgical stress and anesthesia can both alter autonomic and neuroendocrine function. Surgical stress may alter renal function by changing renal blood flow, glomerular filtration rate (GRF), or renal tubular function. Anesthetics can also influence renal function by producing changes in cardiovascular function

and or neuroendocrine activity. For example, an anesthetic could alter renal blood flow indirectly by reducing myocardial contractility and by changing cardiac output, by activating the sympathetic nervous system to increase renal vascular resistance, or by increasing endogenous secretion of ADH to produce renal vasoconstriction. Norepinephrine and epinephrine, released by sympathetic postganglionic fibers and by the adrenal medulla respectively, cause both renal vasoconstriction and renin release from the juxtaglomerular apparatus. Secretion of renin leads to the production of angiotensin II, a potent renal vasoconstrictor. Renin release is also stimulated by arterial baroreceptors and, in the kidney, by a low ratio of serum sodium to serum potassium. A further complicating factor involved in surgical and anesthesia-induced alterations in renal function is the extensive capacity of the kidney to adapt to changing conditions. It is known that low to moderate doses of halothane do not appear not to change the autoregulation of renal blood flow (Prough and Foreman 1989).

Implanting thermocouple arrays in the renal parenchyma following mobilization of the kidney did not result in any significant blood loss or further reduction in renal blood flow. Blood flow to both the mobilized and contralateral kidney remained approximately constant and uniform (Figure 4.6). This observation along with the fact that blood flow to both kidneys remained approximately equal during prolonged anesthesia (Figure 4.5), demonstrates that the microsphere infusion technique used in this study provided uniform mixing of microspheres (Heymann *et al.* 1977) an important prerequisite to measuring blood flow using the microsphere technique.

Although some variation in the vascular response to heating was observed, the general pattern remained that of a progressive decrease in blood flow from the start of heating to the cessation of treatment. In the group of animals receiving hyperthermia treatment without the insertion of thermocouple arrays, both the heated and contralateral control kidney experienced a reduction in blood flow (Figure 4.7). This decrease was not specific to the heated kidney, but was observed in both kidneys. Heating in the group of animals in which thermocouple arrays

were implanted resulted in a similar reduction in total blood flow to both kidneys. The similarity in response between these two groups of animals indicates that thermocouple insertion does not alter the response of the kidney to heat, at least under the conditions of the present study. Hence, the reduction in blood flow during hyperthermia is likely the consequence of heating. On cessation of treatment, while tissue temperatures approached the pre-treatment levels renal blood flow remained depressed.

The vascular response to heat observed in this study is in contrast to that observed by Akyürekli *et al.* (1995b, Chapter 3) during immersion heating of an autoperfused kidney. Heating the autoperfused kidney resulted in a more dramatic reduction in blood flow than observed in this study. Although blood flow to the contralateral kidney also experienced a reduction in flow the decrease was not as dramatic as in the autoperfused kidney. In the particular case of the autoperfused kidney the mechanism responsible for the decrease in blood flow appeared to be specific to the heated organ. In the present study reduction in total blood flow does not appear specific to the heated kidney, but rather affects total blood flow to both kidneys uniformly. The decrease in blood flow in both studies may be the result of vascular damage or collapse caused by heating. However, in the autoperfused kidney hyperthermia may have compounded the ischemic damage resulting in a more dramatic reduction in blood flow.

When the blood flow distribution in the heated (Figure 4.12) and contralateral (Figure 4.11) kidney are compared the indication is that hyperthermia affects the heated kidney differently. More importantly the response of the kidney does not appear to be spatially uniform. Since heating was concentrated to the central region of the kidney (Figure 4.9) the greater reduction in flow observed near the central region of the kidney than at the poles of the kidney may indicate vascular damage to the heated tissue. This is supported by the fact that the similar distribution was not observed in the contralateral kidney during heating. This is in contrast to the response observed by Akyürekli *et al.* (1995b, Chapter 3) in an immersion heated autoperfused kidney. In this case all tissues were elevated to the same temperature,

but tissue samples from the superior and inferior poles of the kidney experienced the greater reduction in blood flow. It was speculated that hyperthermia had a greater effect on cortical than medullary blood flow since tissue samples from the poles are primarily cortical tissue.

A number of mechanisms for the decrease in total renal blood flow during hyperthermia are conceivable. The decrease in flow may be the consequence of a systemic release of precursor substances as discussed above. Alternatively, the observed decrease in blood flow may be reflex in nature since the treated kidney remained enervated and both the treated and contralateral control kidney responded in a similar manner. It would be reasonable to speculate that some direct physical damage to the microvasculature might also cause a decrease in blood flow in the heated kidney. For example severe vascular damage accompanied by extravasation or leakage of blood components, and the ensuing clotting of blood would inevitably impede or occlude the blood flow. An increase in intra-tissue pressure due to heat-induced edema would also result in reduced blood flow. The fact that post-treatment blood flow remained depressed may indicate a form of vascular stasis resulting from the hyperthermia treatment. Vascular stasis has been described previously in different tissues in response to hyperthermia (Endrich, 1988, Endrich *et al.* 1988). Vascular damage may result in the systemic release of vasoconstrictive agents as discussed above, resulting in the decrease in flow seen in the contralateral kidney.

The exact mechanism of heat-induced decrease in renal blood flow requires further investigation using biochemical as well as histopathological techniques. The observation that blood flow decreases during hyperthermia suggests that the therapeutic advantage resulting from differential blood flow between tumors and normal kidney tissue may be lost during treatment. The increase in blood flow to normal tissues surrounding the tumor provide an advantage for delivering hyperthermia doses to these tumors (Prasse 1983). If the normal tissues do not increase their perfusion rates in response to hyperthermia treatments, this advantage is lost.

The results presented in this study have demonstrated the effect of hyperthermia on a porcine kidney. Although the surgical protocol alone had a significant effect on blood flow to both the mobilized and contralateral kidneys, heating resulted in a progressive decrease in renal blood flow. Moreover, heating one kidney resulted in reduced blood flow to both kidneys. The exact mechanisms for this reduced blood flow are not clear, but it is evident that in addition to systemic effects there may also be local factors acting on the heated kidney. Since heating the kidney can result in reduced blood flow, clinical heating for cancer therapy may be contraindicated. Although heating the kidney would be difficult, mainly due to the organ's large blood flow compared to surrounding tissues, caution should be observed when treating tissues in the immediate vicinity of the kidney in order to avoid possible renal damage.

Figure 4.1 Illustration of the 915 MHz microwave applicator, kidney and thermocouple orientation during hyperthermia treatment of the porcine kidney.

Figure 4.2 The transverse computed tomographic image of the kidney were made to record and verify the position of the thermocouple arrays.

Sequence of Microsphere Measurements

Figure 4.3 Illustration of the sequence in which radiolabeled tracer microspheres were administered. The numbers in the figure indicate the order in which microsphere blood flow measurements were made; (1) following mobilization of one kidney, (2) after implanting the thermocouple arrays, (3-5) during heating, and (6) post-hyperthermia.

Figure 4.4. Illustration showing the sectioning of the kidney into transverse tissue samples. Both the autoperfused and contralateral control kidney were removed and sliced into seven or eight 1.5 cm thick transverse sections and weighed.

Figure 4.5 Porcine renal blood flow under halothane anesthesia. Blood flow remained unchanged in both the left and right kidneys in the presence of 1.0-1.5% halothane. Mean of 7 determinations $\pm$ S.E.M are shown.

Figure 4.6 Effect of thermocouple insertion on total renal blood flow. Comparison of the normothermic kidney in which thermocouple were implanted and the contralateral control kidney. Mean of 3 determinations $\pm$ S.E.M are shown.

Figure 4.7 Changes in total blood flow to both the heated non-catheterized kidney and the contralateral control kidney. Mean of 3 determinations $\pm$ S.E.M. are shown.

2 OF 2

PM-1 3½"x4" PHOTOGRAPHIC MICROCOPY TARGET
NBS 1010a ANSI/ISO #2 EQUIVALENT

PRECISIONSM RESOLUTION TARGETS

Figure 4.8 An example of the temperature history in the porcine kidney during heating with a surface cooled 915 MHz microwave dielectrically loaded wave guide. The thermocouple was located 1 cm beneath the anterior surface of the kidney and coincident with the principal axis of the wave guide.

Figure 4.9 An example of the thermal distribution in the porcine kidney during heating. The thermocouple array was located 1 cm beneath the anterior surface of the kidney.

Figure 4.10 Changes in total blood flow to both the heated and contralateral control kidney. Mean of 12 determinations $\pm$ S.E.M are shown.

Figure 4.11 An example of the blood flow distribution in six tissue samples of the contralateral control kidney during hyperthermia.

Figure 4.12 An example of the blood flow distribution in the kidney heated to 43°C.

Chapter 5

Changes in cerebral blood flow during interstitial hyperthermia

5.1 Introduction

In chapter 3 and 4 we examined the effect of hyperthermia in autoperfused and in-situ kidneys respectively. We now turn our attention to the effect of hyperthermia on blood flow in cerebral tissue. Malignant gliomas are devastating cerebral neoplasms for which there is currently no widely successful treatment (Shapiro 1986, Leibel and Sheline 1987). As tumors of the neuroglial tissue, malignant gliomas are highly invasive but rarely metastasize outside the central nervous system (Hochberg and Pruitt 1980, Chourair *et al.* 1986). Radiation is an effective adjuvant to surgical resection (Walker *et al.* 1978, 1979), but radiation-induced necrosis of normal cerebral tissue limits the dose that can be delivered by external beam therapy (Sheline 1980, Brennan *et al.* 1993). The use of brachytherapy sources such as ^{125}I have produced encouraging clinical results in patients with recurrent malignant primary or metastatic brain tumors (Bernstein and Gutin 1981, Gutin *et al.* 1987, 1991, Prados *et al.* 1992). However, despite the effectiveness of interstitial irradiation the problem of normal brain tolerance and peripheral tumor recurrence remains (Gutin *et al.* 1984, Prados *et al.* 1992). Because brain tumors are resistant to other treatment modalities it has been hypothesized that adjuvant interstitial hyperthermia may improve local control of brain tumors, based on biological rationale provided by in vitro studies (Dewey *et al.* 1980) and on encouraging clinical results with external beam radiotherapy and hyperthermia for superficial tumors (Overgaard 1990). However, before the use of hyperthermia can

be optimized, it is essential to determine the effects of heat on normal brain tissue blood flow.

The purpose of this study was to examine the changes in regional cerebral blood flow before, during and after a period of interstitial microwave hyperthermia in a normal porcine brain model using radiolabeled tracer microspheres. In order to isolate the effect of heat on tissue blood flow, control studies were undertaken to establish the baseline cerebral blood flow during prolonged anesthesia and following the insertion of the microwave antenna and thermocouple arrays.

5.2 Materials And Methods

5.2.1. Animals

After obtaining protocol approval from the National Research Council of Canadian Animal Care and Use Committee, studies were conducted on eighteen healthy pigs of either sex and of mixed breeding, weighing between 22 and 47 kg with a mean weight of $36 \text{ kg} \pm 6 \text{ kg}$ (Mean $\pm$ SD). Twelve of these animals were randomly selected for the interstitial microwave hyperthermia treatment, while the remaining six animals were used as a control group to assess the effects of the surgical preparation for hyperthermia treatment. The twelve animals selected for hyperthermia treatment were randomly divided into two groups of six. The first group received interstitial treatment with 23.5 Watts constant net forward power (NFP), while the second group received interstitial treatment with 27.0 Watts constant NFP over the same time period.

5.2.2. Hyperthermia system

Porcine brains were locally heated for a period of 60 minutes by means of a single 915 MHz interstitial helical coil microwave antenna inserted within a blunt close ended polypropylene catheter (Sovereign, 2.3 mm O.D., 1.9 mm I.D.). The design of the helical antenna, illustrated in figure 5.1 has been described previously (Sato and Stauffer 1988) and consists of an eight turn helical coil, 2.5 cm in length shorted to the antenna feed line at the tip with an outer diameter of 1.2 mm.

The thermometry system described in detail in chapter 1. Four type K (Chromel-Alumel) thermocouple arrays were used. Each array consisted of 6 thermocouple junctions with a 5.0 ± 0.5 mm separation. The axial thermocouple arrays were positioned within 18 gauge polypropylene catheters placed parallel to the helical antenna at radial distances of 6.5 mm, 11.5 mm, 16.5 mm and 21.5 mm from the antenna catheter surface. The catheters had a inside diameter of 0.9 mm and a wall thickness of 0.1 mm.

5.2.3. Preparation for heating

A 6.5 cm diameter section of tissue was incised from the frontal scalp region and a custom made Lucite antenna support/template, illustrated in figure 5.2, was mounted securely onto the frontal bone with stainless steel screws. Using the template as a guide, four burr holes were drilled into the cranium for thermometry and a fifth burr hole was drilled for the antenna. The antenna was inserted lateral to the midline to depth of 3.5 cm after which steel stylets were used to insert four 18 gauge catheters parallel to the antenna at radial distances of 6.5 mm, 11.5 mm, 16.5 mm and 21.5 mm from the antenna catheter surface. The template guided all the catheter implantation trajectories and kept the catheters parallel and accurately spaced. The stylets were withdrawn and thermocouples positioned within the catheters. The reference thermocouple junction was positioned 6.5 mm from the antenna surface and at a depth of 2.2 cm from the brain surface which corresponded to the depth of the peak SAR of the antenna. At the conclusion of each experiment the insertion length for the antenna and the thermocouple arrays were re-measured and a series of contiguous 1.0 cm thick transverse CT scans of the brain were obtained in order to verify the position of the catheters (Fig. 5.3).

5.2.4. Blood flow measurement

Microspheres were used to measure blood flow before and after the intracerebral insertion of the antenna and thermocouple arrays, as well as during and following heat treatment. Figure 5.4 illustrates the sequence of microsphere blood flow measurements. The detailed procedure for the measurement of blood

flow with radioactive microspheres has been reported previously (Heymann *et al.* 1977, Akyürekli *et al.* 1985, see chapter 2). At the conclusion of the experiment animals were euthanized and the brain was excised and dissected into coronal sections as illustrated in figure 5.5 and each section weighed. In this manner blood flow in tissue surrounding each thermocouple array could be determined.

5.3 Results

In order to assess the effects of the surgical preparation on regional cerebral blood flow (rCBF), animals in the control group underwent the same surgical preparation as did animals receiving hyperthermia treatment, with the exception that they were not heated. Figure 5.6 shows the mean rCBF for the normothermic ($38.2 \pm 0.1^\circ\text{C}$) control group prior to catheterization, after catheterization, as well as during and after sham heating. The pre- and post-catheterization blood flows were averaged over all 18 animals. In most cases blood flow was slightly higher approximately 10 minutes after intracerebral catheterization followed by a relatively constant blood flow of 53 ml/min/100g for the remainder of the treatment. The mean post-catheterization blood flow was 69.3 ± 2.7 ml/min/100g and the pre-catheterization value was 68.5 ± 1.9 ml/min/100g.

A distinct difference was observed in both the temperature profile and rCBF response between the two groups of animals receiving hyperthermia treatment. For the animals heated with 23.5 Watts constant NFP, the maximum tissue temperature at the position of the reference thermocouple junction ranged from 42.6 – 43.8°C and was reached within 5-10 minutes of applying power. Over the remaining course of treatment the temperature gradually decreased such that at the end of treatment the tissue temperature was on average $0.4 \pm 0.1^\circ\text{C}$ (mean $\pm$ S.E.M.) below the maximum mean temperature. Figure 5.7 shows a representative observation of the change in the temperature profile along with rCBF before, during and after heating. Blood flow increases continuously during the heating period from 89.0 ± 4.2 ml/min/100g to 103.5 ± 2.5 ml/min/100g, reducing the temperature of the tissue at the end of heating by 0.6°C . In contrast figure 5.8 shows a representative observation of the change in both the temperature profile and rCBF before, during

and after heating for the group of animals treated with 27.0 Watts constant NFP. Blood flow initially increased from 73.2 ± 2.6 ml/min/100g to a maximum of 81.3 ± 2.3 ml/min/100g, this was then followed by a dramatic decrease of nearly 30 ml/min/100g. This decrease in rCBF was paralleled by an increase in the tissue temperature of approximately 2°C by the end of the heating period. In these animals the tissue temperature increased continuously over the treatment course reaching maximum values in the range $46.2\text{-}51.3^{\circ}\text{C}$ by the end of the treatment.

Figure 5.6 shows the mean rCBF response for the two groups of treated animals before, during and after heating along with the control group values. For the animals heated with 23.5 Watts constant NFP to temperatures between 42.6°C - 43.8°C , local blood flow increased gradually over the period of heating. As a consequence, blood flow at the end of the treatment was 90.5 ± 6.3 ml/min/100g approximately 32% greater than the pre-treatment value and nearly 70% above the control value at the end of heating. For the animals heated with 27.0 Watts constant NFP to temperatures between $46.2\text{-}51.3^{\circ}\text{C}$, local flow initially increased to a maximum value of 81.4 ± 5.2 ml/min/100g then decreased to 56.8 ± 7.7 ml/min/100g. As a consequence, at the end of heating the mean rCBF was approximately 18% below pre-treatment values and approximately 7% above the control value. On cessation of hyperthermia, tissue temperatures in both groups of interstitially treated animals approached pre-treatment levels of $38\text{-}39^{\circ}\text{C}$. The mean rCBF for both groups of hyperthermically treated animals returned to just below pre-treatment levels, however, they remained approximately 26% above the normothermic control value.

Cardiac output was not measured throughout these experiments, however, arterial blood pressure and pulse rate remained stable during the entire experiment with a mean value of 92.3 ± 5.6 mmHg and 128 ± 15 bpm respectively. Arterial pCO_2 and pO_2 also remained constant during the experiment with a mean value 40.97 ± 1.24 mmHg and 210.3 ± 3.9 mmHg respectively.

5.4 Discussion

Interstitial microwave hyperthermia is currently being considered for use with brachytherapy in the treatment of malignant brain tumors, and initial results suggest that this combination therapy is feasible (Roberts *et al.* 1986, 1987, Sneed *et al.* 1991). Such an approach is appealing in that microwave antennas can be placed into the same catheters as used for the insertion of radioisotopes. However, in attempting to optimize such treatment it is important to consider the effects of temperature on blood flow. The vascular response to the tissue temperature in normal brain is considered important because tissue adjacent to or in the heat treatment field is forced to increase blood flow, which may result in increase in local tissue metabolism as described in the literature (Lilly *et al.* 1984).

The time course of vascular response as determined by perfusion has been extensively documented for skin and muscle (Lokshina *et al.* 1985, Milligan and Panjehpour 1985, Shrivastav *et al.* 1983, Song *et al.* 1987a, 1987b). However, the effects of hyperthermia on cerebral blood flow in large animals have not been extensively studied. The classic response of tissue following heating is vasodilation which increases perfusion to some maximum value after some time into the treatment (Milligan *et al.* 1983). This increase in blood flow is followed by a subsequent decrease back to baseline levels. A slow decrease in temperature over time during heating, and after a quasi-equilibrium has been established, has been documented by many investigators both in normal tissue and in tumors during hyperthermia treatment for cancer therapy (Jain and Ward-Hartly 1984, Waterman *et al.* 1987, Lyons *et al.* 1989). For localized hyperthermia in the brain using an interstitial microwave antenna array, Roberts *et al.* (1986) have shown that power must be increased during a given hyperthermia episode. These changes are believed to be related to blood flow alterations occurring dynamically during a given hyperthermia treatment.

In the present study, rCBF in normal porcine brain was measured during interstitial microwave heating using the microsphere reference sample technique. The response of rCBF to the tissue temperature was determined for both modest

and severe hyperthermia. Results of our study show that rCBF changes dynamically with increase in tissue temperature and that above a "critical threshold" temperature blood flow decreases. When heated with modest power levels (23.5 Watts) temperatures in the range 42.6 - 43.8°C are reached, resulting in a moderate, but gradual increase in blood flow over the entire heating period. Figure 5.6 shows the mean rCBF response for two groups of treated animals before, during and after heating along with the control group values. For the animals heated with 23.5 Watts constant NFP to temperatures between 42.6°C - 43.8°C, local blood flow increased gradually over the period of heating. As a consequence, blood flow at the end of the treatment was 90.5 ± 6.3 ml/min/100g approximately 32% greater than the pre-treatment value and nearly 70% above the control value at the end of heating. Tissues heated to temperatures <44°C have been shown to survive and remain viable 1 week after treatment (Fike *et al.* 1991). This increase in blood flow may result from an increase in the regional tissue metabolism (Lilly *et al.* 1984) which is temperature dependent. In contrast, cerebral blood flow measured in animals heated at higher power levels (27 Watts), to temperatures >46°C, decreased, probably the result of primary vascular damage (intravascular coagulation and endothelial swelling). Figure 5.6 shows that at temperatures >46°C local flow initially increased to a maximum value of 81.4 ± 5.2 ml/min/100g then declined to 56.8 ± 7.7 ml/min/100g. As a consequence, at the end of heating the mean rCBF was approximately 18% below pre-treatment values and approximately 7% above the control value at the end of heating. The decrease in rCBF was paralleled by an increase in the tissue temperature of approximately 2°C by the end of the heating period. The deleterious effect of interstitial hyperthermia at temperatures >46°C induce irreversible vascular injury (necrosis) (Fike *et al.* 1991). It is known that high blood flow rates may increase temperature gradients and narrow thermal profiles within the heated tissues thereby increasing the effects of thermal conduction (Strang and Patterson 1980). Increased blood flow in heated tissue may also increase drug delivery; this is especially critical for lipid soluble drugs such as nitrosoureas (Levin and Kabra 1974). Increased drug delivery to the tissue,

accompanied by biological drug-cell interactions at elevated temperatures (Hahn 1979, Dewey 1984) may enhance the tumor response to chemotherapeutic drugs. However, it should be taken into consideration that in regions of compromised blood flow higher in normal tissue than in tumor tissue (Sato *et al.* 1989), when the tissue surrounding the tumor might be simultaneously heated to the same temperature. In conclusion, the cerebral vascular response to tissue temperature in normal cerebral tissue may play a significant role in the application and modeling of hyperthermia to brain tumor treatment, either when used alone or in conjunction with ionizing radiation and chemotherapeutic agents.

Figure 5.1 Schematic illustration of the 915 MHz helical antenna used to heat the porcine brain.

Figure 5.2. Schematic illustration of the 915 MHz helical antenna support and template.

Figure 5.3. Transverse CT image of the porcine brain showing the position of the thermocouple and antenna catheters.

Figure 5.4. Illustration of the sequence in which radiolabeled tracer microspheres were administered.

Figure 5.5. Schematic illustration of the superior aspect of the brain with position of thermocouples, microwave antenna and transverse tissue slices.

Figure 5.6. Mean blood flow for the control group of animals, and the group of animals heated with 23.5 Watts constant NFP and 27.0 Watts constant NFP.

Figure 5.7. A representative observation showing the gradual increase in blood flow and decrease in temperature for a normal porcine brain during a 60 minute heating with 23.5 Watts constant NFP. The temperature was monitored with the reference thermocouple 6.5 mm from the antenna catheter surface at the depth of the peak SAR.

Figure 5.8. A representative observation showing the gradual increase then decrease in blood flow resulting in the increase in tissue temperature for a normal porcine brain during a 60 minute heating with 27 Watts constant NFP. The temperature was monitored with the reference thermocouple 6.5 mm from the antenna catheter surface at the depth of the peak SAR.

Chapter 6

Porcine muscle blood flow distribution during hyperthermia

6.1 Introduction

Up this point we have examined the effect of thermal insult on blood flow of cerebral and kidney tissues, neither of which has had extensive attention in the literature. This chapter deals with muscle blood flow which has been the subject of extensive studies by other authors for a variety of animal models. This work differs from previous reported work in that better spatial resolution of blood flow is obtained than reported in other hyperthermia work.

The effects of hyperthermia on local and regional blood flow in the muscle of animals (Lokshina *et al.* 1985, Milligan and Panjehpour 1985, Shrivastav *et al.* 1983, Song *et al.* 1987a, 1987b) has been extensively investigated. However, most existing data describing the effects of hyperthermia on muscle blood flow detail only the average blood flow in a single heated volume of tissue and little is known about the heat-induced changes in the spatial distribution of blood flow. The purpose of this study was to examine the spatial distribution of blood flow in normal porcine muscle before, during and following a period of regional microwave hyperthermia. In order to isolate the effect of heat on tissue blood flow, control studies were undertaken to establish the effect of anesthesia and thermocouple insertion on the baseline muscle blood flow.

6.2 Materials And Methods

6.2.1 Animals

Acute experiments were performed on 32 healthy pigs of either sex and of mixed breeding, ranging in weight from 20 to 54 kg with a mean weight of 34 ± 1 kg (Mean $\pm$ S.E.M.). Twenty one of these animals were randomly selected for regional microwave hyperthermia treatment, while the remaining eleven animals were used for control studies. Six animals from the control group were used to establish the muscle blood flow distribution during prolonged anesthesia. These animals were not implanted with thermocouple arrays, but were surgically prepared for the microsphere measurements as described below. The remaining five animals were used to determine the muscle blood flow distribution after implanting the thermocouple arrays.

6.2.2 Hyperthermia and thermometry system

The goal of the experimental protocol was to investigate the vascular response of normal porcine muscle before, during and following hyperthermia to a target temperature of 44.5°C . The lateral aspect of the hind limb was heated with 36 Watts of constant net forward power (NFP) for a period of 60 minutes by means of a single direct contact 915 MHz rectangular wave guide as described in chapter 1. De-ionized water was supplied to the bolus from a constant temperature bath maintained at 23°C using a constant volume pump.

The thermometry system has been characterized and described in detail in chapter 1. Four type K (Chromel-Alumel) thermocouple were used to measure temperature throughout the heating cycle. Each array consisted of 5 thermocouple junctions with a 20.0 ± 0.5 mm separation. The thermocouple arrays were oriented perpendicular to the wave guide E-field. The coordinate system used to specify the location of the applicator and thermocouples is illustrated in figure 6.1. The x-axis was oriented perpendicular to the E-field, and the y-axis parallel to the E-field, while the z-axis was coincident with the principal axis of the wave guide. The origin of the coordinate system was located along the z-axis at the surface of the

skin. In this coordinate system thermocouples were oriented perpendicular to the E-field in the x-z plane.

6.2.3 Preparation for heating

To provide a path for the thermocouple arrays, four 18 gauge stainless-steel trocars, oriented postero-laterally above the femur in a transverse plane were introduced percutaneously into the muscle of the hind limb. With the trocars in position the stylets were withdrawn and replaced by thermocouple arrays. The trocars were then retracted, leaving only the thermocouples in the tissue. The thermocouples were positioned and held securely in place with Teflon buttons attached to the inferior and superior ends of the thermocouple arrays with clips. Thermocouple arrays were inserted beneath the surface of the skin and in the subcutaneous tissue at approximately 0.3 and 0.5 cm respectively and in the muscle at approximately 1.5 and 3.5 cm beneath the surface of the skin. For each thermocouple array, the thermocouple junctions were positioned along the x-axis at approximately -4, -2, 0, 2 and 4 cm from the z-axis.

6.2.4 Preparation for microsphere measurements

Microspheres were used to measure blood flow at six distinct stages during the experiment. The detailed procedure for the measurement of blood flow with radioactive microspheres has been reported previously (Akyürekli *et al.* 1995a, Chapter 2). Figure 6.2 illustrates the sequence of microsphere blood flow measurements. Prior to the first blood flow measurement the animal had been under general anesthesia for approximately four hours. During this time the animal was surgically prepared for the microsphere measurements. Subsequently, microsphere blood flow measurements were made following insertion of the thermocouple arrays, and at approximately 10, 30 and 50 minutes into the heating phase. The final microsphere measurement was made approximately 10 minutes after the cessation of the hyperthermia session. In the case of the control animals used to examine the effect of prolonged anesthesia, blood flow measurements were made

during the final three hours of anesthesia in order to coincide with the time period during which measurements were made in the hyperthermic group of animals.

At the conclusion of each experiment both hind limbs were excised for microsphere assays of blood flow. Subsequently, a series of contiguous 1.0 cm thick transverse computed tomographic images of the heated limb were obtained in order to record and verify the position of the thermocouples (Figure 6.3). From the heated limb, the section of tissue beneath the applicator, measuring approximately 13×7 cm was excised to the level of the femur. This section of tissue was then dissected into five 2 cm thick transverse slices as illustrated in figure 6.1. From each transverse slice, tissue surrounding the thermocouple junctions at a depth of 1.5 cm and 3.5 cm was dissected using a 2 cm diameter curricular punch. The tissue samples were weighed and labeled according to the position of the tissue slice and depth of the tissue sample. Tissue samples surrounding the thermocouples at a depth of 1.5 cm were labeled "A" and those surrounding thermocouples at a depth of 3.5 cm were labeled "B". The contralateral control limb was prepared in an identical manner. The preparation technique provided 10 tissue samples whose central temperature was accurately known and whose average blood perfusion could be accurately calculated using the microsphere reference sample technique (Akyürekli *et al.* 1995a, Chapter 2).

6.3 Results

6.3.1 Effect of prolonged anesthesia

In the group of control animals used to investigate the effect of prolonged anesthesia, a 1-1.5% concentration of halothane did not appear to alter the blood flow distribution during the final three hours. Figures 6.4 and 6.5 show the effect of halothane anesthesia on the blood flow distribution in tissue samples A and B respectively (mean $\pm$ S.E.M., $n=6$). Although there were some variations, the flow distribution appears spatially uniform with no significant difference in flow within or between the A and B tissue samples. During the course of the measurements, blood flow remained relatively constant in both the A and B tissue samples. The

average perfusion rate over all the tissue samples during the observation period was 10 ± 1 ml/min/100g (Mean $\pm$ S.E.M.) and the maximum coefficient of variation within any sample over three hours was less than 11%. The maximum coefficient of variation between all the samples at any time was also less than 11%.

6.3.2 Effect of thermocouple insertion

Figures 6.6 and 6.7 show the blood flow distribution in the control group of anesthetized animals in which thermocouples were implanted (Mean $\pm$ S.E.M., $n=5$). As in the previous control group the blood flow distribution remained spatially uniform with no significant difference in blood flow within or between A and B tissue samples. As well, blood flow remained relatively constant over the course of the measurements. The average perfusion rate over all the tissue samples during the measurement period was 11 ± 1 ml/min/100g (Mean $\pm$ S.E.M.) and the maximum coefficient of variation within any sample was less than 10%. The maximum coefficient of variation between all the samples at any time was less than 12%. There was no significant difference in the magnitude or distribution of muscle blood flow between those animals receiving anesthesia alone (see figures 6.4 & 6.5) and those receiving anesthesia and undergoing the insertion of thermal probes (see figures 6.6 & 6.7).

6.3.3 Effect of hyperthermia after thermocouple insertion

Figure 6.8 shows a representative observation of the thermal profile in muscle during heating for 60 minutes with constant NFP. The temperature of the superficial tissue sample 3A rapidly reached the target temperature and subsequently decreased. The deeper tissue sample 3B exhibited a similar response, with the exception that the initial rate of increase and maximum temperature reached were less than in the superficial tissue sample. During the remainder of the heating cycle the temperature of both A and B samples gradually increased. On cessation of treatment the temperature of both samples approached the pretreatment value. Figure 6.9 shows an example of the rate of heating in tissue samples 1A through to

5A. As can be seen the maximum rate of heating occurs near the center of tissue sample 3A at a depth of 1.5 cm.

Tables 6.1 and 6.2 show the temperature and corresponding blood flow data for both A and B tissue samples respectively. Each data point represents the average of a group of 21 animals. The blood flow distribution data for both the superficial A and deeper B samples before, during and following hyperthermia treatment are shown graphically in figure 6.10. Prior to heating and after insertion of the thermocouple arrays, mean blood flow in the A tissue samples was 10.7 ± 0.6 ml/min/100g and 10.4 ± 0.4 ml/min/100g in B samples. The magnitude and distribution of blood flow following the insertion of the thermal probes appears spatially uniform, with no significant difference in blood flow within or between A and B samples. This is in good agreement with results from the control studies investigating the effect of thermocouple insertion during anesthesia on muscle blood flow (see figures 6.4-6.7).

During heating the greatest increase in blood flow was observed in the region of the applicator's peak SAR where the tissue temperature was greatest; this included tissue samples 2A, 3A and 4A (see figure 6.9). The tissue blood flow peaked approximately 15-30 minutes into treatment, after which flow gradually decreased. The temperature of tissue sample 2A increased from 36.2 ± 0.5 prior to heating to $42.0 \pm 0.3^\circ\text{C}$ during mid-heating and blood flow correspondingly changed from a baseline value of 11.2 ± 3.6 ml/min/100g to 42.3 ± 7.0 ml/min/100g during mid-heating. The temperature of tissue sample 3A increased the most, from a mean pre-treatment value of $36.6 \pm 0.6^\circ\text{C}$ to $44.6 \pm 0.2^\circ\text{C}$ during initial heating. Consequently blood flow in the tissue sample increased approximately 4-fold from a pre-treatment value of 10.5 ± 2.0 ml/min/100g to 43.4 ± 5.4 ml/min/100g during mid-heating. By the end of the hyperthermia session blood flow had decreased to 34.6 ± 4.1 ml/min/100g and the temperature increased to $44.3 \pm 0.8^\circ\text{C}$. In tissue sample 4A the temperature increased from a mean pre-treatment value of $35.4 \pm 0.7^\circ\text{C}$ to $42.6 \pm 0.6^\circ\text{C}$ during the initial heating. As a result blood flow in the tissue sample increased from a pre-treatment value of 9.7 ± 1.5 ml/min/100g to

43.9±7.0 ml/min/100g during mid-heating. Tissue samples 1A and 5A along with the deeper B tissue samples experienced significantly less of an increase in blood flow. Blood flow returned to pre-treatment levels in all the tissue samples within approximately 10 to 15 minutes after cessation of treatment.

The blood flow distribution in the contralateral control limb during hyperthermia is described in figure 6.11. The control limb did not exhibit any increase in flow and the distribution remained relatively uniform with flow averaging 9.1 ± 0.4 ml/min/100g over the treatment period. It should be noted that although cardiac output was not measured throughout these experiments, blood pressure and pulse did not vary significantly during any of the treatment sessions or from animal to animal.

6.4 Discussion

Blood flow is a critical parameter for obtaining satisfactory temperature distributions during clinical hyperthermia. This study examined the changes in blood flow distribution in normal porcine muscle before, during and after 60 minutes of regional microwave heating with constant NFP. In order to isolate the effect of heat on tissue blood flow, control studies were undertaken to establish the effect of general anesthesia and thermocouple insertion on muscle blood flow.

During the final three hours of a seven hour period of stable general anesthesia the blood flow distribution remained uniform with no significant difference between the superficial and deep tissue samples. The insertion of thermal probes into the muscle during general anesthesia did not alter the magnitude or distribution of blood flow from that of the control group receiving anesthesia alone. Under these conditions it is clear that the vascular response of muscle observed during hyperthermia treatment may be attributed to heating. It is known that general anesthesia significantly alters microvascular function (Longnecker and Harris 1979), therefore, the magnitude of the vascular response of muscle to heat may be different under conditions in which general anesthetics are not applied. Several

studies have demonstrated that halothane can alter endothelium-dependent vasodilation (Uggeri *et al.* 1992, Stone and Johns 1989)

The effect of hyperthermia on muscle blood flow observed in this study was in general agreement with observations reported in the literature (Cobb *et al.* 1974, Guy *et al.* 1974, Dickson and Calderwood 1980, Song *et al.* 1980a, b, Milligan and Panjehpour 1985, Rappaport and Song 1984, Song *et al.* 1985, Milligan 1987). Muscle blood flow increased on average by a factor of 4 within 15-30 minutes of heating with constant NFP to a target temperature of 44.5°C. With further heating the elevated level of blood flow could not be sustained, resulting in a gradual increase in the tissue temperature. Before, during and following hyperthermia treatment blood flow in the contralateral control limb was not perturbed and remained approximately constant and uniform. In contrast blood flow to the heated tissue increased significantly and was not uniformly distributed. Although cardiac output was not measured throughout these experiments, blood pressure and pulse did not vary significantly during any of the treatment sessions or from animal to animal. These results indicate that the increase in blood flow during hyperthermia is primarily a regional phenomena and does not result from any significant increase in systemic blood flow. The degree of change in the spatial distribution of blood flow is dependent on the tissue temperature distribution, that is the greater the temperature the greater the blood flow. Thus the regions of tissue in which peak temperatures were measured also correspond to the same regions in which the blood flow increased the most substantially (see tables 6.1 & 6.2). The greatest increase in blood flow was observed in the region of the applicator's peak SAR, which included tissue samples 2A, 3A and 4A (see figure 6.9). The cessation of hyperthermia treatment resulted in the restoration of tissue blood flow and temperature to their pretreatment levels.

A number of mechanisms for the increase in blood flow in the heated muscle are conceivable. The increase in blood flow during heating is caused, at least in part, by the dilation of local vessels and the recruitment of capillaries following stimulation by vasoactive compounds. During rest 20 to 25 per cent of muscle

capillaries are open. However, during heating dormant capillaries open and account for most of the increased blood flow. Cardiac output will increase provided local vasodilatation increases the venous return of blood back to the heart.

Local blood flow is known to be influenced to a certain extent by the metabolic status of the tissue, particularly by the demand for oxygen. Hyperthermic treatment enhances the metabolic activity of the tissue resulting in an increased demand for oxygen, which in turn may cause an increase in blood flow (Granger *et al.* 1975). Indications are that vasoactive compounds play a major role in heat induced increases in blood flow, particularly at high temperatures. Silva and Antonio (1960) reported that bradykinin, a potent vasoactive compound, was released in the paw of the rat upon heating at 43.5-45°C. It is conceivable that our observation in the present study of the initial increase and the subsequent decrease in blood flow in muscle resulted from an initial release and subsequent depletion of vasoactive compounds such as bradykinin and/or histamine (Shea *et al.* 1973, Gabbiani and Badonnel 1975).

It would be reasonable to speculate that, in addition to the above biochemical mediation some direct physical damage to the vessels might cause an increase or decrease in blood flow in the heated tissues. For example, mild damage to the basement membrane may result in vasodilation, which would ease the perfusion of blood through the microvasculature. It is also possible that severe vascular damage accompanied by leakage of blood compounds and the resulting clotting of blood would inevitably impede or occlude blood flow. An increase in interstitial pressure due to heat-induced edema would also reduce blood flow.

Comparison of the temperature profile in muscle (Tables 6.1 & 6.2) and the changes in tissue blood flow during heating confirms the close relationship between tissue temperature and blood flow during heating. The significant drop in tissue temperature following the initial rapid rise (Figure 6.8) results from increased heat dissipation due to blood flow. Some studies have shown that the response of tumor vasculature to heat is different from that of normal tissues. A number of studies have demonstrated that severe vascular damage accompanied by hemorrhage and

stasis are induced in experimental tumors when heated to 41-43°C, whereas the blood flow in normal tissue significantly increases at the same temperature (Song 1984). It would seem reasonable to expect that preferential heating of the tumor, and thus greater damage in the tumor relative to the normal tissue would occur provided the tissue normal tissue surrounding the tumor was heated equally so that the differential response was uniform.

The results presented in this study have demonstrated the effect of hyperthermia on porcine muscle blood flow. General anesthesia alone and thermocouple insertion during anesthesia had no significant effect on the muscle blood flow distribution prior to heating. Regional microwave heating generated a non-uniform blood flow distribution which was a function of the tissue temperature distribution. Blood flow was greater in those tissues samples in which higher temperatures were recorded and less in those sampled further from the peak SAR. The increase in blood flow appears to be primarily a local phenomenon. Although blood flow in muscle may be considered to be uniform prior to heating, this does not hold during hyperthermia treatment. Therefore, the non-uniform nature of the blood distribution during heating should be incorporated into any practical bioheat transfer model.

Figure 6.1 Transverse section through the microwave applicator and muscle showing the coordinate system used to describe the thermocouples' orientation and position during hyperthermia treatment. The muscle beneath the wave guide was sliced into five sections labeled 1 through to 5 and tissue samples labeled "A" and "B" dissected from each slice.

Sequence of Microsphere Measurements

Figure 6.2 Illustration showing the sequence in which radiolabeled tracer microspheres were administered. The numbers in the figure indicate the order in which microsphere blood flow measurements were made; (1) during anesthesia alone (2) after implanting the thermocouple arrays, (3-5) during heating, and (6) post-hyperthermia.

Figure 6.3 An example of the transverse computed tomographic image of the treated limb used to verify the position of the thermocouple arrays.

Figure 6.4 The effects of halothane anesthesia on the muscle blood flow distribution in the superficial "A" tissue samples. Blood flow remained unchanged in presence of 1.0-1.5% halothane. Mean of 6 determinations $\pm$ S.E.M are shown.

Figure 6.5 The effects of halothane anesthesia on the muscle blood flow distribution in the deeper "B" tissue samples. Blood flow remained unchanged in presence of 1.0-1.5% halothane. Mean of 6 determinations $\pm$ S.E.M are shown.

Figure 6.6 The effects of thermocouple insertion during general anesthesia on the muscle blood flow distribution in the superficial "A" tissue samples. Blood flow remained unchanged after insertion of the thermocouples. Mean of 5 determinations $\pm$ S.E.M. are shown.

Figure 6.7 The effects of thermocouple insertion during anesthesia on the muscle blood flow distribution in the deeper "B" tissue the samples. Blood flow remained unchanged after insertion of the thermocouple arrays. Mean of 5 determinations $\pm$ S.E.M. are shown.

Figure 6.8 A representative observation of the temperature history for A and B tissue samples for thermocouples at a depth of 1.5 and 3.5 cm during heating with 40 Watts constant NFP.

Figure 6.9 The initial rate of heating in the superficial "A" tissue samples as a function of the thermocouple position.

Time Course	1 A	2 A	3 A	4 A	5 A
*Baseline	Mean± S.E.M	Mean± S.E.M	Mean± S.E.M	Mean± S.E.M	Mean± S.E.M
Temp.	36.6±0.3	36.2±0.5	36.0±0.6	35.4±0.7	35.1±0.8
Blood Flow	11.4±4.0	11.2±3.6	10.5±2.0	9.7±1.5	10.5±1.8
†Initial-Heating					
Temp.	40.9±0.5	42.8±0.3	44.6±0.2	42.6±0.6	41.1±0.5
Blood Flow	17.4±4.4	37.3±7.8	39.0±5.8	36.2±7.4	20.1±7.1
Mid-Heating					
Temp.	40.8±0.4	42.0±0.3	43.7±0.7	42.0±0.7	40.4±0.5
Blood Flow	19.9±4.5	42.3±7.0	43.4±5.4	43.9±7.0	22.0±4.4
Final-Heating					
Temp.	41.1±0.5	42.3±0.3	44.3±0.8	42.4±0.7	40.8±0.5
Blood Flow	18.3±3.9	36.1±5.3	34.6±4.1	36.4±6.0	22.0±3.9
Cooling					
Temp.	36.7±0.3	36.5±0.4	36.2±0.6	35.6±0.6	35.3±0.7
Blood Flow	9.9±1.8	11.2±1.4	8.9±0.9	10.8±1.1	9.5±0.8

*After thermocouple insertion; † when set point is at the target temperature of 44.5°C.

Table 6.1 Temperature of the superficial "A" tissue samples and corresponding blood flow measured during regional microwave hyperthermia.

Time Course	1B	2B	3B	4B	5B
*Baseline	Mean± S.E.M	Mean± S.E.M	Mean± S.E.M	Mean± S.E.M	Mean± S.E.M
Temp.	37.3±0.3	37.2±0.3	37.2±0.2	36.7±0.4	36.4±0.4
Blood Flow	11.1±2.0	11.3±2.1	10.4±1.8	9.2±1.6	10.1±1.3
Initial- Heating					
Temp.	39.7±0.2	40.9±0.2	42.2± 0.3	41.0±0.2	39.7±0.3
Blood Flow	10.0±1.5	10.9±2.9	17.3±5.0	12.9±3.3	9.8±1.3
Mid-Heating					
Temp.	39.9±0.2	40.8±0.2	41.4±0.2	40.9±0.1	40.0±0.2
Blood Flow	10.6±1.4	12.6±2.7	18.7±3.5	18.5±4.4	12.1±2.0
Final-Heating					
Temp.	40.1±0.2	40.9±0.2	41.7±0.2	41.3±0.3	40.3±0.3
Blood Flow	10.7±1.3	12.9±2.5	18.5±3.2	14.2±2.3	11.4±1.5
Cooling					
Temp.	37.3±0.3	37.2±0.3	37.3±0.2	36.8±0.4	36.7±0.4
Blood Flow	8.0±0.7	9.8±1.4	9.7±1.1	8.3±1.0	9.4±0.9

*After thermocouple insertion; † when set point is at the target temperature of 44.5°C.

Table 6.2 Temperature of the deeper "B" tissue samples and corresponding blood flow measured during regional microwave hyperthermia.

Figure 6.10 The changes in blood flow distribution in "A" and "B" tissue samples from the volume of tissue beneath the microwave applicator. Mean of 21 determinations $\pm$ S.E.M are shown.

Figure 6.11 The effects of hyperthermia on the blood flow distribution to the contralateral control muscle in both superficial and deep tissue samples. Mean of 21 determinations $\pm$ S.E.M are shown.

Chapter 7

Evaluation of a simple thermal model using *in-vivo* tissue blood flow data

7.1 Introduction

One impediment to realizing hyperthermia's true therapeutic potential is the lack of ability to model and plan patient treatment. A fundamental block to developing these planning procedures is the absence of a verified model for accurately predicting temperatures. Recognizing the importance of being able to predict temperatures *in-vivo* in these and other applications, several models have been proposed for calculating temperature distributions in tissue (e.g. Pennes 1948, Chen and Holmes 1980, Weinbaum *et al.* 1984a, b, Weinbaum and Jiji 1985, Lagendijk 1987 Mooibroek and Lagendijk 1991). The model most commonly used for hyperthermic treatment planning is the Pennes bioheat transfer equation (1948), mainly because of its simplicity and practicality. The Pennes equation is an approximation which is intended to predict the overall characteristics of the temperature field and does not have the ability to predict the temperature around isolated large vessels.

In this work we examine the applicability of this model in predicting the temperature distribution in muscle when heated externally with 915 MHz microwave radiation. In order to assess the applicability of this formulation to the experimental case under consideration it is necessary to determine the effects of various parameters such as perfusion and the surface heat transfer coefficient on the tissue temperature distribution. Furthermore, it is necessary to compare these predictions to *in-vivo* temperature measurements while using blood flow data from

the same experiment. At present, there is little data in the literature comparing the predictions of bioheat transfer models to detailed *in-vivo* temperature measurements. For the Pennes bioheat transfer equation there are several scattered comparisons to animal data for hyperthermia situations (Roemer *et al.* 1989, Winget *et al.* 1986) and some comparisons to clinical human hyperthermia tumor treatments (Strohbehn *et al.* 1986). The most comprehensive of these comparisons is the *in-vivo* study of Sekins and Emery (1982) and Sekins *et al.* (1984) for normal muscle tissues in humans. In all these cases, the general experimental trends presented were well represented by the simple Pennes bioheat transfer equation predictions, as were Pennes' original normothermic experimental data (Pennes, 1948).

The purpose of this study was to examine the effects of muscle blood flow and surface cooling on the tissue temperature distribution. The study also compares the predictions of the one dimensional Pennes bioheat equation to *in-vivo* temperature measurements made under hyperthermic conditions. The blood flow values used in the model were the corresponding *in-vivo* values. Finally a comparison is made between the predictions of the one dimensional Pennes bioheat equation and a two dimensional finite element model.

Temperature distributions in the treatment field were measured by twenty thermocouples in arrays of five thermocouples per array (Gerig *et al.* 1992). Temperature was measured at a sampling rate of 3 kHz, before during and after heating. The tissue blood flow was measured using the microsphere reference technique (Heymann *et al.* 1977). This technique provided a map of blood perfusion distribution at six different times during the experiment.

In this present work the measured blood flow and power distribution from the porcine muscle experiment were used as input data to the Pennes model and the subsequent predictions were compared to the measured temperature profiles. The values of blood flow measured in these experiments and used as input data for the Pennes model under normothermic and hyperthermic conditions are given in Table 7.1.

7.2 The animal model

The temperature profiles calculated in this study are compared to temperature measurements obtained during regional heating of the porcine thigh muscle (Akyürekli *et al.* 1995c, Chapter 6). The goal of the experimental protocol was to measure the changes in muscle blood flow before, after and following hyperthermia treatment. The lateral aspect of the porcine hind limb was heated with 36 Watts constant net forward power for a period of 60 minutes by means of a single direct contact 915 MHz rectangular wave guide. The applicator was constructed from WR 430 wave guide (internal dimensions 10.92×5.46 cm) and filled with alumina powder. In order to improve the applicator-tissue coupling and to insure effective surface cooling a temperature controlled de-ionized water bolus was interposed between the aperture of the applicator and the tissue surface.

7.3 The semi-infinite slab model

7.3.1 The one-dimensional bioheat equation

The differential equation describing the temperature distribution in a homogenous and isotropic slab of tissue can be written as follows (Pennes 1948, Hodson 1979 *et al.*, Bowman 1981, Roemer 1988)

$$k \frac{\partial^2 T'(x,t)}{\partial x^2} - w_b c_b [T'(x,t) - T_\infty] + S(x,t) = \rho_t c_t \frac{\partial T'(x,t)}{\partial t} \quad (7.1)$$

where $T'(x,t)$ [$^{\circ}\text{C}$] is the local tissue temperature, T_∞ [$^{\circ}\text{C}$] is the perfusing arterial blood temperature which is assumed to be a constant and equal to the ambient body core temperature, t is time in seconds, k [$\text{Wm}^{-1}\text{C}^{-1}$] is the tissue thermal conductivity, w_b is the mass flow rate of blood per unit volume of tissue [$\text{kgm}^{-3}\text{s}^{-1}$] which is equal to $\rho_t \rho_b \omega$ where, ρ_t the tissue density [kgm^{-3}], ρ_b is the density of blood [kgm^{-3}] and ω is the volumetric blood flow rate per unit mass of tissue [m^3/kg]; c_b is the specific heat of blood [$\text{Jkg}^{-1}\text{C}^{-1}$], $S(x,t)$ [Wm^{-3}] is the applied

power density, and c_t [$\text{Jkg}^{-1}\text{C}^{-1}$] is the specific heat of the tissue. The term on the right of equation (7.1) describes the rate of change in the stored internal energy of the tissue, while the first term on the left of equation 7.1 describes heat transfer due to conduction. The blood perfusion term, $w_b\rho_b c_b [T'(x,t) - T_\infty]$ is modeled as a uniform heat-sink due to microvascular heat transfer, with blood supplied by larger vessels into the tissue volume at temperature T_∞ and exiting at the same temperature. The inherent assumptions of the bioheat equation have been reviewed in the literature (Hodson *et al.* 1979, Bowman 1981 Roemer 1988).

The applied power density $S(x,t)$ [Wm^{-3}] was modeled as

$$S(x,t) = S_0 \exp(-x/\delta) H(t) \quad (7.2)$$

the penetration or skin depth, δ , is the distance at which the power density has decreased to $1/e$ (37%) of its value at the surface (Marion and Heald 1980). $H(t)$ is the unit step function and S_0 is the power density at the surface of the tissue. The validity of equation (7.2) follows from the solution of Maxwell's equations for plane waves.

For convenience let $T(x,t) = T'(x,t) - T_\infty$, therefore equation (7.1) becomes

$$\beta \frac{\partial T}{\partial t} = \frac{\partial^2 T}{\partial x^2} - \xi T + q \quad (7.3)$$

where $\beta = \rho_t c_t / k$, $\xi = w_b c_b / k$, and $q = S(x,t) / k$.

In the case of a semi-infinite plane of tissue the boundary conditions are:

$$T(x,0) = 0 \quad \forall x \geq 0 \quad (7.4)$$

$$T(\infty,t) = \lim_{x \rightarrow \infty} T(x,t) = 0, \quad t \geq 0 \quad (7.5)$$

$$\left. \frac{\partial T(x,t)}{\partial x} \right|_{x=0} = \frac{h}{k} (T + T_{\infty} - T_s), \quad t \geq 0 \quad (7.6)$$

where T_s is the temperature of the water bolus cooling the surface of the tissue. Initially the entire tissue plane is assumed to be at a uniform temperature equal to T_{∞} (equation 7.4). The third boundary condition describes the convective heat loss from the tissue surface. The coefficient h is the surface heat transfer coefficient given in units of $\text{W/m}^2\text{C}$. When water circulates through the bolus the heat transfer coefficient between the water and the bolus membrane is approximately 5000 $[\text{W/m}^2\text{C}]$. The plastic membrane represents a heat transfer coefficient $h = k_{\text{silicon}}/d_{\text{silicon}}$, for a membrane thickness of $d_{\text{silicon}} = 0.5 \text{ mm}$ and a thermal conductivity of $k_{\text{silicon}} = 0.2 [\text{W/mC}]$ the heat transfer coefficient is approximately 400 $\text{W/m}^2\text{C}$. The total heat transfer coefficient between water and the tissue surface is determined by the membrane thickness and is between 200-1000 $\text{W/m}^2\text{C}$. Provided there is optimal contact between the membrane and tissue surface, the tissue surface temperature approximately equals that of the circulating water. The thermophysical parameters of various tissues shown in table 7.1 were obtained from the literature (Chen *et al.* 1990). The perfusion values used in this study were obtained experimentally (Akyürekli *et al.* 1995c, Chapters 6) and are listed in table 7.2.

7.3.2 Steady-state solution when $q=0$

Prior to heating (i.e. $q=0$), the tissue temperature distribution is described by the steady-state solution to equation 7.3 (see Appendix A):

$$T(x,0) = \frac{h(T_s - T_{\infty})}{h + k\sqrt{\xi}} e^{-x\sqrt{\xi}}, \quad t \leq 0, x \geq 0 \quad (7.7)$$

Figure 7.2 presents the effect of blood perfusion on the steady-state temperature distribution in the x direction. The temperature profiles were calculated for muscle

where $T_{\infty}=38^{\circ}\text{C}$, $T_s=26^{\circ}\text{C}$ and $h=500\text{ W/m}^2\text{C}$. As would be expected the temperature gradient near the tissue surface varies significantly with perfusion, where steeper thermal gradients are generated when the tissue perfusion rate increases. The heat transfer coefficient as well as the tissue surface temperature T_s have a similar effect on temperature gradients near the surface. The larger the heat transfer coefficient and the greater the difference between the core (T_{∞}) and surface (T_s) tissue temperatures the steeper are the thermal gradients near the surface.

7.3.3 Steady state solution when $q \neq 0$

In the steady-state $\partial T / \partial t = 0$, and equation 7.3 reduces to the following

$$\frac{d^2 T}{dx^2} = \xi T - q_0 e^{-x/\delta} \quad (7.8)$$

where $q_0 = S_0/k$. The solution to equation 7.8, subject to the initial condition given by equation 7.4 and boundary conditions 7.5 and 7.6, is given (see Appendix A) as

$$T(x, \infty) = \frac{q_0}{\xi - 1/\delta^2} \left[e^{-x/\delta} - \frac{1/\delta + h/k}{\sqrt{\xi} + h/k} e^{-x\sqrt{\xi}} \right] + \frac{h(T_s - T_{\infty})}{h + k\sqrt{\xi}} e^{-x\sqrt{\xi}} \quad (7.9)$$

The microwave penetration depth δ in muscle is 3 cm at 915 MHz. The net forward power used in these calculations was 36 Watts and the power density is described by equation 7.2 and illustrated in figure 7.3. Figure 7.4 shows the solution (equation 7.9) to equation 7.8 as a function of blood perfusion when the heat transfer coefficient is $500\text{ W/m}^2\text{C}$. The perfusion values chosen are those measured experimentally and range from 5 to 50 ml/min/100g. In the absence of blood flow the surface temperature rise is proportional to k/h , and is relatively large. Examination of equation 7.9 shows that in this case the heating pattern is also greatly altered. The steady-state tissue temperature increase becomes an exponentially decreasing function of distance from the surface, with a decay

parameter equal to the energy penetration depth δ . The perfusion rate affects not only the maximum tissue temperature attained during heating but also the depth at which the maximum tissue temperature occurs. With greater tissue perfusion the temperature maximum occurs nearer the tissue surface. Figure 7.5 depicts the effect of changes in the heat transfer coefficient on the temperature profile in muscle when perfusion is 10 ml/min/100g. As the heat transfer coefficient increases the tissue surface temperature decreases as does the maximum tissue temperature. When $h=0$ W/m²°C a maximum tissue temperature of 61.1°C occurs at the surface, while with a heat transfer coefficient of 500 W/m²°C a maximum tissue temperature of 50.5°C occurs at a depth of 2.0 cm. As the heat transfer coefficient increases the thermal gradients near the surface also increase. Figure 7.6 shows the effect of the heat transfer coefficient on the temperature profile when muscle blood flow is 40 ml/min/100g. When $h=0$ W/m²°C a maximum tissue temperature of 44.5°C occurs at the surface, while with a heat transfer coefficient of 500 W/m²°C a maximum tissue temperature of 42.1°C occurs at a depth of 1.5 cm.

The position of the maximum tissue temperature can be controlled by choosing appropriate microwave wavelengths, and by cooling the tissue surface. The location of the temperature maximum as a function of δ can be determined by differentiating equation 7.9. If the tissue surface is maintained at a temperature T_s ,

$$x_{max} = \frac{l}{\sqrt{\xi} - l/\delta} \left\{ \ln(\delta \sqrt{\xi}) + \ln \left[1 + \left(\frac{\xi - l/\delta^2}{q_0} \right) (T_\infty - T_s) \right] \right\} \quad (7.10)$$

In the case of 915 MHz microwaves $\delta \leq l/\sqrt{\xi} \approx 3\text{cm}$ and $x_{max} \leq l/\sqrt{\xi}$. As δ decreases, or as the surface heat transfer coefficient decreases, the temperature maximum shifts nearer to the surface. The position of the temperature maximum can vary by approximately 1 cm and by less at shorter wave lengths. This follows from the fact that the longer the wavelength of the microwave radiation the greater

the penetration and hence the heating pattern is spread over a greater volume of tissue.

7.3.4 The transient solution

The transient solution for the tissue temperature is obtained from the Fourier transformation of equation 7.3 with respect to time and is given as:

$$\frac{d^2 g(x, i\omega)}{dx^2} - \eta^2 g(x, i\omega) + h_0 e^{-x/\delta} = 0 \quad (7.11)$$

where $\eta^2 = \xi + i\omega\beta$, $h_0 = iq_0/\omega\sqrt{2\pi}$ and $g_\omega(x, i\omega)$ is the Fourier transform of $T(x, t)$. Note that the Fourier transform of the unit step function $H(t)$ is $i/\omega\sqrt{2\pi}$. The solution to this equation has been calculated and is given by equation 7.9, where $T(x, t)$ is replaced by $g_\omega(x, i\omega)$, ξ by η^2 and q_0 by h_0 . Transforming this solution back into the time domain (Campbell and Foster 1948) gives:

$$\begin{aligned} T(x, t) = & \frac{q_0}{\xi - 1/\delta^2} \left\{ e^{-x/\delta} + \frac{1/\delta + h/k}{2(\sqrt{\xi} - h/k)} e^{x\sqrt{\xi}} \operatorname{erfc} \left(\frac{x}{2} \sqrt{\beta/t} + \sqrt{\frac{\xi t}{\beta}} \right) \right. \\ & - \frac{1/\delta + h/k}{2(\sqrt{\xi} + h/k)} e^{-x\sqrt{\xi}} \operatorname{erfc} \left(\frac{x}{2} \sqrt{\beta/t} - \sqrt{\xi t/\beta} \right) - \frac{1}{2} e^{\left(\frac{1}{\beta} (1/\delta^2 - \xi) - x/\delta \right)} \\ & \bullet \operatorname{erfc} \left(\frac{1}{\delta} \sqrt{t/\beta} - \frac{x}{2} \sqrt{\beta/t} \right) - \frac{1/\delta + h/k}{2(1/\delta - h/k)} \bullet e^{\frac{1}{\beta} (1/\delta^2 - \xi) + x/\delta} \\ & \bullet \operatorname{erfc} \left(\frac{1}{\delta} \sqrt{t/\beta} + \frac{x}{2} \sqrt{\beta/t} \right) + \frac{\frac{h}{k} (\xi + 1/\delta^2)}{(\xi - (h/k)^2)(1/\delta - h/k)} \bullet e^{\frac{(h/k)x + ((h/k)^2 - \xi)t}{\beta}} \\ & \left. \bullet \operatorname{erfc} \left(\frac{x}{2} \sqrt{\beta/t} + \frac{h}{k} \sqrt{t/\beta} \right) \right\} + \frac{h(T_s - T_\infty)}{h + k\sqrt{\xi}} e^{-x\sqrt{\xi}} \end{aligned} \quad (7.12)$$

where it is assumed that ξ and h/k are constant with respect to time. Figure 7.7 shows the variation in the temperature profile with time when the tissue perfusion is 10 ml/min/100g and the heat transfer coefficient is 500 W/m²°C. Figure 7.8 shows the variation in the temperature profile with time when the tissue perfusion is 40 ml/min/100g using the same heat transfer coefficient. Tissues with greater perfusion equilibrate more rapidly than those with lesser blood flow. The time required to reach equilibrium is determined by the time constants β/ξ and $\beta\delta^2$, hence as blood flow increases the time constant decreases. Moreover, temperature gradients are steeper near the surface of tissues with greater flow.

7.3.5 The finite element solution

In order to assess the validity of the one dimensional transient analytic solution presented in this study we will compare its predictions to those of a finite element method (FEM) solution of a previously reported two dimensional heat transfer problem Akyürekli (1988). The two dimensional problem consisted of a cylindrical volume of muscle tissue heated by a 1 MHz plane wave ultrasound transducer such that heating was along the z-axis of the cylinder. The thermophysical properties were assumed to be homogeneous, $k=0.5$ W/m°C and $\rho c=3.4 \times 10^6$ J/m³°C as was the perfusion $w_b=1.5$ kg/m³s. The surface of tissue in contact with the transducer face was maintained at $T_s=35^\circ\text{C}$ and the ambient initial temperature was $T_\infty=37^\circ\text{C}$. All other surfaces of the cylinder were maintained at an ambient temperature of $T_\infty=37^\circ\text{C}$. A FEM solution to the problem was generated using axially graded four node quadrilateral elements.

7.3.6 Comparison with experimental results

The transient solution can be compared with our experimental results (Akyürekli *et al.* 1985c, Chapter 6) provided tissue blood flow remains approximately constant during heating. During hyperthermia tissue blood flow peaks approximately 15-30 minutes into treatment, after which flow gradually decreases from an average of 43 to 34 ml/min/100g. During this period the tissue blood flow may be assumed to be

in a quasi-steady state and the theoretical results may be compared to the experimental values. This is done in figure 7.9 in the case of a thermal sensor located at a depth of 1.5 cm beneath the tissue surface. The difference between the theoretical and experimental temperature histories is shown in figure 7.10. The model reproduces the major features of the temperature-time profile, however, during initial heating the model overestimates the temperature while during equilibrium phase, slightly underestimates the temperature.

7.3.7 Comparison with the two dimensional finite element model

Figure 7.11 compares the one dimensional analytic solution to the two dimensional FEM solution of the two dimensional heat transfer problem. The solution represents the axial temperature distribution in the cylinder. There appears to be excellent agreement between the theoretical and the FEM solution. The analytic model reproduces the temperature profile generated by the FEM with only minor underestimates of the temperature. It should be noted that both solutions are for conditions of homogenous perfusion and at $t = 3600s$.

7.4 Discussion

The non-homogeneous, anisotropic and generally highly complex composition, structure and function of mammalian tissues make them among the most difficult of mediums to describe analytically. A successful estimation of the temperature field requires a thermal model which incorporates all relevant heat transfer mechanisms. In this study the Pennes bioheat equation for a semi-infinite volume of tissue irradiated by microwave radiation was used to study the effects surface heat transfer and thermal convection due to blood flow on the tissue thermal profile. A steady-state solution exists only because blood flow and convective heat transfer from the tissue surface both act as heat sinks, cooling the tissue.

During hyperthermia, muscle blood flow peaks approximately 15-30 minutes into treatment (Akyürekli *et al.* 1985c, Chapter 6), after which flow gradually decreases from an average of 43 to 34 ml/min/100g. It is during this period that tissue blood flow may be assumed to be in a quasi-steady state and the theoretical

results may be compared to the *in-vivo* temperature profiles. The tissue model used in this study is a simplified one. Furthermore, the bioheat equation used in these simulations is an approximation and the results are at best qualitative. In spite of this, the model gives some insight into the effect of blood perfusion and surface cooling. The results of this study have shown that the role of blood flow and surface cooling in hyperthermia are critical. Blood flow increases the time taken to reach a desired steady-state temperature and places limits on the temperature achieved for a given input power. The surface temperature and heat transfer coefficient (boundary conditions) also have a strong influence on thermal gradients near the surface of the tissue. The greater the heat transfer coefficient the steeper the thermal gradient near the surface.

Given the complexity of blood flow patterns, it is surprising that the simple linear term introduced by Pennes (1948) has been as good a predictor as it has been shown to be. For example, it has been used by Pennes to accurately predict the one dimensional temperature profiles in the resting forearm, and in a more complex extremity study it has been used to simultaneously and accurately predict both the two dimensional profiles and the transient perfusion as measured at one location (Emery and Sekins 1982, Sekins *et al.* 1982). And as seen in figure 7.9 the model reproduced the major features of the experimental temperature-time profile, although, during initial heating the model overestimates the temperature while during equilibrium phase, it slightly underestimates the temperature. It is surprising however, that the simple analytic model was able to reproduce the temperature distribution calculated using the FEM with such good agreement. Although it is known that the two dimensional approximations of three dimensional heat transfer problems can lead to significant errors (Chen *et al.* 1990) a one dimensional approximation of a two dimensional problem appears to give very reasonable estimates. The analytic model reproduces the temperature profile generated by the FEM with only minor underestimates of the temperature. Under conditions which lead to steep temperature gradients in the radial direction the accuracy of the one dimensional model will be unacceptable. Furthermore, for two dimensional

geometry's with non-uniform blood perfusion the one dimensional approximation will give the wrong temperature profile.

One of the most difficult problems of estimating the temperature field in tissue is the assessment of the effects of blood perfusion. In the model presented here the collective effect of blood flow in the microvasculature is accounted for by a term based on Fick's principle. This principle when applied to biological systems, can be stated as follows, "the amount of substance taken up by an organ or tissue per unit time is equal to the arterial level of the substance minus the venous level times the rate of blood flow". First suggested in this connection by Pennes (1948), the amount of thermal energy convected by the blood stream from some control volume is given by the term $w_b c_b [T - T_\infty]$ (see equation 7.1), where perfusion is in general a function of the tissue temperature and type.

The validity of the convective term due to blood flow is not general. It may provide a good approximation for the amount of heat transported by the blood perfusing the tissue when most of the flow is through the capillary vessels or maybe even through the venules. For the larger blood vessels the assumption of local thermal equilibrium gets progressively worse as the size and velocities increase. Chen and Holmes (1980) performed an analysis showing that the real difficulties arise from the intermediate size vessels, since the largest blood vessels have relatively little heat transfer with the surrounding tissue. The physical interpretation of the blood convective term is as follows : Blood is supplied to tissues from the arterial side from a major vessel which gradually branches off into the smaller vessels leading to the capillaries. The arterial blood, at some known temperature, flows through the capillary bed where complete thermal equilibrium between blood and tissue is attained. Consequently, the blood stream either warms or cools during its passage through the tissue as a function of the temperature difference prevalent at the specific tissue under consideration. Upon draining into the venous side, the temperature of the blood, at each point in the tissue where the drainage occurs is equal to the temperature of the tissue.

The blood perfusion term in the Pennes equation is not capable in describing the dynamics of the vascular response of tissue to thermal stress, since the vascular response of tissue to heat is, in general, position, temperature and tissue dependent (Akyürekli *et al.* 1995c, Chapter 6). Hence the important step in being able to model heat transport in tissue is to incorporate the vascular and spatial response of tissue during hyperthermia.

Most realistic hyperthermia problems require numerical solutions. A widely used technique in solving heat transfer problems is the method of finite element analysis. There are a number of attributes of the finite element method (FEM) which make it convenient for modeling hyperthermia treatments. Perhaps the most important, from the point of view of calculating the temperature fields within a tissue system, is that the method takes into consideration the heterogeneous tissue properties and irregular boundaries as a fundamental part of the mathematical formulation. In some approaches, (e.g. finite difference methods) these problems require special attention. In addition, the FEM facilitates putting fine detail (small elements) in regions where there may be high thermal gradients and gross detail (large elements) in other regions. This feature conserves CPU time.

In conclusion, a major problem limiting the progress in hyperthermia treatment planning is the lack of adequate data to input into the simulations. Clearly blood flow is the most important physiological parameter. Accurate modeling of hyperthermia treatments will require employing three dimensional numerical models. However, as we have demonstrated, for some tissue sites where geometric simplification can be employed and where the power density distribution is easily modeled then the one dimensional Pennes model can provide a relatively accurate prediction of the temperature profile.

Figure 7.1 Transverse section through the microwave applicator and muscle showing the coordinate system used to describe the thermocouples orientation and position during hyperthermia treatment.

Tissue	Thermal conductivity ($\text{kg/m}^\circ\text{C}$)	Density ρ (kg/m^3)	Specific heat $C_p(\text{J/kg}^\circ\text{C})$
Muscle	0.64	1080	3800

Table 7.1 Thermophysical properties of muscle used in calculations of tissue temperature (*Chen et al. 1990*).

Tissue	Baseline blood flow (ml/min/100g)	Blood flow during hyperthermia (ml/min/100g)	Baseline flow ($\text{kg/m}^3\text{s}$)	Blood flow ($\text{kg/m}^3\text{s}$)
Muscle	10	50	16.7	83.5

Table 7.2 Experimentally obtained muscle blood flow values both baseline and during hyperthermia treatment.

Figure 7.2 The variation in the steady-state muscle temperature distribution with perfusion (w kg/m³s) prior to applying power ($q=0$) and were $h=500$ (W/m²°C).

Figure 7.3 The applied power density as given by equation 7.2 where the net forward power is 36 Watts.

Figure 7.4 Solution to equation 7.8 (i.e. eqn. 7.9) illustrating the effect of perfusion on the steady-state temperature distribution in microwave irradiated muscle when the heat transfer coefficient is $500 \text{ W/m}^2\text{°C}$.

Figure 7.5 Solution to equation 7.8 depicting the effect of changes in the heat transfer coefficient on the steady-state muscle temperature distribution when perfusion is constant at 10 ml/min/100g.

Figure 7.6 Solution to equation 7.8 describing the variation in the steady-state muscle temperature distribution with the heat transfer coefficient when perfusion is constant at 40 ml/min/100g.

Figure 7.7 Plot of equation 7.12 describing the variation in the muscle temperature profile with time when perfusion is constant at $w=10\text{ml/min}/100\text{g}$, and the heat transfer coefficient is $h=500\text{ W/m}^2\text{°C}$.

Figure 7.8 Plot of equation 7.11 describing the variation in the muscle temperature profile with time when perfusion is constant at $w=40\text{ml/min/100g}$, and the heat transfer coefficient is $h=500\text{ W/m}^2\text{°C}$.

Figure 7.9 Comparison of the transient solution and measured temperature histories in muscle for a thermocouple 1.5 cm beneath the tissue surface.

Figure 7.10 The difference between the transient solution and measured temperature history for a thermocouple 1.5 cm beneath the muscle surface.

Figure 7.11 Comparison of the 1 dimensional transient analytic solution to a two dimensional transient finite element solution (Akyürekli 1988).

Chapter 8

Conclusion

Blood flow assumes a critical role in determining the effectiveness of adjuvant hyperthermia, since it is the primary mechanism of heat dissipation in tissue and can significantly influence the tissue thermal distribution during treatment. Consequently an understanding of the thermally induced vascular changes during hyperthermia is important for the effective use of this modality. The goal of this study was to investigate the changes in blood flow in various porcine organs and tissues before, during and following hyperthermia treatment. The organs and tissues studied included the porcine kidney, brain and muscle.

In order to temporally isolate the changes in perfusion the radioactive microsphere reference sample technique was employed to sequentially measure the local and regional blood flow distribution under a variety of hemodynamic conditions. In this study a tissue preparation technique was developed to provide the consistent, reproducible counting geometry necessary when planar Ge(Li) detectors are used to make multiple blood flow measurements. It was demonstrated that this technique gave results which were consistent with an independent measurement technique, a calibrated ultrasound blood flow probe.

The vasculature of the organs and tissues investigated in this study have striking differences, which are closely related to their function. Consequently, it is not surprising that their vascular responses to hyperthermia are also different. The classical description of tissue response to heating has been an increase in blood flow. Our measurements have revealed that this is not always the case. Elevated temperatures in the autoperfused kidney resulted in reduced blood flow to both the

heated and contralateral control kidneys. Furthermore, the reduction in blood flow to the heated kidney was significantly greater than that observed in the contralateral control kidney or in the normothermic control group. In the autoperfused kidney the mechanism responsible for the decrease in blood flow appeared to be specific to the heated organ. The decrease in blood flow may be the result of vascular damage or collapse caused during heating. Moreover, ischemic damage incurred prior to reperfusion may have made the kidney more susceptible to thermal damage.

In contrast to the autoperfused kidney was the vascular response of the *in-situ* kidney, which showed a reduction in total blood flow to both kidneys during heating. When the blood flow distribution in the *in-situ* heated kidney and its respective contralateral control kidney are compared, the indication is that hyperthermia affects the heated kidney differently. More importantly the response of the kidney does not appear to be uniform. Since heating was primarily concentrated to the central region of the kidney the greater reduction in flow observed near this region than at the poles may indicate vascular damage. This is supported by the fact that a similar blood flow distribution was not observed in the contralateral kidney during heating. This is in contrast to the response observed in the heated autoperfused kidney, in which the superior and inferior poles of the kidney experienced the greater reduction in blood flow. It was speculated that hyperthermia had a greater effect on cortical than medullary blood flow. The exact mechanisms responsible for the reduced renal blood flow are not clear, but it is evident that in addition to systemic effects there may also be local factors acting on the heated kidney. Since heating the kidney can result in reduced blood flow, clinical heating for cancer therapy may be contraindicated. Although heating the kidney would be difficult, mainly due to the organ's large blood flow compared to surrounding tissues, caution should be observed when treating tissues in the immediate vicinity of the kidney in order to avoid possible renal damage.

A distinctive difference was observed between the thermal vascular response of the porcine kidney and brain. In control studies cerebral blood flow showed a continual decrease following the insertion of the interstitial antenna and thermal

probes such that after one hour the mean regional cerebral blood flow was 80% of the pre-insertion value. In contrast the insertion of thermocouple arrays into the renal parenchyma did not result in any significant change in blood flow.

Interstitial heating of the porcine brain to temperatures in the range of 42.6 - 43.8°C resulted in a moderate, but gradual increase in blood flow over the entire heating period. As a consequence, blood flow at the end of the heating period was approximately 32% greater than the blood flow prior to inserting the antenna and thermal probes. Further, this blood flow was nearly 70% greater than the blood flow measured one hour after insertion of the antenna and thermal probes in the control sham heated animals. In comparison, cerebral blood flow in animals heated to temperatures >46°C, showed the same initial increase, but with further heating blood flow decreased, probably the result of primary vascular damage (i.e. intravascular coagulation and endothelial swelling). Following cessation of heating the mean regional cerebral blood flow was approximately 18% below the value prior to the insertion of the antenna and thermal probes, but still approximately 7% greater than the blood flow measured one hour after the insertion of the antenna and thermal probes in the control sham heated animals. Since these were acute experiments it is not possible to predict whether this apparent vascular damage was transient, permanent or progressive. Our study indicates that cerebral blood flow is sensitive to both mechanical and thermal intervention, therefore in clinical applications minimally invasive techniques are required to guarantee cerebral blood flow integrity and care must be taken to insure that the temperature of normal tissues does not exceed some critical value such as the 46°C observed in porcine brain.

Regional heating of muscle generated a non-uniform blood flow distribution which was a function of the tissue temperature distribution. Blood flow was greater in those tissue samples in which higher temperatures were recorded and less in those sampled further from the peak SAR. Our results indicate that the increase in blood flow during hyperthermia is primarily a regional phenomena and does not result from any significant increase in systemic blood flow. Although blood flow in

muscle may be considered to be uniform prior to heating, this assumption does not hold during hyperthermia treatment. Therefore, the non-uniform nature of the blood distribution during heating should be incorporated into any practical bioheat transfer model.

In this study the one dimensional bioheat equation was used to model the effect of blood flow and surface heat transfer on the temperature distribution in muscle. A steady-state solution existed only because blood flow and convective heat transfer from the tissue surface both acted as heat sinks, cooling the tissue. The most difficult problem in estimating the temperature field in tissue is the assessment of the effects of blood perfusion. In the model presented in this study the collective effect of blood flow in the microvasculature is accounted for by a perfusion term which is assumed constant. In general the perfusion term will be a function of the tissue temperature and type, hence the bioheat equation will be non-linear. Although the standard perfusion term provides a good approximation for the amount of heat transported by the blood its validity is not general. For the larger blood vessels the assumption of local thermal equilibrium gets progressively worse as the size and velocities increase.

Given the complexity of blood flow patterns in general, it is surprising that the simple linear term introduced by Pennes is as good a predictor as it has shown to be. In this study the model reproduced the major experimentally measured features of the temperature-time profile, although, during initial heating the model overestimates the temperature while during equilibrium phase, it slightly underestimates the temperature. The blood perfusion term in the Pennes equation is not capable of describing the dynamics of the vascular response of tissue to thermal stress, since the vascular response of tissue to heat is in general temperature and tissue dependent. Hence the important step in being able to model heat transport in tissue is to determine the vascular changes during hyperthermia and incorporate them within a realistic heat transfer model.

The non-homogeneous, anisotropic and generally highly complex composition, structure and function of mammalian tissues make them among the most difficult of mediums to describe analytically. A successful estimation of the temperature field requires a thermal model which incorporates all relevant heat transfer mechanisms. Therefore it is vital that the vascular response of the tissue within the treatment field be incorporated into the treatment plan. It cannot be assumed that the normal tissue adjacent to the tumor will have a constant uniform blood flow distribution during treatment, as demonstrated in this work. Also, it may not be assumed that all tissues or organs will respond with an increase in blood flow. The vascular response during hyperthermia is tissue specific, therefore the thermal model must also be tissue specific. In conclusion, it is evident that the vascular response during hyperthermia is not only temperature and time dependent, but also spatial and tissue dependent. Subsequently the effective clinical use of adjuvant hyperthermia will depend on the careful application of these biological principles from experimental work.

It is hoped that in the future the data presented in this work will be used to establish a dynamic perfusion term and hence a more practical model of bioheat transfer. Moreover, predictive models used to simulate the tissue temperature distribution must incorporate the tissue specific as well as non-uniform nature of the vascular response during hyperthermia. Although this study examined the changes in the spatial distribution of blood flow in various tissues, it is important that future studies investigate these changes with greater spatial resolution; this implies sampling smaller volumes of tissue. Furthermore, greater insight into the mechanisms responsible for the changes in blood flow are necessary. Histological examination and correlation to the tissue sample temperature and perfusion in addition to the effect of vasodilator mechanisms need to be investigated. Finally, tissues and organs of clinical importance such as the lung, spin and prostate as well as neoplastic tissues need to be investigated, since the clinical effectiveness of hyperthermia may depend on this information.

Appendix A

The one dimensional form of the heat equation first proposed by Pennes (1948) and the inherent assumptions in this equation are outlined by Hodson *et al.* (1979). The differential equation describing the temperature distribution in an semi-infinite slab of tissue is given by equation 7.3 and is as follows

$$\rho \frac{\partial T}{\partial t} = \frac{\partial^2 T}{\partial x^2} - \xi T + q \quad (1A)$$

Steady-state solution without heating

For the particular case $S(x,t) = 0$ and $\frac{\partial T}{\partial t} = 0$ the bioheat equation (1A) is written as follows

$$\frac{d^2 T}{dx^2} - \xi T = 0 \quad \forall t \geq 0, x \geq 0 \quad (2A)$$

and $\xi = \frac{w_b c_b}{k}$ with boundary conditions given as

$$T(x)|_{x=\infty} = 0 \quad \frac{\partial T}{\partial x} \Big|_{x=0} = \frac{h}{k} (T - T_\infty - T_a) \quad \forall t \geq 0 \quad (3A)$$

Let $\frac{d}{dx} = D$ then equation 2A can be rewritten as

$$f(D) = (D^2 - \xi)T = 0 \quad (4A)$$

The auxiliary equation $m^2 - \xi = 0$ has roots $m = \pm\sqrt{\xi}$, hence the general solution to equation 4A may be written as

$$T(x) = c_1 e^{-x\sqrt{\xi}} + c_2 e^{x\sqrt{\xi}} \quad (5A)$$

Since $T(x) \rightarrow 0$ as $x \rightarrow \infty$, then $c_2 = 0$. The solution 4A then reduces to

$$T(x) = c_1 e^{-x\sqrt{\xi}} \quad (7.6A)$$

Applying the second boundary condition the second coefficient is given as

$$c_1 = \frac{h(T_s - T_\infty)}{h + k\sqrt{\xi}} \quad (7A)$$

Therefore the solution to equation 2A is given as

$$T(x) = \frac{h(T_s - T_\infty)}{h + k\sqrt{\xi}} e^{-x\sqrt{\xi}} \quad (8A)$$

Steady-state solution with heating

Rewriting equation 1A

$$\frac{d^2 T}{dx^2} - \xi T = -q_0 e^{-x/\delta} \quad (9A)$$

$$q_0 = \frac{(P_i - P_r)}{ab\delta k} \quad (10A)$$

were P_i and P_r are the incident and reflected power in watts, a and b are the wave guide aperture dimensions, δ the microwave penetration depth and k the thermal conductivity of the tissue. Letting $D^2 = \frac{d^2}{dx^2}$, equation 9A can be rewritten as

$$f(D) = (D^2 - \xi)T = -q_0 e^{-x/\delta} \quad (11A)$$

The general solution of this nonhomogeneous equation is the sum of the complementary and particular solutions, i.e.

$$T = T_c + T_p \quad (12A)$$

The solution to the homogeneous equation $(D^2 - \xi)T_c = 0$ is given by equation 5A. The right hand side of equation 11A $R(x) = -q_0 e^{-x/\delta}$ is a particular solution of a homogeneous linear differential equation whose auxiliary function $f(m')$ has the root $m' = -1/\delta$. Therefore the function $R(x)$ is a particular solution of the equation

$$\left(D + \frac{1}{\delta}\right)R(x) = 0 \quad (13A)$$

The operator $(D + 1/\delta)$ will annihilate the right side of equation 11A. Therefore applying the operator to equation 11A

$$\left(D + \frac{1}{\delta}\right)(D^2 - \xi)T = 0 \quad (14A)$$

Any solution of 9A must be a particular solution of 14A. The general solution of 14A can be written from the roots of it's auxiliary equation, i.e. $m' = -1/\delta$ and $m = \pm\sqrt{\xi}$. Hence the general solution is given as

$$T(x) = c_1 e^{-x\sqrt{\xi}} + c_2 e^{x\sqrt{\xi}} + c_3 e^{-x/\delta} \quad (15A)$$

where the particular solution is $T_p = c_3 e^{-x/\delta}$. By applying the boundary conditions (3A) the constants are determined to be

$$c_1 = - \left(\frac{Q_0}{\xi - \frac{l}{\delta^2}} \right) \left(\frac{\frac{l}{\delta} + \frac{h}{k}}{\frac{h}{k} + \sqrt{\xi}} \right) + \frac{\frac{h}{k}(T_s - T_\infty)}{\left(\frac{h}{k} + \sqrt{\xi} \right)} \quad (16A)$$

$$c_2 = 0$$

$$c_3 = \frac{S_0}{\xi - \frac{l}{\delta^2}}$$

Hence the solution is given as follows

$$T(x, \infty) = \frac{q_0}{\xi - l/\delta^2} \left\{ e^{-x/\delta} - \frac{l/\delta + h/k}{h/k + \sqrt{\xi}} e^{-x/\sqrt{\xi}} \right\} + \frac{(h/k)(T_s - T_\infty)}{h/k + \sqrt{\xi}} e^{-x/\sqrt{\xi}} \quad (17A)$$

Transient solution with heating

The Fourier transform with respect to time is given as

$$F[T(x, t)] = \frac{l}{\sqrt{2\pi}} \int_{-\infty}^{+\infty} T(x, t) e^{-i\omega t} dt \quad (18A)$$

$$F[T(x, t)] = g(x, i\omega) \quad (19A)$$

where $T=T(x,t)$ is the temperature

$$F\left[\frac{\partial T(x,t)}{\partial t}\right] = \frac{1}{\sqrt{2\omega}} \int_{-\infty}^{+\infty} \frac{\partial T(x,t)}{\partial t} e^{-i\omega t} dt \quad (20A)$$

$$= \frac{1}{\sqrt{2\pi}} \left\{ T(x,t) e^{-i\omega t} \Big|_{-\infty}^{+\infty} - i\omega \int_{-\infty}^{+\infty} T(x,t) e^{-i\omega t} dt \right\} \quad (21A)$$

$$\Rightarrow F\left[\frac{\partial T(x,t)}{\partial t}\right] = i\omega g(x, i\omega) \quad (22A)$$

$$F\left[\frac{\partial^2 T(x,t)}{\partial x^2}\right] = \frac{1}{\sqrt{2\pi}} \int_{-\infty}^{+\infty} \frac{\partial^2 T(x,t)}{\partial x^2} e^{-i\omega t} dt \quad (23A)$$

$$= \frac{\partial^2}{\partial x^2} \left\{ \frac{1}{\sqrt{2\pi}} \int_{-\infty}^{+\infty} T(x,t) e^{-i\omega t} dt \right\}$$

$$\Rightarrow F\left[\frac{\partial^2 T(x,t)}{\partial x^2}\right] = \frac{\partial^2 g(x, i\omega)}{\partial x^2}$$

Therefore the Fourier transform of equation 1A is given as

$$\frac{d^2 g(x, i\omega)}{dx^2} - \xi^2 g(x, i\omega) + h_0 e^{-x/\delta} = 0 \quad (24A)$$

where $\beta^2 = \xi + i\eta\omega$ and $h_0 = iq_0/\omega\sqrt{2\pi}$. Solution of equation 24A is given by equation

$$g(x, i\omega) = \frac{h_0}{\xi - 1/\delta^2} \left\{ e^{-x/\delta} - \frac{1/\delta + h/k}{h/k + \xi} e^{-x\sqrt{\xi}} \right\} + \frac{(h/k)(T_\infty - T_0)}{h/k + \xi} \quad (25A)$$

using the inverse transforms given by Campbell and Foster (1948) the transient solution is given as follows

$$\begin{aligned}
 T(x, t) = & \frac{q_0}{\xi - 1/\delta^2} \left\{ e^{-x/\delta} + \frac{1/\delta + h/k}{2(\sqrt{\xi} - h/k)} e^{x\sqrt{\xi}} \operatorname{erfc} \left(\frac{x}{2} \sqrt{\beta/t} + \sqrt{\frac{\xi t}{\beta}} \right) \right. \\
 & - \frac{1/\delta + h/k}{2(\sqrt{\xi} + h/k)} e^{-x\sqrt{\xi}} \operatorname{erfc} \left(\frac{x}{2} \sqrt{\beta/t} - \sqrt{\xi t/\beta} \right) - \frac{1}{2} e^{\left(\frac{1}{\beta} (1/\delta^2 - \xi) - x/\delta \right)} \\
 & \bullet \operatorname{erfc} \left(\frac{1}{\delta} \sqrt{t/\beta} - \frac{x}{2} \sqrt{\beta/t} \right) - \frac{1/\delta + h/k}{2(1/\delta - h/k)} \bullet e^{\frac{1}{\delta} (1/\delta^2 - \xi) + x/\delta} \\
 & \bullet \operatorname{erfc} \left(\frac{1}{\delta} \sqrt{t/\beta} + \frac{x}{2} \sqrt{\beta/t} \right) + \frac{\frac{h}{k} (\xi + 1/\delta^2)}{(\xi - (h/k)^2)(1/\delta - h/k)} \bullet e^{(h/k)x + ((h/k)^2 - \xi) \frac{t}{\beta}} \quad (7.26A) \\
 & \left. \bullet \operatorname{erfc} \left(\frac{x}{2} \sqrt{\beta/t} + \frac{h}{k} \sqrt{t/\beta} \right) \right\} + \frac{h(T_i - T_\infty)}{h + k\sqrt{\xi}} e^{-x\sqrt{\xi}}
 \end{aligned}$$

References

ABE, M., HIRAOKA, M., TAKAHASHI, M., EGAWA, S., MATSUDA, C., ONOYAMA, Y., MORITA, K., KAKEHI, M., and SUGAHARA, T., 1986, Multi- institutional studies on hyperthermia using an 8-MHz radiofrequency capacitive heating device Thermotron RF-8 in combination with radiation for cancer therapy. *Cancer*, **58**, 1589-1595.

AKYÜREKLI, D. U., 1988, Simulation of ultrasound hyperthermia using the finite element method. M.Sc thesis, Carleton University, Ottawa, Ontario Canada.

AKYÜREKLI, D. U., GERIG, L. H. AND RAAPHORST, G. P., 1995a, (Chapter 2), A Tissue preparation technique for microsphere assays of blood flow. *Animal Laboratory Science* Vol. 45, No. 4 August.

AKYÜREKLI, D. U., GERIG, L. H. and RAAPHORST, G. P., 1995b, (Chapter 3), Changes in blood flow in a hyperthermic autoperfused kidney. (Submitted for publication)

AKYÜREKLI, D. U., GERIG, L. H., RAAPHORST, G. P., 1985c, (Chapter 6), Porcine muscle blood flow distribution during hyperthermia. (Submitted for publication)

ARCANGELI, G., CIVIDALLI, A., NERVI, C., and CRETON, G., 1983, Tumor control and therapeutic gain with different schedules of combined radiotherapy and local external hyperthermia in human cancer. *International Journal of Radiation Oncology, Biology and Physics*, **9**, 1125-1134.

ARCANGELI, G., GUERRA, A., LOVISOLO, G., CIVIDALLI, A., MARINO, C., and MAURO, F., 1985, Tumor response to heat and radiation: Prognostic variables in the treatment of neck node metastases from head and neck cancer. *International journal of hyperthermia*, **1**, 207-217.

ARCANGELI, G., OVERGAARD, J., GONZALEZ, D. G., and SHRIVASTAVA, P. N., 1988, Hyperthermia trials. *International Journal of Radiation Oncology, Biology and Physics*, **14**, S93-S109.

ARCHIE, J. P., FIXLER, D. E., ULLYOT, D. J., HOFFMAN, J. I. E., UTLEY, J. R., and CARLSON, E. L., 1973, Measurement of cardiac output with and organ trapping of radioactive microspheres. *J. Appl. Physiol.*, **35**, 148-154

AUSTIN, R. E., HAUCK, W. W., ALDEA, G. S., FLYNN, A. E., COGGINS, D. L., and HOFFMAN, J. I. E., 1989, Quantitating error in blood flow measurements with radioactive microspheres. *Am. J. Physiol.*, **257**, H280-H288.

BARCROFT, H and EDHOM, O. G., 1946, Temperature and blood flow in the human forearm. *Journal of Physiology*, **104**, 366-376.

BASSINGTHWAIGHTE, J. B., MALONE, M. A., MOFFETT, T. C., KING, R. B., LITTLE, S. E., LINK, J. M., and KROHN, K. A., 1987, Validity of microsphere depositions for regional myocardial flows. *Am. J. Physiol.*, **253**, H184-H193.

BERNSTEIN, M., and GUTIN, P. H., 1981, Interstitial irradiation of brain tumors: A review. *Neurosurgery*, **9**, 91-99.

BOWMAN, H. F., 1982, Heat transfer mechanisms and thermal dosimetry. *National Cancer Institute Monograph*, **61**, 437-445.

BOWMAN, H. F., 1981, Heat transfer and thermal dosimetry. *Journal of Microwave Power*, **16**, pp. 121-133

BRACCO, M., and CURTI, P. C., 1954, The vasoconstrictor factor of platelets. *Experientia*, **10**, 71-72.

BRENNAN, K. M., ROOS, M. S., BUDINGER, T. F., HIGGINS, R. J., WONG, S. T. M., and BRISTOL, K. S., 1993, A study of radiation necrosis and edema in the canine brain using positron emission tomography and magnetic resonance imaging. *Radiation Research*, **134**, 43-53.

BRICKFORD, R. G. , and WINTON, F. R., 1937, The influence of temperature on the isolated kidney. *Journal of Physiology*, **89**, 198.

BUCKBERG, G. D., LUCK, J. C., PAYNE, D. B., HOFFMAN, J. I. E., ARCHIE, J. P., and FIXLER, D. E., 1971, Some sources of error in measuring regional blood flow with radioactive microspheres. *J. Appl. Physiol.*, **31**, 598-604.

BURTON, M. A., KELLEHER, D. K., CODDE, J. P. and GRAY, B. N., 1991, Changes in hepatic blood flow during regional hyperthermia. *International Journal of Hyperthermia*, **7**, 271-277.

CAMPBELL, G. A., and FOSTER, R. M., 1948, *Fourier Integrals for Practical Applications*. Van Nostrand, New York.

CHEN, M. M., AND HOLMES, K. R., 1979, Thermal characteristics of tumors. Conference Proceedings of the New York Academy of Science and National Cancer Institute.

CHEN, M. M., AND HOLMES, K. R., 1980, Microvascular contributions in tissue heat transfer. *Annals of the New York Academic Sciences*, **335**, 137-150.

CHEN, Z. P., MILLER, W. H., ROEMER, R. B., CETAS, T. C., 1990, Errors between two- and three-dimensional thermal model predictions of hyperthermia treatments. *International Journal of Hyperthermia*, **6**, 175-191.

CHOURAIR, A. K., LEVIN, V. A., and GUTIN, P. H., 1986, Development of multiple lesions during radiation therapy and chemotherapy in patients with gliomas. *Journal of Neurosurgery*, **65**, 654-658.

CREZEE, J., and LAGENDIJK, J. J. W., 1990, Experimental verification of bioheat transfer theories: measurement of temperature profiles around large artificial vessels in perfused tissue. *Phys. Med. Biol.*, **35**, 905-923.

DEBERTIN, K., and HELMER, R. G., 1988, Gamma and X-ray spectrometry with semiconductor detectors. North-Holland, New York.

DEWEY, W. C., FREEMAN, M. L., RAAPHORST, G. P., CLARK, E. P., WONG, R. S. L., HIGHFIELD, D. P., SPIRO, I. J., TAMOSOVIC, S. P., DENMAN, D. L., and CROSS, R. S., 1980, Cell biology of hyperthermia and radiation. Radiation Biology and Cancer Research, Edited by R. E. Meyn and H. R. Whithers (New York: Raven Press), pp. 589-612.

DEWEY, W. C., and HOLAHAN, E. V., 1984, Hyperthermia: Basic biology: *Brain Tumor Therapy, Progress in Experimental Tumor Research*, Vol. 28, M. L. Rosenblum and C. B. Wilson editors, Basel, Karger. pp. 198-219.

DICKSON, J. A., and CALDERWOOD, S. K., 1980, Temperature range and selective sensitivity of tumors to hyperthermia: A critical review. *Annals of the New York Academy of Sciences*, **335**, 180-205.

DOLE, W. P., JACKSON, D. L., ROSENBLATT, J. I., and THOMPSON, W. L., 1982, Relative error and variability in blood flow measurements with radiolabeled microspheres. *Am. J. Physiol.*, **243**, H371-H378.

EMERY, A. F., DELATEUR, B. J., and NELP, W. B., 1984, Local muscle blood flow and temperature responses to 915 MHz diathermy as simultaneously measured

and numerical predicted. *Archives of Physical Medicine and Rehabilitation*, **65**, 1-7.

EMERY, A. F., and SEKINS, K. M., 1982, *International Journal of Heat and Mass Transfer*, **25**, 823-834.

ENDRICH, B., 1988, Hyperthermia and microcirculatory effects of heat in animal tumors. *Recent Results in Cancer Research*, **109**, 96-108.

ENDRICH, B., HAMMERSEN, F., and MESSMER, K., 1988, Hyperthermia induced changes in tumor microcirculation. *Recent Results in Cancer Research*, **107**, 44-59.

FIELD, S. B., 1979, Hyperthermia in the treatment of cancer. *Cancer Treatment Review*, **6**, 63-94.

FIKE, J. R., GOBBEL, G. T., SATOH, T., and STAUFFER, P. R., 1991, Normal brain response after interstitial microwave hyperthermia. *International Journal of Hyperthermia*, **7**, 795-808.

GABBIANI, G., and BADONNEL, M. C., 1975, Early changes of endothelial clefts after thermal injury. *Microvascular Research*, **10**, 65-75.

GARDNER, R. P., VERGHESE, K., and LEE, H., 1980, The average solid angle subtended by a circular detector coaxial to a circular isotropic source. *Nuclear Inst. Meth.*, **176**, 615-617.

GERWECK, L. E., 1985, Environmental and vascular effect. *Hyperthermic Oncology 1984*, vol. 2, edited by J. Overgaard (London: Taylor & Francis), pp. 253-262.

GELMAN, S., 1987, Anesthesia and organ transplantation. Suannders Company, 78-100.

GERIG, L. H., DANJOUX, C. E., RAAPHORST, G. P., HAUDERGICZ, Z. and HOLFORD, S., 1988, A computerized thermocouple-based clinical thermometry system. *Endocuriether. Hypertherm. Oncol.*, **4**, 31-37.

GERIG, L. H., SZANTO, J. and RAAPHORST, G. P., 1992, On the spatial resolution of clinical thermometers. *Medical Physics*, **19** (3), 679-684.

GERWECK, L. E., 1985, Environmental and vascular effects. *Hyperthermic Oncology 1984*, vol. 2, edited by J. Overgaard (London: Taylor & Francis), pp. 253-262.

GILLENWATER, J. Y., FROHLICH, E. D., and KELLER, A. D., 1964, Differential effects of heat on splenic and renal vascular beds. *American Journal of Physiology*, **207**, 133-136

GOLDMAN, H., 1993, Techniques for the measurement of cerebral blood flow. *The regulation of cerebral blood flow*, edited by J. W. Phillis, (Toronto: CRC Press).

GRANGER, H. J., GOODMAN, A. H., and COOK, B. H., 1975, Metabolic models of microcirculation to hyperthermia. *Federation Proceedings*, **34**, 2025-2030.

GREENBLATT, E. P., LOEB, A. L., LONGNECKER, D. E., 1992, Endothelium-dependent circulatory control- A mechanism for the differing peripheral vascular effects of isoflurane versus halothane. *Anesthesiology*, **77**, 1178-1185.

GUTIN, P. H., LEIBEL, S. A., WARA, W. M., CHOUCAIR, A., LEVIN, V. A., PHILLIPS, T. L., SILVER, P., DA SILVA, V., EDWARDS, M. S. B., DAVIS, R. L., WEAVER, K. A., and LAMB, S., 1987, Recurrent malignant gliomas: Survival following interstitial brachytherapy with high-activity iodine-125 sources. *Journal of Neurosurgery*, **67**, 864-873.

GUTIN, P. H., PHILLIPS, T. L., WARA, W. M., LEIBEL, S. A., HOSOBUCHI, Y., LEVIN, V. A., WEAVER, K. A., and LAMB, S., 1984, Brachytherapy of recurrent malignant brain tumors with removable high activity iodine-125 sources, *Journal of Neurosurgery*, **60**, 61-68.

GUTIN, P. H., PRADOS, M. D., PHILLIPS, T. L., WARA, W. M., LARSON, D. A., LEIBEL, S. A., SNEED, P. K., LEVIN, V. A., WEAVER, K. A., SILVER, P., LAMBORN, K., LAMB, S., and HAM, B., 1991, External irradiation followed by an interstitial high activity iodine- 125 implant boost in the initial treatment of malignant gliomas. NCOG study 6G-82-2. *International Journal of Radiation Oncology Biology and Physics*, **21**, 601-606.

GUY, A. W., LEHMANN, J. F., and STONEBRIDGE, J. B., 1974, Therapeutic applications of electromagnetic power. *IEEE Transactions on Biomedical Engineering*, **62**, 55-75.

GUY, A. W., 1978, Development of a 915 MHz direct contact applicator for therapeutic heating of tissues. *IEEE Transactions on Microwave Theory and Techniques*, **26**, 550-556.

HAHN, G. M., 1979, Potential for therapy of drugs and hyperthermia. *Cancer Research*, **39**, 2264-2268.

HAHN, G. M., 1982, *Hyperthermia and Cancer*, (New York: Plenum Press).

HALL, E. J., 1978, *Radiobiology for the Radiologist* (New York: Harper & Row).

HANDLEY, C., and MOYER, J., 1954, Unilateral renal adrenergic blockade and renal response to vasopressor agents and to hemorrhage. *J. Pharmacol. Exp. Ther.*, **112**, 1-7.

HARVEY, H. B., 1959, Effect of temperature on the function of the isolated kidney. *American Journal of Physiology*, **197**, 181.

HEYMANN, M. A., PAYNE, D. B., HOFFMAN, J. I. E., and RUDOLPH, A. M., 1977, Blood flow measurements with radionuclide-labeled particles. *Prog. Cardiovasc. Dis.*, **20**, 55-79.

HOCHBERG, F. H., and PRUITT, A., 1980, Assumptions in the radiotherapy of glioblastoma. *Neurology*, **30**, 907-911.

HODSON, D. A., EASON, G., and BARBENEL, J. C., 1979, Modeling transient heat transfer through the skin and superficial tissues: surface insulation. Transactions of the American Society of Mechanical Engineering, *Journal of Biomechanical Engineering*, **108**, pp. 183.

HOLMES, K. R., RYAN, W., WEINSTEIN, P. and CHEN, M. M., 1984, A fixation technique for organs to be used as perfused tissue phantoms in bioheat transfer studies. *Advances in Bioengineering*, Ed R. L., Spiker, New York: ASME, 9-10.

HUGANDER, A., BOLMSIO, M., HAFSTROM, L., and PERSSON, B., 1983 Liver blood flow studies during local hyperthermia. An experimental study in rats. *Clinical Oncology*, **9** 303-310.

JAIN, R. K., 1980, Temperature distributions in normal and neoplastic tissues during normothermia and hyperthermia. *Annals of the New York Academy of Sciences*, **334**, 48-64.

JAIN, R. K., and WARD-HARTLEY, K., 1984, Tumor blood flow: characterization, modifications and role in hyperthermia. *IEEE Transactions in Sonics and Ultrasonics*, **SU-31**, 504-526.

JOHNSON, J. M., BRENGELMAN, G. L., and ROWELL, L. B., 1976 Interaction between local and reflex influence on the human forearm skin blood flow. *Journal of Applied Physiology*, **41**, 826-831.

KATZ, M. A., and BLANTZ, R. C., 1972, Geometric error in tissue gamma-counting: methods for minimization. *J. Appl. Physiol.*, **32**, 533-534.

KIM, J. H., HAHN, E. W., and AHMED, S. A., 1982, Combination hyperthermia and radiation therapy for malignant melanoma. *Cancer*, **50**, 478-482.

LAGENDIJK., 1987, Heat transfer in tissues. In; *Physics and Technology of Hyperthermia*, edited by S. B. Field and C. Franconi (Martinus Nijhoff, Dordrecht, The Netherland), pp.517-552.

LAGENDIJK, J. J. W., and MOOIBROEK, J., 1986, Hyperthermia treatment planning. *Recent Results in Cancer Research*, **101**, 119-131.

LEIBEL, S. A., and SHELINE, G. E., 1987, Radiation therapy for neoplasms of the brain. *Journal of Neurosurgery*, **66**, 1-22.

LEVIN, V. A., and KABRA, P., Effectiveness of the nitrosourea as a function of their lipid solubility in chemotherapy of experimental rat brain tumors. *Cancer Chemotherapy Reports, Part 1*, **58**, 787-792.

LILLY, M. B., NG, T. C., EVANOCHKO, W. T., KATHOLI, C. R., KUMAR, N. G., ELGAVISH, G. A., DURANT, J. R., HIRAMOTO, R., GHANTA, V., and GLICKSON, J. D., 1984, Loss of high-energy phosphate following hyperthermia demonstrated by in vivo ^{31}P -nuclear magnetic resonance spectroscopy. *Cancer Research*, **44**, 633-638.

LINDHOLM, C. E., KJELLEN, E., NILSSON, P., and HERTZMAN, S., 1987, Microwave induced hyperthermia and radiotherapy in human superficial tumors: Clinical results with a comparative study of combined treatment versus radiotherapy alone. *International Journal of Hyperthermia*, **5**, 393-411.

LOKSHINA, A. M., SONG, C. W., RHEE, J. G. and LEVITT, S. H., 1985, Effect of fractionated heating on the blood flow in normal tissues. *International Journal of Hyperthermia*, **1**, 117-129.

LUK, K. H., FRANCIS, M. E., PEREZ, C. A., and JOHNSON, R. J., 1984, Combined radiation and hyperthermia: Comparison of two treatment schedules based on data from registry established by the radiation therapy oncology group (RTOG). *International Journal of Radiation Oncology, Biology, and Physics*, **10**, 801-809.

LUSTINEC, K., 1965, Vasoconstrictor properties of blood in perfusion experiments. *Physiol. Bohemoslov*, **14**, 583-593.

LYONS, B. E., SAMULSKI, T. V., COX, R. S., and FESSENDEN, P., 1989, Heat loss and blood flow during hyperthermia in normal canine brain I: Empirical study and analysis. *International Journal of Hyperthermia*, **5**, 225-247.

MAKOWSKI, E. L., MESCHIA, G., and DROEGEMUELLER, W., 1968, Measurement of umbilical arterial blood flow to the sheep placenta and fetus in utero. *Circ Res.*, **23**, 623-631.

MARION, J. B., and HEALD, M. A., 1980, *Classical electromagnetic radiation*, (Academic Press, New York).

MARMOR, J. B., 1979, Interactions of hyperthermia and chemotherapy in animals. *Cancer Research*, **39**, 2269-2276.

MERCER, P., 1971, Effect of bilateral vagotomy on renal function in the rat. *Canadian Journal of Physiology and Pharmacology*, **49**, 420-426.

MERNAGH, J. R., SPIERS, E. W., and ADISESHIAH, M., 1976, The measurement of radioactive microspheres in biological samples. *Phys. Med. Biol.*, **21**, 646-650.

MILLIGAN, A. J., 1987, Canine muscle blood flow during fractionated hyperthermia. *International Journal of Hyperthermia*, **3**, 353-359.

MILLIGAN, A. J., and PANJEHPOUR, M., 1985, Canine normal and tumor tissue estimated blood flow during fractionated hyperthermia. *International Journal of Radiation Oncology, Biology and Physics*, **11**, 1879-1884.

MILLIGAN, A. J., CONRAN, P.B., ROPAR, M. A., MCCULLOCH, H. A., AHUJA, R. K., and DOBELBOWERR, R., 1983, Predictions of blood flow from thermal clearance during regional hyperthermia. *International Journal of Radiation Oncology, Biology and Physics*, **9**, 1335-1343.

MOOIBROEK, J., and LAGENDIJK, J. J. W., 1991, A fast and simple algorithm for the calculation of convective heat transfer by large vessels in three-dimensional inhomogeneous tissues. *IEEE Transactions on Biomedical Engineering*, **38**, 490-501.

NAGASAKA, T., HIRATA, K., NUNOMURA, T. and CABANAC, M., 1987, The effects of local heating on blood flow in the forearm skin. *Canadian Journal of Physiology and Pharmacology*, **65**, 1329-1332.

NAKAJIMA, T., RHEE, J. G., SONG, C. W. and ONOYAMA, Y., 1992, Effect of a second heating on rat liver blood flow. *International Journal of Hyperthermia*, **8**, 679-687.

NAKAJIMA, T., OSBORN, J., RHEE, J. G. and SONG, C. W., 1990, Effect of regional heating of the upper body on the liver blood flow in rats. *International Journal of Hyperthermia*, **6**, 1-14.

NIZET, A. Y., 1967, Control by plasma potassium concentration of sodium excretion isolated perfused dog kidney. *Arch. Ges. Physiol*, **297**, 162-165.

NIZET, A. Y., CUYPERS, L., MASSILLON, L., and LAMBERT, S., 1957, Demonstration of factors reducing blood flow liberated by erythrocytes. *Arch. Intern. Physiol.*, **65**, 568.

NOSE, Y., NAKAMURA, T., and NAKAMURA, M., 1985, The microsphere method facilitates statistical assessment of regional blood flow. *Basic Research in Cardiology*, **80**, 417-429.

OLESON, J. R., SIM, D. A., AND MANNING, M. R., 1984, Analysis of prognostic variables in hyperthermia treatment of 161 patients. *International Journal of Radiation Oncology, Biology, and Physics*, **10**, 2231-2239.

OVERGAARD, J., 1980, Hyperthermia modification of the radiation response in solid tumors. *Biological Bases and Clinical Implications of Tumor Radioresistance*, edited by G. Fletcher and H. Withers (New York: Masson), pp. 337-352.

OVERGAARD, J., 1981, Fractionated radiation and hyperthermia: experimental and clinical studies. *Cancer*, **48**, 1116-1123.

OVERGAARD, J., 1985, Rationale and problems in the design of clinical studies. *Hyperthermic Oncology 1984*, vol. 2, edited by J. Overgaard (London: Taylor & Francis), pp. 325-338.

OVERGAARD, J., 1990, The current and potential role of hyperthermia in radiotherapy. *International Journal of Radiation Oncology Biology and Physics*, **16**, 535-549.

PATTERSON, J., and STRANG, R., 1979, The role of blood flow in hyperthermia. *International Journal of Radiation Oncology, and Biology, Physics*, **5**, 235-241.

PEARCE, J., 1968, Renal nervous and spinal pathways and the reflex regulation of extracellular fluid volume. *Fed. Proc*, **27**, 1142-1148.

PENNES, H. H., 1948, Analysis of tissue and arterial blood temperatures in the resting human forearm, *Journal of Applied Physiology*, **1**, 93-122.

PEREZ, C. A., KUSKE, R. R., EMAMI, B., and FINEBERG, B., 1986, Irradiation alone or combined with hyperthermia in the treatment of recurrent carcinoma of the breast in the chest wall: a non-randomized study. *International journal of hyperthermia*, **2**, 179-188.

PEREZ, C. A., NUSSBAUM, G., EMAMI, B., and VONGERICHTEN, D., 1983, Clinical results of irradiation combined with local hyperthermia. *Cancer*, **52**, 1597-1603.

PRADOS, M. D., GUTIN, P. H., PHILLIPS, T. L., WARA, W. M., SNEED, P. K., LARSON, D. A., LAMB, S. A., HAM, B., MALEC, M. K., and WILSON, C. B., 1992, Interstitial brachytherapy for newly diagnosed patients with malignant gliomas: The UCSF experience, *International Journal of Radiation Oncology Biology and Physics*, **24**, 593-597.

RAAPHORST, G. P., 1990, Fundamental aspects of hyperthermic biology. *An Introduction to The Practical Aspects Of Clinical Hyperthermia*, edited by S. B. Field & J. W. Hand, (London: Taylor and Francis), 10-53.

RAAPHORST, G. P., and FEELEY, M. M., 1990, Comparison of recovery from potentially lethal damage after exposure to hyperthermia and radiation. *Radiation Research*, **121**, 107-110.

RAAPHORST, G. P., FREEMAN, M. L., and DEWEY, W. C., 1979a, Radiosensitivity and recovery from radiation damage in cultured CHO cells exposed to hyperthermia at 42.5°C or 45.5°C. *Radiation Research*, **79**, 390-402.

RAAPHORST, G. P., ROMANO, S. L., MITCHELL, J. B., BEDFORD, J. S., and DEWEY, W. C., 1979b, Intrinsic differences in heat and/or X-ray sensitivity of seven mammalian cell lines cultured and treated under identical conditions. *Cancer Research*, **39**, 396-401.

REINHOLD, H. S., WIKE-HOOLEY, J. L., VAN DEN BERG, A. P., and VAN DEN BERG-BLOK, A., 1985, Environmental factors, blood flow and microcirculation. *Hyperthermic Oncology 1984*, vol. 2, edited by J. Overgaard (London: Taylor & Francis), pp. 41-52.

REINHOLD, H. S., and ENDRICH, B., 1986, Tumor microcirculation as a target for hyperthermia. *International Journal of Hyperthermia*, **2**, 111-137.

REINHOLD, H. S., 1988, Physiological effects of hyperthermia. *Recent Advances in Cancer Research*, **107**, 32-43.

ROBERTS, D. W., COUGHLIN, C. T., WONG, T. Z., FRATKIN, J. D., DOUPLE, E. B., and STROHBEHN, J. W., 1986, Interstitial hyperthermia and

iridium brachytherapy in treatment of malignant glioma. *Journal of Neurosurgery*, **64**, 581-587.

ROBERTS, D. W., STROHBEHN, J. W., COUGHLIN, C. T., RYAN, T. P., LYONS, B. E., and DOUPLE, E. B., 1987, Iridium-192 brachytherapy in combination with microwave-induced hyperthermia for malignant glioma. *Applied Neurophysiology*, **50**, 287-291.

ROBINSON, J. E., HARRISON, G. H., McCREDY, W. A., and SAMARAS, G. M., 1978, Good thermometry is essential to good hyperthermia research. *British Journal of Radiology*, **51**, 532-534.

ROEMER, R. B., 1988, Heat transfer in hyperthermia treatments: basic principles and applications. *Biological, Physical and Clinical Aspects of Hyperthermia*, edited by B. R. Paliwal, F. W. Hetzel, and M. W. Dewhirst (New York: AIP), pp. 210-242.

ROEMER, R. B., MOROS, E. G., HYNYNEN, K., 1989, A comparison of bioheat transfer and effective conductivity equation predictions to experimental hyperthermia data. In: *Bioheat Transfer-Applications in Hyperthermia, Emerging Horizons in Instrumentation and Modeling*, edited by R. B. Roemer, J. J. McGrath and H. F. Bowman. (American Society of Mechanical Engineers, Bioengineering Division, New York, USA) **BED-12**, pp. 11-15.

SATOH, T., NAKASONE, S., and NISHIMOTO, A., 1989, Cerebral blood flow response to the tissue temperature in tumor and brain tissues. *International Journal of Hyperthermia*, **5**, 683-696.

SATOH, T., STAUFFER, P. R., FIKE, J. R., 1988, Thermal distributions studies of helical coil microwave antennas for interstitial hyperthermia. *International Journal of Radiation Oncology Biology and Physics*, **15**, 1209-1218.

SEELY, J. F., and BOULPAEP, E. L., 1971, Renal function studies on the isobaric autoperfused dog kidney. *American Journal of Physiology*, **221**, 1075-1083.

SEKINS, K. M., and EMERY, A., 1982, Thermal science for physical medicine. In: *Therapeutic Heat and Cold*, edited by J. Lehmann (Baltimore: Williams and Wilkins), pp. 70-132.

SEKINS, K. M., EMERY, A. F., LEHMANN, J. F., and MACDOUGALL, J. A., 1984, *Am. Soc. Mech. Eng. J. Biomech. Eng.*, **104**, 272-279.

SCOTT, R. S., JOHNSON, R. J. R., STORY, K. V., and CLAY, L., 1984, Local hyperthermia in combination with definitive radiotherapy: Increased tumor

clearance, reduced recurrence rate in extended follow-up. *International Journal of Radiation Oncology, Biology, and Physics*, **10**, 2119-2123.

SHAPIRO, W. R., 1986, Therapy of adult malignant brain tumors: What have the clinical trails taught us ? *Seminars in Oncology*, **13**, 38-45.

SHEA, S. M., CAULFIELD, J. B., and BURKE, J. F., 1973, Microvascul ultrastructure in thermal injury: a reconsideration of the role of mediators. *Microvascular Research*, **5**, 87-96.

SHELINE, G. E., 1980, Irradiation injury of the human brain: A review of clinical experience. *Radiation Damage to the Central Nervous System* (H. A. Gilbert and A. R. Kagan, Eds.) pp 39-58. Raven, New York.

SHRIVASTAV, S., KAEIN Jr., W. G., JOINES, W. T. and JIRTLE, R. L., 1983, Microwave hyperthermia and its effects on tumor blood flow in rats. *Cancer Research*, **43**, 4665-4669.

SILVA, M. R. E., and ANTONIO, A., 1960, Release of bradykinin and the mechanism of production of a 'thermic edema (45°C) in the rat's paw. *Medicina experimentalis*, **3**, 371-382.

SMITH, H. W., 1951, *Structure and function in health and Disease*. Oxford Press, New York, 411-460.

SNEED, P. K., GUTIN, P. H., STAUFFER, P. R., PHILLIPS, T. L., SUEN, S., WEAVER, K. A., LAMB, S. A., HAM, B., PARADOS, M. D., LARSON, D. A., and WARA, W. M., 1992, Thermoradiotherapy of recurrent malignant brain tumors. *International Journal of Radiation Oncology Biology and Physics*, **23**, 853-861.

SNEED, P. K., STAUFFER, P. R., GUTIN, P. H., PHILLIPS, T. L., SUEN, S., WEAVER, K. A., LAMB, S. A., HAM, B., PARADOS, M. D., LARSON, D. A., and WARA, W. M., 1991, Interstitial irradiation and hyperthermia for the treatment of recurrent malignant brain tumors. *Neurosurgery*, **28**, 206-214.

SONG, C. W., 1984, Effect of hyperthermia on blood flow and microenvironment: A review. *Cancer Research*, **44** (Suppl.) 4721s-4730s.

SONG, C. W., KANG, M. S., RHEE, J. G., and LEVITT, S. H., 1980a, Effect of hyperthermia on vascular function in normal and neoplastic tissues. *Annals of the New York Academy of Sciences*, **335**, 35-47.

SONG, C. W., PATTEN, M., CHELSTROM, L. M., RHEE, J. G. and LEVITT, S. H., 1987a, Effect of multiple heatings on the blood flow in RIF-1 tumors, skin and muscle of C3H mice. *International Journal of Hyperthermia*, **3**, 535-545.

SONG, C. W., RHEE, J. G. and HAUMSCHILD, D. J., 1987b, Continuous and noninvasive quantitation of heat-induced changes in blood flow in the skin and RIF-1 tumor of mice by laser doppler flowmetry. *International Journal of Hyperthermia*, **3**, 71-77.

SONG, C. W., RHEE, J. G., and LEVITT, S. H., 1980b, Blood flow in normal tissues and tumors during hyperthermia. *Journal of the National Cancer Institute*, **64**, 119-124.

STRANG, R., and PATTERSON, J., 1980, The role of thermal conduction in hyperthermia. *International Journal of Radiation Oncology Biology and Physics*, **6**, 729-735.

STROHBEHN, J. W., PAULSEN, K. D., and LYNCH, D. R., 1986, Use of the finite difference method in computerized thermal dosimetry. In: *Handbook of Techniques for Clinical Hyperthermia*, edited by J. W. Hand and J. R. James (Chichester: Research Studies Press, Ltd), pp. 383-451.

TAKEUCHI, J., AOKI, S., NOMURA, G., MIZUMURA, Y., SHIMIZU, H., AND KUBO, T., 1971, Nervous control of renal circulation: on the existence of sympathetic cholinergic fibers. *Journal of Applied Physiology*, **31**, 686-692.

THUNING, C. A., BAKIR, N. A., and WARREN, J., 1980, Synergistic effect of combined hyperthermia and a nitrosourea in treatment of a murine ependymoblastoma. *Cancer Research*, **40**, 2726-2729.

TRANQUILLI, W. J., MANOHAR, M., PARKS, C. M., THURMAN, J. C., THEODORAKIS, C., and BENSON, G. J., 1982, Systematic and regional blood flow distribution in unanesthetized swine and swine anesthetized with halothane + nitrous oxide, halothane or enflurane. *Anesthesiology*, **56**, 369-379.

TSOUFLANIDIS, N., 1983, Measurement and detection of radiation. McGraw-Hill Series in Nuclear Engineering, Toronto.

VALVANO, J. W., ALLEN, J. T., and BOWMAN, H. F., 1984, *Journal of biomechanical Engineering*, **10**, 192-197.

VAN DER ZEE, J., VAN RHOON, G. C., WIKE-HOOLEY, J. L., and REINHOLD, H. S., 1985, Clinical derived dose effect relationship for hyperthermia given in combination with low dose radiotherapy. *British Journal of Radiology*, **58**, 243-250.

- VAN DER ZEE, J., VAN DER PUTTEN, W. J. L., VAN DER BERG, A. P., VAN RHOON, G. C., WIKE-HOOLEY, J. L., BROEKMEYER-REURINK, M. P., and REINHOLD, H. S., 1986, Retrospective analysis of the response of tumors in patients treated with a combination of radiotherapy and hyperthermia. *International Journal of Hyperthermia*, **2**, 337-350.
- VAUPEL, P., and KALLINOWSKI, F., 1987, Physiological effects of hyperthermia. *Recent Results in Cancer Research*, **104**, 71-108.
- WALKER, M. D., ALEXANDER, J. Jr., HUNT, W. E., MACCARTY, C. S., MAHALEY, M. S. Jr., NORRELL, H. A., OWENS, G., RANSOHOFF, J., WILSON, C. B., GEHAN, E. A., and STRIKE, T. A., 1978, Evaluation of BCNU and/or radiotherapy in the treatment of anaplastic gliomas. *Journal of Neurosurgery*, **49**, 333-343.
- WALKER, M. D., STRIKE, T. A., SHELINE, G. E., 1979, An analysis of dose-effect relationship in the radiotherapy of malignant gliomas. *International Journal of Radiation Biology*, **19**, 467-477.
- WATERMAN, F. J., NERLINGER, R. E., MOYLAN, D. J., and LEEPER, D. E., 1987, Response of human tumor blood flow to local hyperthermia. *International Journal of Radiation Oncology Biology and Physics*, **13**, 75-82.
- WEINABAUM, S., JIJI, L. M., 1985, A new simplified bioheat equation for the effect of blood flow on local average tissue temperature. *ASME Journal of Biomechanical Engineering*, **107**, 131-139.
- WEINABAUM, S., JIJI, L. M., and LEMONS, D. E., 1984a, Theory and experiment for the effect of vascular microstructure on surface tissue heat transfer. Part I: anatomical foundation and model conceptualization. *Journal of Biomechanical Engineering*, **106**, 321-330.
- WEINABAUM, S., JIJI, L. M., and LEMONS, D. E., 1984b, Theory and experiment for the effect of vascular microstructure on surface tissue heat transfer. Part II: model formulation and solution. *Journal of Biomechanical Engineering*, **106**, 331-341.
- WEBER, A., 1991, Comparison of four methods of measuring blood flow. M.Sc thesis, Carleton University, Ottawa, Ontario Canada.
- WEXLER, I., and KAO, F., 1970, Neural and humoral factors affecting canine renal blood flow during induced muscular work. *American Journal of Physiology*, **218**, 755-761.

WINGET, J. M., DEWHIRST, M. W., ENGLER, M. J., and OLESON, J. R., 1986, The use of limited temperature observations to predict complete temperature fields. *IEEE Eighth Annual Conference of the Engineering in Medicine and Biology Society*, pp. 1507-1511.

END

3 1 - 0 5 - 9 | 6

FIN